


Benzaldehyde

Response Ratio




$$\text{Response} = 6.030\text{e-}001 * \text{Amt} + 3.146\text{e-}001$$

Coef of Det (r^2) = 0.990967 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010521.M


Calibration Table Last Updated: Tue Jan 05 16:45:30 2021

Benzoic acid




$\text{Response} = 2.267\text{e-}001 * \text{Amt} - 5.179\text{e-}002$
 Coef of Det (r^2) = 0.990633 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010521.M
 Calibration Table Last Updated: Tue Jan 05 16:45:30 2021

Hexachlorocyclopentadiene



$R = 6.369e-002 A^2 + 4.320e-003 A + 8.145e-003$
Coef of Det (r^2) = 0.997796 Curve Fit: Quadratic
Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010521.M
Calibration Table Last Updated: Tue Jan 05 16:45:30 2021

2,4-Dinitrophenol



$R = 3.681e-002 A^2 + 6.858e-002 A + 3.951e-003$
 Coef of Det (r^2) = 0.993332 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010521.M
 Calibration Table Last Updated: Tue Jan 05 16:45:30 2021