

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010524\
 Data File : BF136967.D
 Acq On : 05 Jan 2024 10:16
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/06/2024
 Supervised By :mohammad ahmed 01/06/2024

Quant Time: Jan 05 10:50:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	75943	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	284941	20.000	ng	0.00
39) Acenaphthene-d10	9.822	164	142739	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	256194	20.000	ng	0.00
76) Chrysene-d12	13.974	240	133421	20.000	ng	0.00
86) Perylene-d12	15.463	264	144248	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	393101	80.764	ng	0.00
7) Phenol-d6	6.445	99	487436	77.288	ng	0.00
23) Nitrobenzene-d5	7.357	82	443356	79.620	ng	0.00
42) 2,4,6-Tribromophenol	10.622	330	126480	75.238	ng	0.00
45) 2-Fluorobiphenyl	9.145	172	869367	81.918	ng	0.00
79) Terphenyl-d14	12.910	244	768146	80.457	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.540	88	81758	40.315	ng	95
3) Pyridine	3.305	79	195032	39.416	ng	96
4) n-Nitrosodimethylamine	3.287	42	98051	37.107	ng	96
6) Aniline	6.457	93	229422	36.384	ng	# 26
8) 2-Chlorophenol	6.581	128	212798	39.952	ng	96
9) Benzaldehyde	6.345	77	133589	37.239	ng	98
10) Phenol	6.457	94	255049	37.883	ng	91
11) bis(2-Chloroethyl)ether	6.528	93	197854	38.643	ng	99
12) 1,3-Dichlorobenzene	6.728	146	233035	40.106	ng	98
13) 1,4-Dichlorobenzene	6.804	146	235131	39.604	ng	97
14) 1,2-Dichlorobenzene	6.957	146	220360	39.888	ng	99
15) Benzyl Alcohol	6.940	79	161045	37.850	ng	98
16) 2,2'-oxybis(1-Chloropr...	7.063	45	310321	37.072	ng	76
17) 2-Methylphenol	7.051	107	166906	37.825	ng	96
18) Hexachloroethane	7.292	117	85923	39.850	ng	91
19) n-Nitroso-di-n-propyla...	7.204	70	145773	38.129	ng	95
20) 3+4-Methylphenols	7.210	107	214975	38.997	ng	# 54
22) Acetophenone	7.198	105	291609	40.830	ng	# 89
24) Nitrobenzene	7.375	77	223722	40.107	ng	93
25) Isophorone	7.610	82	395309	39.015	ng	99
26) 2-Nitrophenol	7.687	139	114489	42.899	ng	98
27) 2,4-Dimethylphenol	7.728	122	133876	41.204	ng	95
28) bis(2-Chloroethoxy)met...	7.816	93	247556	40.195	ng	100
29) 2,4-Dichlorophenol	7.934	162	171047	40.891	ng	96
30) 1,2,4-Trichlorobenzene	8.004	180	191685	41.129	ng	98
31) Naphthalene	8.087	128	638732	40.803	ng	99
32) Benzoic acid	7.875	122	119275m	37.941	ng	
33) 4-Chloroaniline	8.139	127	222147	38.436	ng	98
34) Hexachlorobutadiene	8.198	225	117377	41.452	ng	98
35) Caprolactam	8.528	113	57609	41.748	ng	91
36) 4-Chloro-3-methylphenol	8.634	107	179746	38.170	ng	99
37) 2-Methylnaphthalene	8.781	142	401047	39.809	ng	100
38) 1-Methylnaphthalene	8.881	142	394765	39.941	ng	99
40) 1,2,4,5-Tetrachloroben...	8.945	216	189664	42.905	ng	100
41) Hexachlorocyclopentadiene	8.928	237	63874	47.441	ng	98
43) 2,4,6-Trichlorophenol	9.063	196	114060	40.233	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF010524\
 Data File : BF136967.D
 Acq On : 05 Jan 2024 10:16
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/06/2024
 Supervised By :mohammad ahmed 01/06/2024

Quant Time: Jan 05 10:50:18 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.116	196	128033	40.527	ng	99
46) 1,1'-Biphenyl	9.245	154	490418	40.943	ng	98
47) 2-Chloronaphthalene	9.269	162	376298	41.325	ng	100
48) 2-Nitroaniline	9.375	65	108539	38.064	ng	96
49) Acenaphthylene	9.686	152	557229	40.036	ng	100
50) Dimethylphthalate	9.545	163	403041	38.657	ng	99
51) 2,6-Dinitrotoluene	9.616	165	92123	38.933	ng	91
52) Acenaphthene	9.857	154	341599	39.594	ng	99
53) 3-Nitroaniline	9.786	138	91387	36.459	ng	99
54) 2,4-Dinitrophenol	9.904	184	47944	38.588	ng	96
55) Dibenzofuran	10.028	168	510350	39.023	ng	99
56) 4-Nitrophenol	9.969	139	58360	34.490	ng	97
57) 2,4-Dinitrotoluene	10.028	165	113222	36.859	ng	94
58) Fluorene	10.375	166	407269	39.247	ng	98
59) 2,3,4,6-Tetrachlorophenol	10.157	232	87585	36.015	ng	97
60) Diethylphthalate	10.251	149	408509	37.763	ng	97
61) 4-Chlorophenyl-phenyle...	10.363	204	199375	39.527	ng	99
62) 4-Nitroaniline	10.410	138	84394m	34.926	ng	
63) Azobenzene	10.528	77	374879	37.103	ng	97
65) 4,6-Dinitro-2-methylph...	10.439	198	58247	39.671	ng	85
66) n-Nitrosodiphenylamine	10.486	169	346657	41.654	ng	99
67) 4-Bromophenyl-phenylether	10.857	248	120016	42.182	ng	98
68) Hexachlorobenzene	10.928	284	136111	42.927	ng	92
69) Atrazine	11.022	200	81877	38.465	ng	98
70) Pentachlorophenol	11.128	266	56795	38.429	ng	98
71) Phenanthrene	11.339	178	524641	39.910	ng	99
72) Anthracene	11.392	178	524828	39.423	ng	99
73) Carbazole	11.557	167	485525	38.682	ng	99
74) Di-n-butylphthalate	11.880	149	607005	39.632	ng	100
75) Fluoranthene	12.533	202	521934	37.119	ng	97
77) Benzidine	12.663	184	180200	45.806	ng	99
78) Pyrene	12.769	202	520994	39.778	ng	99
80) Butylbenzylphthalate	13.386	149	192939	40.484	ng	97
81) Benzo(a)anthracene	13.963	228	374835	40.459	ng	98
82) 3,3'-Dichlorobenzidine	13.927	252	114218	43.411	ng	95
83) Chrysene	14.004	228	356179	39.314	ng	99
84) Bis(2-ethylhexyl)phtha...	13.945	149	265645	44.302	ng	100
85) Di-n-octyl phthalate	14.568	149	436425	47.052	ng	99
87) Indeno(1,2,3-cd)pyrene	16.986	276	443195	42.554	ng	99
88) Benzo(b)fluoranthene	15.027	252	355393	36.888	ng	100
89) Benzo(k)fluoranthene	15.057	252	343066	37.684	ng	99
90) Benzo(a)pyrene	15.398	252	319127	39.241	ng	98
91) Dibenzo(a,h)anthracene	17.004	278	363523	42.545	ng	97
92) Benzo(g,h,i)perylene	17.451	276	369218	42.190	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.787 min Scan# 79
Delta R.T. -0.005 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:152 Resp: 7594
Ion Ratio Lower Upper
152 100
150 156.8 127.0 190.6
115 57.8 49.0 73.4

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#2
1,4-Dioxane
Concen: 40.315 ng
RT: 2.540 min Scan# 77
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion: 88 Resp: 81758
Ion Ratio Lower Upper
88 100
58 79.2 67.4 101.0
43 33.3 27.7 41.5






#3
Pyridine
Concen: 39.416 ng
RT: 3.305 min Scan# 207
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :




SSTDCCC040




Tgt Ion: 79 Resp: 19503
Ion Ratio Lower Upper
79 100
52 81.9 67.7 101.5
51 39.5 34.0 51.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#4
n-Nitrosodimethylamine
Concen: 37.107 ng
RT: 3.287 min Scan# 204
Delta R.T. 0.012 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion: 42 Resp: 98051
Ion Ratio Lower Upper
42 100
74 124.9 96.0 144.0
44 7.6 5.4 8.2





#5
2-Fluorophenol
Concen: 80.764 ng
RT: 5.428 min Scan# 50
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:112 Resp: 39310
Ion Ratio Lower Upper
112 100
64 60.8 51.1 76.7
63 30.0 26.1 39.1

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#6
Aniline
Concen: 36.384 ng
RT: 6.457 min Scan# 743
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16


Tgt Ion: 93 Resp: 229422
Ion Ratio Lower Upper
93 100
66 90.9 34.6 51.8#
65 57.6 17.8 26.8#





#7
Phenol-d6
Concen: 77.288 ng
RT: 6.445 min Scan# 741
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 99 Resp: 487430
Ion Ratio Lower Upper
99 100
42 18.7 16.1 24.1
71 31.3 27.4 41.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#8
2-Chlorophenol
Concen: 39.952 ng
RT: 6.581 min Scan# 764
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Tgt Ion:128 Resp: 212798
Ion Ratio Lower Upper
128 100
130 31.4 11.3 51.3
64 48.0 32.5 72.5





#9
Benzaldehyde
Concen: 37.239 ng
RT: 6.345 min Scan# 71
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion: 77 Resp: 133589
Ion Ratio Lower Upper
77 100
105 104.8 84.8 124.8
106 100.9 77.8 117.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#10
Phenol
Concen: 37.883 ng
RT: 6.457 min Scan# 743
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion: 94 Resp: 255049
Ion Ratio Lower Upper
94 100
65 38.9 16.6 56.6
66 61.4 32.6 72.6





#11

bis(2-Chloroethyl)ether

Concen: 38.643 ng

RT: 6.528 min Scan# 70

Delta R.T. -0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion: 93 Resp: 197854
 Ion Ratio Lower Upper
 93 100
 63 81.6 61.2 101.2
 95 33.1 11.6 51.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#12

1,3-Dichlorobenzene

Concen: 40.106 ng




RT: 6.728 min Scan# 789


Delta R.T. -0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Tgt Ion:146 Resp: 233035
 Ion Ratio Lower Upper
 146 100
 148 63.2 50.5 75.7
 75 29.1 26.2 39.4






#13
1,4-Dichlorobenzene
Concen: 39.604 ng
RT: 6.804 min Scan# 803
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:146 Resp: 23513
Ion Ratio Lower Upper
146 100
148 63.9 49.2 73.8
111 42.0 35.2 52.8

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#14
1,2-Dichlorobenzene
Concen: 39.888 ng
RT: 6.957 min Scan# 828
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16


Tgt Ion:146 Resp: 220360
Ion Ratio Lower Upper
146 100
148 61.7 49.6 74.4
111 44.7 36.2 54.2





#15
Benzyl Alcohol
Concen: 37.850 ng
RT: 6.940 min Scan# 81
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion: 79 Resp: 16104
Ion Ratio Lower Upper
79 100
108 78.8 61.0 91.6
77 62.1 49.8 74.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#16
2,2'-oxybis(1-Chloropropane)
Concen: 37.072 ng
RT: 7.063 min Scan# 846
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion: 45 Resp: 310321
Ion Ratio Lower Upper
45 100
77 23.5 0.0 34.3
79 21.2 0.0 31.4






#17
2-Methylphenol
Concen: 37.825 ng
RT: 7.051 min Scan# 844
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:107 Resp: 166900
Ion Ratio Lower Upper
107 100
108 113.5 90.3 135.5
77 54.9 38.7 58.1
79 52.2 37.8 56.6

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#18
Hexachloroethane
Concen: 39.850 ng
RT: 7.292 min Scan# 885
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:117 Resp: 85923
Ion Ratio Lower Upper
117 100
119 99.5 75.8 113.6
201 98.5 69.1 103.7






#19
n-Nitroso-di-n-propylamine
Concen: 38.129 ng
RT: 7.204 min Scan# 81
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion: 70 Resp: 14577
Ion Ratio Lower Upper
70 100
42 62.1 54.2 81.4
101 10.0 7.6 11.4
130 21.6 17.1 25.7

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#20
3+4-Methylphenols
Concen: 38.997 ng
RT: 7.210 min Scan# 871
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:107 Resp: 214975
Ion Ratio Lower Upper
107 100
108 94.3 66.0 106.0
77 41.0 114.2 154.2#
79 30.3 7.2 47.2






#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.063 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:136 Resp: 28494
Ion Ratio Lower Upper
136 100
137 10.6 9.0 13.4
54 8.6 7.2 10.8
68 5.8 4.9 7.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#22
Acetophenone
Concen: 40.830 ng
RT: 7.198 min Scan# 869
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:105 Resp: 291609
Ion Ratio Lower Upper
105 100
71 5.0 4.8 7.2
51 31.1 32.8 49.2#
120 23.1 17.0 25.6





#23

Nitrobenzene-d5

Concen: 79.620 ng

RT: 7.357 min Scan# 896

Delta R.T. 0.000 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion: 82 Resp: 443350
 Ion Ratio Lower Upper
 82 100
 128 48.6 35.3 52.9
 54 52.5 44.5 66.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024
 Supervised By :mohammad ahmed 01/06/2024



#24

Nitrobenzene

Concen: 40.107 ng




RT: 7.375 min Scan# 899


Delta R.T. 0.000 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Tgt Ion: 77 Resp: 223722
 Ion Ratio Lower Upper
 77 100
 123 51.6 36.7 55.1
 65 13.7 11.8 17.6






#25
Isophorone
Concen: 39.015 ng
RT: 7.610 min Scan# 910
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion: 82 Resp: 395309
Ion Ratio Lower Upper
82 100
95 6.7 5.5 8.3
138 18.5 14.5 21.7


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#26
2-Nitrophenol
Concen: 42.899 ng
RT: 7.687 min Scan# 952
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:139 Resp: 114489
Ion Ratio Lower Upper
139 100
109 27.0 21.9 32.9
65 46.0 38.6 58.0






#27
2,4-Dimethylphenol
Concen: 41.204 ng
RT: 7.728 min Scan# 91
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:122 Resp: 133870
Ion Ratio Lower Upper
122 100
107 108.2 92.0 138.0
121 56.9 47.5 71.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#28
bis(2-Chloroethoxy)methane
Concen: 40.195 ng
RT: 7.816 min Scan# 974
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion: 93 Resp: 247556
Ion Ratio Lower Upper
93 100
95 32.3 25.5 38.3
123 11.5 9.2 13.8





#29

2,4-Dichlorophenol

Concen: 40.891 ng

RT: 7.934 min Scan# 994

Delta R.T. 0.000 min

Lab File: BF136967.D


Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:	162	Resp:	17104
Ion Ratio	Lower	Upper	
162	100		
164	63.6	41.7	81.7
98	35.9	19.5	59.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#30

1,2,4-Trichlorobenzene


Concen: 41.129 ng

RT: 8.004 min Scan# 1006



Delta R.T. -0.012 min


Lab File: BF136967.D

Acq: 05 Jan 2024 10:16



Tgt Ion:	180	Resp:	191685
Ion Ratio	Lower	Upper	
180	100		
182	93.6	75.4	113.0
145	32.0	23.1	34.7






#31
Naphthalene
Concen: 40.803 ng
RT: 8.087 min Scan# 1021
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:128 Resp: 638732
Ion Ratio Lower Upper
128 100
129 10.5 8.7 13.1
127 13.5 10.8 16.2


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#32
Benzoic acid
Concen: 37.941 ng m
RT: 7.875 min Scan# 984
Delta R.T. 0.018 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:122 Resp: 119275
Ion Ratio Lower Upper
122 100
105 122.2 100.8 140.8
77 89.9 70.9 110.9





#33

4-Chloroaniline

Concen: 38.436 ng

RT: 8.139 min Scan# 1029

Delta R.T. -0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:	127	Resp:	22214
Ion	Ratio	Lower	Upper
127	100		
129	31.4	26.3	39.5
65	31.2	25.8	38.8
92	19.4	15.8	23.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#34

Hexachlorobutadiene

Concen: 41.452 ng




RT: 8.198 min Scan# 1039


Delta R.T. -0.012 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Tgt Ion:	225	Resp:	117377
Ion	Ratio	Lower	Upper
225	100		
223	61.5	50.6	76.0
227	63.9	51.9	77.9









#35
Caprolactam
Concen: 41.748 ng
RT: 8.528 min Scan# 1094
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:113 Resp: 57609
Ion Ratio Lower Upper
113 100
55 169.6 163.1 203.1
56 130.0 121.7 161.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#36
4-Chloro-3-methylphenol
Concen: 38.170 ng
RT: 8.634 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16


Tgt Ion:107 Resp: 179746
Ion Ratio Lower Upper
107 100
144 24.9 20.2 30.2
142 78.3 62.3 93.5





#37
2-Methylnaphthalene
Concen: 39.809 ng
RT: 8.781 min Scan# 1139
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:142 Resp: 40104
Ion Ratio Lower Upper
142 100
141 88.8 71.1 106.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#38
1-Methylnaphthalene
Concen: 39.941 ng
RT: 8.881 min Scan# 1155
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:142 Resp: 394765
Ion Ratio Lower Upper
142 100
141 92.7 73.6 110.4






#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.822 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:164 Resp: 142739
Ion Ratio Lower Upper
164 100
162 99.2 81.6 122.4
160 44.3 34.9 52.3

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#40
1,2,4,5-Tetrachlorobenzene
Concen: 42.905 ng
RT: 8.945 min Scan# 1166
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:216 Resp: 189664
Ion Ratio Lower Upper
216 100
214 77.9 62.3 93.5
179 20.7 16.4 24.6
108 16.8 14.0 21.0






#41
Hexachlorocyclopentadiene
Concen: 47.441 ng
RT: 8.928 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:237 Resp: 63874
Ion Ratio Lower Upper
237 100
235 64.7 43.2 83.2
272 15.5 0.0 34.2


Manual Integrations

APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#42
2,4,6-Tribromophenol
Concen: 75.238 ng
RT: 10.622 min Scan# 1451
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Tgt Ion:330 Resp: 126480
Ion Ratio Lower Upper
330 100
332 94.4 77.4 116.0
141 27.3 23.4 35.2





#43
2,4,6-Trichlorophenol
Concen: 40.233 ng
RT: 9.063 min Scan# 1186
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:196 Resp: 114060
Ion Ratio Lower Upper
196 100
198 97.1 75.1 112.7
200 30.6 22.5 33.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#44
2,4,5-Trichlorophenol
Concen: 40.527 ng
RT: 9.116 min Scan# 1195
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:196 Resp: 128033
Ion Ratio Lower Upper
196 100
198 95.5 77.0 115.6
97 43.5 36.1 54.1
132 27.1 21.8 32.6





#45
2-Fluorobiphenyl
Concen: 81.918 ng
RT: 9.145 min Scan# 1200
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

Tgt Ion:172 Resp: 86936
Ion Ratio Lower Upper
172 100
171 35.4 29.0 43.6
170 23.2 19.0 28.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#46
1,1'-Biphenyl
Concen: 40.943 ng
RT: 9.245 min Scan# 1217
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:154 Resp: 490418
Ion Ratio Lower Upper
154 100
153 39.8 21.1 61.1
76 11.7 0.0 31.7






#47
2-Chloronaphthalene
Concen: 41.325 ng
RT: 9.269 min Scan# 1221
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:162 Resp: 376298
Ion Ratio Lower Upper
162 100
127 41.8 33.6 50.4
164 32.4 25.7 38.5

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#48
2-Nitroaniline
Concen: 38.064 ng
RT: 9.375 min Scan# 1239
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion: 65 Resp: 108539
Ion Ratio Lower Upper
65 100
92 67.7 53.2 79.8
138 107.2 81.1 121.7





#49
Acenaphthylene
Concen: 40.036 ng
RT: 9.686 min Scan# 1293
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F





ClientSampleId :

SSTDCCC040

Tgt Ion:152 Resp: 557229
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4
153 13.0 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#50
Dimethylphthalate
Concen: 38.657 ng
RT: 9.545 min Scan# 1268
Delta R.T. -0.012 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:163 Resp: 403041
Ion Ratio Lower Upper
163 100
194 3.8 3.3 4.9
164 10.2 7.8 11.6





#51
2,6-Dinitrotoluene
Concen: 38.933 ng
RT: 9.616 min Scan# 1280
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:165 Resp: 9212
Ion Ratio Lower Upper
165 100
63 56.9 39.8 59.6
89 53.9 38.8 58.2

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#52
Acenaphthene
Concen: 39.594 ng
RT: 9.857 min Scan# 1321
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:154 Resp: 341599
Ion Ratio Lower Upper
154 100
153 114.5 90.7 136.1
152 54.1 43.7 65.5






#53
3-Nitroaniline
Concen: 36.459 ng
RT: 9.786 min Scan# 11
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:138 Resp: 9138
Ion Ratio Lower Upper
138 100
108 11.3 8.6 12.8
92 117.6 95.2 142.8

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#54
2,4-Dinitrophenol
Concen: 38.588 ng
RT: 9.904 min Scan# 1329
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:184 Resp: 47944
Ion Ratio Lower Upper
184 100
63 54.3 40.9 61.3
154 55.8 42.7 64.1






#55
Dibenzenofuran
Concen: 39.023 ng
RT: 10.028 min Scan# 11
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:168 Resp: 510350
Ion Ratio Lower Upper
168 100
139 36.2 29.3 43.9
169 13.6 11.0 16.4

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#56
4-Nitrophenol
Concen: 34.490 ng
RT: 9.969 min Scan# 1340
Delta R.T. 0.012 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:139 Resp: 58360
Ion Ratio Lower Upper
139 100
109 61.8 45.8 85.8
65 99.5 77.8 117.8






#57
2,4-Dinitrotoluene
Concen: 36.859 ng
RT: 10.028 min Scan# 11
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:165 Resp: 11322
Ion Ratio Lower Upper
165 100
63 58.9 41.0 61.6
89 73.0 59.9 89.9
182 2.7 2.0 3.0

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#58
Fluorene
Concen: 39.247 ng
RT: 10.375 min Scan# 1409
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:166 Resp: 407269
Ion Ratio Lower Upper
166 100
165 94.0 73.7 110.5
167 13.2 10.4 15.6





#59
2,3,4,6-Tetrachlorophenol
Concen: 36.015 ng
RT: 10.157 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:232 Resp: 8758
Ion Ratio Lower Upper
232 100
131 41.9 35.1 52.7
130 1.9 1.6 2.4
166 28.1 23.6 35.4

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#60
Diethylphthalate
Concen: 37.763 ng
RT: 10.251 min Scan# 1388
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16


Tgt Ion:149 Resp: 408509
Ion Ratio Lower Upper
149 100
177 22.9 16.9 25.3
150 12.3 9.3 13.9





#61
4-Chlorophenyl-phenylether
Concen: 39.527 ng
RT: 10.363 min Scan# 1407
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:204 Resp: 199379
Ion Ratio Lower Upper
204 100
206 32.5 25.6 38.4
141 59.5 48.4 72.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#62
4-Nitroaniline
Concen: 34.926 ng m
RT: 10.410 min Scan# 1415
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:138 Resp: 84394
Ion Ratio Lower Upper
138 100
92 43.8 32.1 72.1
108 58.6 50.1 90.1






#63
Azobenzene
Concen: 37.103 ng
RT: 10.528 min Scan# 1436
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion: 77 Resp: 374879
Ion Ratio Lower Upper
77 100
182 27.2 6.0 46.0
105 24.2 2.3 42.3
51 29.7 11.1 51.1

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.316 min Scan# 1569
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:188 Resp: 256194
Ion Ratio Lower Upper
188 100
94 8.8 7.5 11.3
80 9.6 8.1 12.1









#65
4,6-Dinitro-2-methylphenol
Concen: 39.671 ng
RT: 10.439 min Scan# 1419
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tgt Ion:198 Resp: 5824
Ion Ratio Lower Upper
198 100
51 41.9 34.1 74.1
105 38.8 27.3 67.3




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#66
n-Nitrosodiphenylamine
Concen: 41.654 ng
RT: 10.486 min Scan# 1428
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:169 Resp: 346657
Ion Ratio Lower Upper
169 100
168 65.9 53.8 80.6
167 36.4 28.7 43.1





#67
4-Bromophenyl-phenylether
Concen: 42.182 ng
RT: 10.857 min Scan# 1491
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:248 Resp: 120010
Ion Ratio Lower Upper
248 100
250 97.2 77.8 116.6
141 69.4 59.2 88.8

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#68
Hexachlorobenzene
Concen: 42.927 ng
RT: 10.928 min Scan# 1503
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:284 Resp: 136111
Ion Ratio Lower Upper
284 100
142 23.6 23.4 35.0
249 25.1 22.3 33.5





#69

Atrazine

Concen: 38.465 ng

RT: 11.022 min Scan# 11

Delta R.T. 0.000 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tgt Ion:200 Resp: 8187
Ion Ratio Lower Upper
200 100
173 27.2 7.6 47.6
215 49.8 28.3 68.3

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#70

Pentachlorophenol

Concen: 38.429 ng




RT: 11.128 min Scan# 1537


Delta R.T. 0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16


Tgt Ion:266 Resp: 56795
Ion Ratio Lower Upper
266 100
268 67.2 52.6 78.8
264 62.8 49.0 73.4





#71
Phenanthrene
Concen: 39.910 ng
RT: 11.339 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040



Tgt Ion:178 Resp: 52464
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.4 11.9 17.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#72
Anthracene
Concen: 39.423 ng
RT: 11.392 min Scan# 1582
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:178 Resp: 524828
Ion Ratio Lower Upper
178 100
176 18.9 14.9 22.3
179 15.6 11.8 17.6






#73
 Carbazole
 Concen: 38.682 ng
 RT: 11.557 min Scan# 1610
 Delta R.T. 0.000 min
 Lab File: BF136967.D
 Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:167 Resp: 485529
 Ion Ratio Lower Upper
 167 100
 166 21.5 17.3 25.9
 139 11.7 9.8 14.6

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#74
 Di-n-butylphthalate
 Concen: 39.632 ng
 RT: 11.880 min Scan# 1665
 Delta R.T. -0.006 min
 Lab File: BF136967.D
 Acq: 05 Jan 2024 10:16




Tgt Ion:149 Resp: 607005
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.2
 104 4.3 3.6 5.4





#75
Fluoranthene
Concen: 37.119 ng
RT: 12.533 min Scan# 1776
Delta R.T. -0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:202 Resp: 521934
Ion Ratio Lower Upper
202 100
101 11.8 0.0 29.3
203 17.5 0.0 37.1

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.974 min Scan# 2021
Delta R.T. 0.006 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Tgt Ion:240 Resp: 133421
Ion Ratio Lower Upper
240 100
120 10.6 10.3 15.5
236 24.5 19.4 29.2





#77
Benzidine
Concen: 45.806 ng
RT: 12.663 min Scan# 1798
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040




Tgt Ion:184 Resp: 180200
Ion Ratio Lower Upper
184 100
185 13.7 11.2 16.8
183 10.7 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024




#78
Pyrene
Concen: 39.778 ng
RT: 12.769 min Scan# 1816
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:202 Resp: 520994
Ion Ratio Lower Upper
202 100
200 19.3 15.9 23.9
203 17.1 13.8 20.6





#79

Terphenyl-d14

Concen: 80.457 ng

RT: 12.910 min Scan# 1841

Delta R.T. -0.006 min

Lab File: BF136967.D


Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040



Tgt Ion:244 Resp: 768140




Ion	Ratio	Lower	Upper
244	100		
212	6.3	5.0	7.4
122	11.7	9.0	13.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#80

Butylbenzylphthalate

Concen: 40.484 ng

RT: 13.386 min Scan# 1921




Delta R.T. -0.006 min


Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Tgt Ion:149 Resp: 192939

Ion	Ratio	Lower	Upper
149	100		
91	70.4	54.1	81.1
206	21.2	17.1	25.7






#81
Benzo(a)anthracene
Concen: 40.459 ng
RT: 13.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:228 Resp: 374835
Ion Ratio Lower Upper
228 100
226 27.1 21.4 32.0
229 19.8 14.9 22.3

Manual Integrations


APPROVED

Reviewed By :Yogesh Patel 01/06/2024



Supervised By :mohammad ahmed 01/06/2024




#82
3,3'-Dichlorobenzidine
Concen: 43.411 ng
RT: 13.927 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16



Tgt Ion:252 Resp: 114218
Ion Ratio Lower Upper
252 100
254 61.2 52.2 78.4
126 10.7 8.0 12.0





#83

Chrysene

Concen: 39.314 ng

RT: 14.004 min Scan# 2026

Delta R.T. 0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:228 Resp: 356179

Ion Ratio Lower Upper

228 100

226 30.2 24.7 37.1





229 19.9 15.4 23.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#84

Bis(2-ethylhexyl)phthalate

Concen: 44.302 ng

RT: 13.945 min Scan# 2016

Delta R.T. -0.006 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16




Tgt Ion:149 Resp: 265645


Ion Ratio Lower Upper

149 100

167 26.0 20.9 31.3

279 3.7 2.9 4.3






#85
Di-n-octyl phthalate
Concen: 47.052 ng
RT: 14.568 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040






Tgt Ion:149 Resp: 436429
Ion Ratio Lower Upper
149 100
167 1.7 1.2 1.8
43 10.6 8.9 13.3

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#86
Perylene-d12
Concen: 20.000 ng
RT: 15.463 min Scan# 2274
Delta R.T. 0.018 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:264 Resp: 144248
Ion Ratio Lower Upper
264 100
260 23.0 18.3 27.5
265 21.7 17.0 25.6





#87
Indeno(1,2,3-cd)pyrene
Concen: 42.554 ng
RT: 16.986 min Scan# 21
Delta R.T. 0.035 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :





SSTDCCC040

Tgt Ion:276 Resp: 44319
Ion Ratio Lower Upper
276 100
138 22.8 18.7 28.1
277 25.1 20.2 30.2

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 01/06/2024
Supervised By :mohammad ahmed 01/06/2024



#88
Benzo(b)fluoranthene
Concen: 36.888 ng
RT: 15.027 min Scan# 2200
Delta R.T. 0.012 min
Lab File: BF136967.D
Acq: 05 Jan 2024 10:16

Tgt Ion:252 Resp: 355393
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 9.2 7.4 11.2





#89

Benzo(k)fluoranthene

Concen: 37.684 ng

RT: 15.057 min Scan# 21

Delta R.T. 0.012 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:252 Resp: 343060

Ion Ratio Lower Upper

252 100

253 22.1 17.0 25.6





125 10.0 8.2 12.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#90

Benzo(a)pyrene

Concen: 39.241 ng

RT: 15.398 min Scan# 2263

Delta R.T. 0.012 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16




Tgt Ion:252 Resp: 319127


Ion Ratio Lower Upper

252 100

253 22.3 17.0 25.6

125 10.3 8.8 13.2





#91

Dibenzo(a,h)anthracene

Concen: 42.545 ng

RT: 17.004 min Scan# 21

Delta R.T. 0.035 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tgt Ion:278 Resp: 36352

Ion Ratio Lower Upper

278 100

139 14.4 12.2 18.4





279 24.3 17.8 26.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 01/06/2024

Supervised By :mohammad ahmed 01/06/2024



#92

Benzo(g,h,i)perylene

Concen: 42.190 ng

RT: 17.451 min Scan# 2612

Delta R.T. 0.041 min

Lab File: BF136967.D

Acq: 05 Jan 2024 10:16

Tgt Ion:276 Resp: 369218

Ion Ratio Lower Upper

276 100

277 23.8 18.3 27.5

138 20.7 15.6 23.4

