

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
 Data File : BF111897.D
 Acq On : 8 Jan 2019 00:42
 Operator : JU/SJ
 Sample : K1012-03 2X
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EO-02-010419-B

Manual Integrations
 APPROVED

Sohil
 1/8/2019 11:20:52 AM

Quant Time: Jan 08 05:03:20 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 07 14:33:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	138095	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	385985	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	175699	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	339718	20.00	ng	0.00
76) Chrysene-d12	14.06	240	387219	20.00	ng	0.00
87) Perylene-d12	15.54	264	236868	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	413787	51.26	ng	0.00
7) Phenol-d6	6.53	99	482480	46.38	ng	0.00
23) Nitrobenzene-d5	7.45	82	280860	38.89	ng	-0.01
42) 2,4,6-Tribromophenol	10.72	330	122553	53.63	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	414568	34.41	ng	0.00
79) Terphenyl-d14	13.00	244	559385	27.42	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.63	163	33288	2.522	ng	97
71) Phenanthrene	11.44	178	140450	7.733	ng	99
72) Anthracene	11.49	178	37956	2.059	ng	97
75) Fluoranthene	12.63	202	256317	13.747	ng	99
78) Pyrene	12.86	202	275473	10.339	ng	98
81) Benzo(a)anthracene	14.04	228	139318	6.252	ng	96
83) Chrysene	14.08	228	110048	5.179	ng	97
86) Indeno(1,2,3-cd)pyrene	17.03	276	32184	2.078	ng	95
88) Benzo(b)fluoranthene	15.10	252	99034m	6.989	ng	
89) Benzo(k)fluoranthene	15.13	252	31838m	2.390	ng	
90) Benzo(a)pyrene	15.47	252	60994	4.896	ng	# 93
92) Benzo(a,h,i)perylene	17.48	276	28443	2.902	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
Data File : BF111897.D
Acq On : 8 Jan 2019 00:42
Operator : JU/SJ
Sample : K1012-03 2X
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EO-02-010419-B

Manual Integrations
APPROVED

Sohil
1/8/2019 11:20:52 AM

Quant Time: Jan 08 05:03:20 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 07 14:33:51 2019
Response via : Initial Calibration

