

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
 Data File : BF111900.D
 Acq On : 8 Jan 2019 2:04
 Operator : JU/SJ
 Sample : K1064-01 10X
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P-1419-S

Quant Time: Jan 08 05:09:50 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 07 14:33:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	113768	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	379678	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	158909	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	315047	20.00	ng	0.00
76) Chrysene-d12	14.06	240	291134	20.00	ng	0.00
87) Perylene-d12	15.54	264	214422	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	81261	12.22	ng	0.00
7) Phenol-d6	6.53	99	101001	11.78	ng	0.00
23) Nitrobenzene-d5	7.45	82	59274	8.34	ng	-0.01
42) 2,4,6-Tribromophenol	10.72	330	23617	11.43	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	100827	9.25	ng	0.00
79) Terphenyl-d14	13.00	244	125623	8.19	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.19	128	58685	3.256	ng	98
71) Phenanthrene	11.44	178	36178	2.148	ng	97
75) Fluoranthene	12.63	202	35590	2.058	ng	74
84) Bis(2-ethylhexyl)phthalate	14.03	149	44729	4.022	ng	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
Data File : BF111900.D
Acq On : 8 Jan 2019 2:04
Operator : JU/SJ
Sample : K1064-01 10X
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P-1419-S

Quant Time: Jan 08 05:09:50 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 07 14:33:51 2019
Response via : Initial Calibration

