

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
 Data File : BF111903.D
 Acq On : 8 Jan 2019 3:26
 Operator : JU/SJ
 Sample : J6426-10DL 10X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P001-SS013A-2430-01DL

Quant Time: Jan 08 06:08:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 07 14:33:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	104213	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	336527	20.00	ng	0.00
39) Acenaphthene-d10	9.93	164	156971	20.00	ng	0.00
64) Phenanthrene-d10	11.42	188	258817	20.00	ng	0.00
76) Chrysene-d12	14.06	240	271742	20.00	ng	0.00
87) Perylene-d12	15.55	264	203260	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	39309	6.45	ng	0.00
7) Phenol-d6	6.53	99	44158	5.62	ng	0.00
23) Nitrobenzene-d5	7.45	82	63268	10.05	ng	-0.01
42) 2,4,6-Tribromophenol	10.72	330	22650	11.09	ng	0.00
45) 2-Fluorobiphenyl	9.25	172	83039	7.71	ng	0.00
79) Terphenyl-d14	13.00	244	112239	7.84	ng	0.00
Target Compounds						
12) 1,3-Dichlorobenzene	6.83	146	19676	2.509	ng	98
13) 1,4-Dichlorobenzene	6.91	146	126778	15.933	ng	99
14) 1,2-Dichlorobenzene	7.06	146	45818	6.024	ng	96
30) 1,2,4-Trichlorobenzene	8.12	180	572688	99.555	ng	99
31) Naphthalene	8.20	128	596868	37.362	ng	100
37) 2-Methylnaphthalene	8.89	142	346753	35.030	ng	99
38) 1-Methylnaphthalene	8.99	142	211556	22.282	ng	98
40) 1,2,4,5-Tetrachlorobenzene	9.06	216	61090	12.208	ng	# 98
46) 1,1'-Biphenyl	9.35	154	90607	6.901	ng	98
71) Phenanthrene	11.44	178	156491	11.309	ng	97
84) Bis(2-ethylhexyl)phthalate	14.04	149	284490	27.404	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010719\
Data File : BF111903.D
Acq On : 8 Jan 2019 3:26
Operator : JU/SJ
Sample : J6426-10DL 10X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P001-SS013A-2430-01DL

Quant Time: Jan 08 06:08:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 07 14:33:51 2019
Response via : Initial Calibration

