

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010819\
 Data File : BF111920.D
 Acq On : 8 Jan 2019 21:12
 Operator : JU/SJ
 Sample : K1044-04
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 CPSB-11(20-25)

Manual Integrations
 APPROVED

Sohil
 1/9/2019 10:36:49 AM

Quant Time: Jan 09 01:25:57 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 07 14:33:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.89	152	121997	20.00	ng	0.00
21) Naphthalene-d8	8.17	136	352501	20.00	ng	0.00
39) Acenaphthene-d10	9.92	164	137012	20.00	ng	-0.01
64) Phenanthrene-d10	11.41	188	281390	20.00	ng	0.00
76) Chrysene-d12	14.05	240	255461	20.00	ng	0.00
87) Perylene-d12	15.54	264	175688	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	570016	79.94	ng	0.00
7) Phenol-d6	6.53	99	741050	80.63	ng	0.00
23) Nitrobenzene-d5	7.44	82	430456	65.26	ng	-0.02
42) 2,4,6-Tribromophenol	10.71	330	158753	89.08	ng	-0.01
45) 2-Fluorobiphenyl	9.24	172	644964	68.64	ng	-0.01
79) Terphenyl-d14	13.00	244	760766	56.53	ng	0.00
Target Compounds						
10) Phenol	6.54	94	43083	4.313	ng	Qvalue # 78
50) Dimethylphthalate	9.63	163	92515	8.989	ng	99
52) Acenaphthene	9.96	154	23024	2.686	ng	91
58) Fluorene	10.47	166	29142	3.212	ng	100
71) Phenanthrene	11.43	178	147984	9.836	ng	99
72) Anthracene	11.49	178	48316	3.164	ng	98
75) Fluoranthene	12.63	202	192344	12.454	ng	99
78) Pyrene	12.86	202	188497	10.724	ng	99
81) Benzo(a)anthracene	14.04	228	74614	5.075	ng	96
83) Chrysene	14.08	228	68543m	4.890	ng	
86) Indeno(1,2,3-cd)pyrene	17.03	276	20555	2.011	ng	95
88) Benzo(b)fluoranthene	15.10	252	53145m	5.057	ng	
90) Benzo(a)pyrene	15.47	252	37161	4.021	ng	# 79
92) Benzo(a,h,i)perylene	17.48	276	26251	3.611	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF010819\
Data File : BF111920.D
Acq On : 8 Jan 2019 21:12
Operator : JU/SJ
Sample : K1044-04
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
CPSB-11(20-25)

Manual Integrations
APPROVED

Sohil
1/9/2019 10:36:49 AM

Quant Time: Jan 09 01:25:57 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 07 14:33:51 2019
Response via : Initial Calibration

