


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF010919\
 Data File : BF111947.D
 Acq On : 9 Jan 2019 13:38
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF010719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 07 14:33:51 2019



AutoFind: Scans 1627, 1628, 1629; Background Corrected with Scan 1622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.8	41376	PASS
68	69	0.00	2	1.8	656	PASS
69	198	0.00	100	36.7	37157	PASS
70	69	0.00	2	0.3	116	PASS
127	198	10	80	46.4	46965	PASS
197	198	0.00	2	0.7	668	PASS
198	198	100	100	100.0	101309	PASS
199	198	5	9	7.3	7349	PASS
275	198	10	60	26.8	27113	PASS
365	198	1	100	4.0	4030	PASS
441	198	0.01	100	14.2	14392	PASS
442	198	50	100	88.9	90029	PASS
443	442	15	24	18.6	16741	PASS