


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF010919\
 Data File : BF111972.D
 Acq On : 10 Jan 2019 2:32
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF010719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 07 14:33:51 2019



AutoFind: Scans 1628, 1629, 1630; Background Corrected with Scan 1623

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	23399	PASS
68	69	0.00	2	1.4	297	PASS
69	198	0.00	100	34.8	20986	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	45.8	27596	PASS
197	198	0.00	2	1.1	669	PASS
198	198	100	100	100.0	60248	PASS
199	198	5	9	6.9	4131	PASS
275	198	10	60	27.0	16246	PASS
365	198	1	100	3.8	2293	PASS
441	198	0.01	100	14.6	8785	PASS
442	198	50	100	92.1	55506	PASS
443	442	15	24	19.4	10753	PASS