

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011019\  
 Data File : BF111986.D  
 Acq On : 10 Jan 2019 14:44  
 Operator : JU/SJ  
 Sample : K1082-11  
 Misc :  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 SP-1-6

Quant Time: Jan 11 01:08:17 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Jan 07 14:33:51 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.87  | 152  | 123900   | 20.00  | ng    | -0.02     |
| 21) Naphthalene-d8          | 8.16  | 136  | 469301   | 20.00  | ng    | -0.02     |
| 39) Acenaphthene-d10        | 9.91  | 164  | 236415   | 20.00  | ng    | -0.02     |
| 64) Phenanthrene-d10        | 11.40 | 188  | 435845   | 20.00  | ng    | -0.02     |
| 76) Chrysene-d12            | 14.04 | 240  | 300612   | 20.00  | ng    | -0.02     |
| 87) Perylene-d12            | 15.52 | 264  | 299007   | 20.00  | ng    | -0.02     |
| System Monitoring Compounds |       |      |          |        |       |           |
| 5) 2-Fluorophenol           | 5.51  | 112  | 701657   | 96.89  | ng    | -0.01     |
| 7) Phenol-d6                | 6.52  | 99   | 917461   | 98.29  | ng    | -0.02     |
| 23) Nitrobenzene-d5         | 7.43  | 82   | 576817   | 65.69  | ng    | -0.02     |
| 42) 2,4,6-Tribromophenol    | 10.70 | 330  | 282760   | 91.95  | ng    | -0.02     |
| 45) 2-Fluorobiphenyl        | 9.23  | 172  | 1060663  | 65.42  | ng    | -0.02     |
| 79) Terphenyl-d14           | 12.99 | 244  | 1045828  | 66.04  | ng    | -0.01     |
| Target Compounds            |       |      |          |        |       |           |
| 50) Dimethylphthalate       | 9.62  | 163  | 288538   | 16.247 | ng    | Qvalue 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011019\  
Data File : BF111986.D  
Acq On : 10 Jan 2019 14:44  
Operator : JU/SJ  
Sample : K1082-11  
Misc :  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
SP-1-6

Quant Time: Jan 11 01:08:17 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Mon Jan 07 14:33:51 2019  
Response via : Initial Calibration

