

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011019\
 Data File : BF112000.D
 Acq On : 10 Jan 2019 21:05
 Operator : JU/SJ
 Sample : K1062-01
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-(1-10)-COMP

Manual Integrations
 APPROVED

Sohil
 1/11/2019 4:42:58 PM

Quant Time: Jan 11 02:02:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 07 14:33:51 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	89302	20.00	ng	-0.02
21) Naphthalene-d8	8.16	136	305279	20.00	ng	-0.02
39) Acenaphthene-d10	9.91	164	129629	20.00	ng	-0.02
64) Phenanthrene-d10	11.40	188	243388	20.00	ng	-0.02
76) Chrysene-d12	14.04	240	234007	20.00	ng	-0.02
87) Perylene-d12	15.52	264	211549	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	444727	85.20	ng	-0.02
7) Phenol-d6	6.52	99	559047	83.10	ng	-0.02
23) Nitrobenzene-d5	7.43	82	340012	59.53	ng	-0.02
42) 2,4,6-Tribromophenol	10.70	330	128540	76.24	ng	-0.02
45) 2-Fluorobiphenyl	9.23	172	519954	58.49	ng	-0.02
79) Terphenyl-d14	12.98	244	552775	44.84	ng	-0.02
Target Compounds						
10) Phenol	6.53	94	24241	3.315	ng	Qvalue 81
50) Dimethylphthalate	9.62	163	50988	5.236	ng	99
71) Phenanthrene	11.42	178	224064	17.219	ng	97
72) Anthracene	11.47	178	42919	3.249	ng	98
73) Carbazole	11.63	167	28224	2.197	ng	98
75) Fluoranthene	12.62	202	355452	26.608	ng	99
78) Pyrene	12.84	202	300635	18.672	ng	98
80) Butylbenzylphthalate	13.45	149	22941	3.370	ng	99
81) Benzo(a)anthracene	14.03	228	157576	11.700	ng	98
83) Chrysene	14.07	228	157913	12.298	ng	96
84) Bis(2-ethylhexyl)phthalate	14.02	149	50813	5.684	ng	# 100
86) Indeno(1,2,3-cd)pyrene	17.00	276	54977	5.873	ng	97
88) Benzo(b)fluoranthene	15.09	252	207184m	16.372	ng	
89) Benzo(k)fluoranthene	15.11	252	61839m	5.199	ng	
90) Benzo(a)pyrene	15.46	252	110300	9.913	ng	# 96
92) Benzo(g,h,i)perylene	17.46	276	56188	6.419	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011019\
Data File : BF112000.D
Acq On : 10 Jan 2019 21:05
Operator : JU/SJ
Sample : K1062-01
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleID :
TP-(1-10)-COMP

Manual Integrations
APPROVED

Sohil
1/11/2019 4:42:58 PM

Quant Time: Jan 11 02:02:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 07 14:33:51 2019
Response via : Initial Calibration

