

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011119\
 Data File : BF112034.D
 Acq On : 11 Jan 2019 23:49
 Operator : JU/SJ
 Sample : K1094-07 10X
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :

Quant Time: Jan 12 01:00:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 11 21:37:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	6183	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	20479	20.00	ng	0.00
39) Acenaphthene-d10	9.91	164	10035	20.00	ng	0.00
64) Phenanthrene-d10	11.40	188	28510	20.00	ng	0.00
76) Chrysene-d12	14.04	240	16839	20.00	ng	0.00
87) Perylene-d12	15.52	264	15607	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.51	112	39261	108.63	ng	0.00
7) Phenol-d6	6.53	99	34120	73.25	ng	0.02
23) Nitrobenzene-d5	7.43	82	19544	51.01	ng	0.00
42) 2,4,6-Tribromophenol	10.70	330	8826	67.62	ng	0.00
45) 2-Fluorobiphenyl	9.23	172	40113	58.29	ng	0.00
79) Terphenyl-d14	12.98	244	73176	82.49	ng	0.00


Target Compounds				Qvalue		
4) n-Nitrosodimethylamine	3.34	42	9003	44.178	ng	# 1
6) Aniline	6.50	93	12312	21.860	ng	# 1
9) Benzaldehyde	6.42	77	379107	Below Cal		# 1
11) bis(2-Chloroethyl)ether	6.58	93	10034	25.619	ng	# 1
15) Benzyl Alcohol	6.95	79	221053	599.779	ng	# 6
17) 2-Methylphenol	7.13	107	1168	3.621	ng	# 1
18) Hexachloroethane	7.42	117	69912	424.141	ng	# 1
19) n-Nitroso-di-n-propylamine	7.43	70	2649	8.753	ng	# 87
20) 3+4-Methylphenols	7.27	107	1596	3.976	ng	# 1
22) Acetophenone	7.27	105	324317	625.534	ng	# 50
24) Nitrobenzene	7.42	77	61839	159.536	ng	# 21
25) Isophorone	7.43	82	19544	31.498	ng	# 65
31) Naphthalene	8.18	128	5240	5.390	ng	# 98
50) Dimethylphthalate	9.62	163	8863	11.757	ng	# 91
56) 4-Nitrophenol	10.02	139	391	3.065	ng	# 1
60) Diethylphthalate	10.32	149	2177	2.984	ng	# 27
62) 4-Nitroaniline	10.44	138	432	2.523	ng	# 92
66) n-Nitrosodiphenylamine	10.57	169	1993	2.083	ng	# 85
67) 4-Bromophenyl-phenylether	10.93	248	764	2.069	ng	# 1
70) Pentachlorophenol	11.20	266	759	3.259	ng	# 40
84) Bis(2-ethylhexyl)phthalate	14.02	149	3243	5.041	ng	# 51


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011119\
Data File : BF112034.D
Acq On : 11 Jan 2019 23:49
Operator : JU/SJ
Sample : K1094-07 10X
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :

Quant Time: Jan 12 01:00:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 11 21:37:26 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.88 min Scan# 815
Delta R.T. 0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Instrument :
BNA_F
ClientSampled :

Tot Ion:152 Resp: 6183
Ion Ratio Lower Upper
152 100
150 162.7 122.7 184.1
115 63.0 45.2 67.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#4


n-Nitrosodimethylamine
Concen: 44.178 ng
RT: 3.34 min Scan# 213
Delta R.T. -0.02 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt Ion: 42 Resp: 9003
Ion Ratio Lower Upper
42 100
74 0.0 87.0 130.6#
44 45.0 5.5 8.3#



Abundance Ion 42.00 (41.70 to 42.70): BF1
Ion 74.00 (73.70 to 74.70): BF1
Ion 44.00 (43.70 to 44.70): BF1









#5
 2-Fluorophenol
 Concen: 108.634 ng
 RT: 5.51 min Scan# 582
 Delta R.T. 0.01 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:112 Resp: 39261
 Ion Ratio Lower Upper
 112 100
 64 158.2 48.3 72.5#
 63 509.3 26.1 39.1#



#6
 Aniline
 Concen: 21.860 ng
 RT: 6.50 min Scan# 750
 Delta R.T. -0.04 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49

Tgt Ion: 93 Resp: 12312
 Ion Ratio Lower Upper
 93 100
 66 94.2 60.0 90.0#
 65 799.7 37.5 56.3#





#7

Phenol-d6

Concen: 73.249 ng

RT: 6.53 min Scan# 756

Delta R.T. 0.02 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampled :


Tot Ion: 99 Resp: 34120

Ion Ratio Lower Upper

99 100

42 18.8 17.9 26.9




71 35.7 28.7 43.1



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#9

Benzaldehyde

Concen: Below Cal

RT: 6.42 min Scan# 737

Delta R.T. -0.00 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion: 77 Resp: 379107

Ion Ratio Lower Upper

77 100

105 733.8 80.6 120.6#



106 67.3 74.6 114.6#




Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): E

Ion 106.00 (105.70 to 106.70): E









#11
bis(2-Chloroethyl)ether
Concen: 25.619 ng
RT: 6.58 min Scan# 764
Delta R.T. -0.03 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Instrument :
BNA_F
ClientSampleId :

Tot Ion: 93 Resp: 10034
Ion Ratio Lower Upper
93 100
63 999.3 65.2 105.2#
95 0.0 11.9 51.9#




Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 63.00 (62.70 to 63.70): BF1
Ion 95.00 (94.70 to 95.70): BF1






#15
Benzyl Alcohol
Concen: 599.779 ng
RT: 6.95 min Scan# 827
Delta R.T. -0.06 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt Ion: 79 Resp: 221053
Ion Ratio Lower Upper
79 100
108 0.2 61.4 92.0#
77 140.7 50.4 75.6#



Abundance Ion 79.00 (78.70 to 79.70): BF1
Ion 108.00 (107.70 to 108.70): E
Ion 77.00 (76.70 to 77.70): BF1





#17

2-Methylphenol

Concen: 3.621 ng

RT: 7.13 min Scan# 857

Delta R.T. -0.01 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampleId :

Tot Ion:107 Resp: 1168


Ion Ratio Lower Upper

107 100

108 0.0 89.7 134.5#

77 3986.1 40.6 60.8#

79 2153.2 40.2 60.2#




Abundance Ion 107.00 (106.70 to 107.70): E


Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1



Time-->



#18

Hexachloroethane

Concen: 424.141 ng

RT: 7.42 min Scan# 906

Delta R.T. 0.03 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion:117 Resp: 69912

Ion Ratio Lower Upper

117 100

119 1021.1 76.2 114.2#


201 0.0 75.7 113.5#






Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E







Time-->



#19
n-Nitroso-di-n-propylamine
Concen: 8.753 ng
RT: 7.43 min Scan# 909
Delta R.T. 0.15 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49


Instrument :
BNA_F
ClientSampleId :


Tot Ion:	70	Resp:	2649
Ion	Ratio	Lower	Upper
70	100		
42	65.8	53.6	80.4
101	0.0	7.3	10.9#
130	0.0	16.8	25.2#



#20
3+4-Methylphenols
Concen: 3.976 ng
RT: 7.27 min Scan# 882
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt Ion:	107	Resp:	1596
Ion	Ratio	Lower	Upper
107	100		
108	0.0	71.1	111.1#
77	2327.8	85.4	125.4#
79	1768.8	12.4	52.4#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.16 min Scan# 1032
Delta R.T. -0.00 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49




Instrument :
BNA_F
ClientSampleId :


Tot Ion:136	Resp:	20479
Ion Ratio	Lower	Upper
136 100		
137 12.0	9.1	13.7
54 7.3	7.2	10.8
68 4.9	4.4	6.6



#22
Acetophenone
Concen: 625.534 ng
RT: 7.27 min Scan# 882
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt	Ion:105	Resp:	324317
Ion	Ratio	Lower	Upper
105	100		
71	0.0	4.6	6.8#
51	4.3	32.2	48.2#
120	0.3	17.5	26.3#









#23
 Nitrobenzene-d5
 Concen: 51.005 ng
 RT: 7.43 min Scan# 909
 Delta R.T. -0.01 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49

Instrument :
 BNA_F
 ClientSampleID :

Tot Ion: 82 Resp: 19544
 Ion Ratio Lower Upper
 82 100
 128 41.4 35.7 53.5
 54 56.5 45.8 68.8




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#24
 Nitrobenzene
 Concen: 159.536 ng
 RT: 7.42 min Scan# 906
 Delta R.T. -0.04 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49

Tgt Ion: 77 Resp: 61839
 Ion Ratio Lower Upper
 77 100
 123 0.0 38.1 57.1#
 65 56.8 11.0 16.4#



Abundance Ion 77.00 (76.70 to 77.70): BF1
 Ion 123.00 (122.70 to 123.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 31.498 ng

RT: 7.43 min Scan# 909

Delta R.T. -0.26 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampleId :


Tot Ion: 82 Resp: 19544

Ion Ratio Lower Upper

82 100

95 0.0 5.8 8.6#




138 0.0 14.4 21.6#



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 95.00 (94.70 to 95.70): BF1

Ion 138.00 (137.70 to 138.70): BF1



#31

Naphthalene

Concen: 5.390 ng

RT: 8.18 min Scan# 1036

Delta R.T. -0.00 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion: 128 Resp: 5240

Ion Ratio Lower Upper

128 100

129 12.9 9.0 13.4



127 13.4 10.7 16.1




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.91 min Scan# 1330
 Delta R.T. -0.01 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:164 Resp: 10035
 Ion Ratio Lower Upper
 164 100
 162 108.8 82.2 123.2
 160 45.7 35.5 53.3



#42
 2,4,6-Tribromophenol
 Concen: 67.619 ng
 RT: 10.70 min Scan# 1465
 Delta R.T. -0.00 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49

Tgt Ion:330 Resp: 8826
 Ion Ratio Lower Upper
 330 100
 332 102.6 76.9 115.3
 141 38.0 23.7 35.5#









#45
2-Fluorobiphenyl
Concen: 58.289 ng
RT: 9.23 min Scan# 1214
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Instrument :
BNA_F
ClientSampleId :

Tot Ion:172 Resp: 40113
Ion Ratio Lower Upper
172 100
171 34.8 29.6 44.4
170 22.7 19.8 29.8




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 11.757 ng
RT: 9.62 min Scan# 1281
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt Ion:163 Resp: 8863
Ion Ratio Lower Upper
163 100
194 7.7 3.4 5.2#
164 13.6 8.5 12.7#



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#56
4-Nitrophenol
Concen: 3.065 ng
RT: 10.02 min Scan# 1349
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Instrument :
BNA_F
ClientSampled :

Tot Ion:139 Resp: 391
Ion Ratio Lower Upper
139 100
109 599.1 52.2 92.2#
65 77.4 83.1 123.1#




Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF1






#60
Diethylphthalate
Concen: 2.984 ng
RT: 10.32 min Scan# 1400
Delta R.T. -0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Tgt Ion:149 Resp: 2177
Ion Ratio Lower Upper
149 100
177 58.9 18.3 27.5#
150 40.6 10.2 15.4#



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E





#62

4-Nitroaniline

Concen: 2.523 ng

RT: 10.44 min Scan# 1420

Delta R.T. -0.04 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampleId :


Tot Ion:138 Resp: 432

Ion Ratio Lower Upper

138 100

92 38.6 31.0 71.0




108 73.1 54.7 94.7



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF1

Ion 108.00 (107.70 to 108.70): E



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.40 min Scan# 1584

Delta R.T. 0.01 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion:188 Resp: 28510

Ion Ratio Lower Upper

188 100

94 23.9 6.6 10.0#



80 23.9 7.7 11.5#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1





#66

n-Nitrosodiphenylamine

Concen: 2.083 ng

RT: 10.57 min Scan# 1442

Delta R.T. -0.00 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampleId :


Tot Ion:169 Resp: 1993

Ion Ratio Lower Upper

169 100

168 55.5 53.8 80.6




167 27.6 29.1 43.7#



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#67

4-Bromophenyl-phenylether

Concen: 2.069 ng

RT: 10.93 min Scan# 1504

Delta R.T. -0.01 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion:248 Resp: 764

Ion Ratio Lower Upper

248 100

250 96.7 76.8 115.2



141 426.0 57.0 85.6#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#70

Pentachlorophenol

Concen: 3.259 ng

RT: 11.20 min Scan# 1550

Delta R.T. -0.01 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49

Instrument :

BNA_F

ClientSampled :


Tot Ion:266 Resp: 759

Ion Ratio Lower Upper

266 100

268 0.0 50.6 76.0#




264 90.5 49.3 73.9#



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.04 min Scan# 2032

Delta R.T. -0.01 min

Lab File: BF112034.D

Acq: 11 Jan 2019 23:49


Tgt Ion:240 Resp: 16839

Ion Ratio Lower Upper

240 100

120 16.3 8.2 12.2#



236 30.0 20.2 30.4




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#79
 Terphenyl-d14
 Concen: 82.490 ng
 RT: 12.98 min Scan# 1852
 Delta R.T. -0.01 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49




Instrument :
 BNA_F
 ClientSampleId :


Tot Ion:244 Resp: 73176
 Ion Ratio Lower Upper
 244 100
 212 8.4 5.8 8.8
 122 14.2 9.5 14.3



#84
 Bis(2-ethylhexyl)phthalate
 Concen: 5.041 ng
 RT: 14.02 min Scan# 2028
 Delta R.T. -0.00 min
 Lab File: BF112034.D
 Acq: 11 Jan 2019 23:49

Tgt Ion:149 Resp: 3243
 Ion Ratio Lower Upper
 149 100
 167 54.4 21.5 32.3#
 279 10.7 3.0 4.6#






#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.52 min Scan# 2284
Delta R.T. 0.01 min
Lab File: BF112034.D
Acq: 11 Jan 2019 23:49

Instrument :
BNA_F
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
264	100		
260	22.8	19.2	28.8
265	24.1	17.1	25.7



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

