

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011419\
 Data File : BF112047.D
 Acq On : 14 Jan 2019 13:40
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Jan 14 15:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 11 21:37:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	158874	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	563367	20.00	ng	0.00
39) Acenaphthene-d10	9.91	164	293414	20.00	ng	0.00
64) Phenanthrene-d10	11.40	188	560208	20.00	ng	0.00
76) Chrysene-d12	14.04	240	455559	20.00	ng	0.00
87) Perylene-d12	15.52	264	334723	20.00	ng	0.00

System Monitoring Compounds

5) 2-Fluorophenol	5.50	112	732479	78.88	ng	0.00
7) Phenol-d6	6.52	99	875994	73.19	ng	0.00
23) Nitrobenzene-d5	7.43	82	990291	93.95	ng	0.00
42) 2,4,6-Tribromophenol	10.70	330	303968	79.65	ng	0.00
45) 2-Fluorobiphenyl	9.23	172	1615447	80.28	ng	0.00
79) Terphenyl-d14	12.98	244	2195789	91.49	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.66	88	206521	50.861	ng	99
3) Pyridine	3.42	79	554279	52.665	ng	98
4) n-Nitrosodimethylamine	3.37	42	256571	48.998	ng	96
6) Aniline	6.53	93	531728	36.741	ng	92
8) 2-Chlorophenol	6.66	128	392769	38.090	ng	97
9) Benzaldehyde	6.42	77	262389	39.098	ng	98
10) Phenol	6.53	94	479956	36.894	ng	93
11) bis(2-Chloroethyl)ether	6.60	93	391069	38.858	ng	97
12) 1,3-Dichlorobenzene	6.82	146	478609	40.037	ng	99
13) 1,4-Dichlorobenzene	6.89	146	520599	42.916	ng	98
14) 1,2-Dichlorobenzene	7.05	146	470413	40.568	ng	97
15) Benzyl Alcohol	7.02	79	367527	38.809	ng	99
16) 2,2'-oxybis(1-Chloropropan	7.15	45	764476	44.553	ng	98
17) 2-Methylphenol	7.13	107	380346	45.893	ng	95
18) Hexachloroethane	7.38	117	204628	48.314	ng	99
19) n-Nitroso-di-n-propylamine	7.29	70	324994	41.790	ng	97
20) 3+4-Methylphenols	7.29	107	432679	41.949	ng	# 89
22) Acetophenone	7.28	105	637786	44.717	ng	# 99
24) Nitrobenzene	7.45	77	486299	45.605	ng	96
25) Isophorone	7.69	82	635392	37.224	ng	100
26) 2-Nitrophenol	7.77	139	213580	40.679	ng	98
27) 2,4-Dimethylphenol	7.81	122	306731	40.405	ng	99
28) bis(2-Chloroethoxy)methane	7.90	93	454950	39.931	ng	100
29) 2,4-Dichlorophenol	8.02	162	352450	40.376	ng	98
30) 1,2,4-Trichlorobenzene	8.10	180	392606	40.769	ng	99
31) Naphthalene	8.17	128	1081747	40.449	ng	100
32) Benzoic acid	7.95	122	171234	34.759	ng	96
33) 4-Chloroaniline	8.22	127	461067	39.882	ng	96
34) Hexachlorobutadiene	8.29	225	244736	40.823	ng	99
35) Caprolactam	8.60	113	117017	41.029	ng	95
36) 4-Chloro-3-methylphenol	8.72	107	352243	40.585	ng	98
37) 2-Methylnaphthalene	8.87	142	712861	43.018	ng	99
38) 1-Methylnaphthalene	8.97	142	690547	43.447	ng	97
40) 1,2,4,5-Tetrachlorobenzene	9.04	216	387851	41.463	ng	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011419\
 Data File : BF112047.D
 Acq On : 14 Jan 2019 13:40
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Quant Time: Jan 14 15:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF010719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 11 21:37:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Hexachlorocyclopentadiene	9.02	237	233292	43.134	nq	100
43) 2,4,6-Trichlorophenol	9.15	196	251761	39.681	nq	99
44) 2,4,5-Trichlorophenol	9.20	196	263715	39.672	nq	99
46) 1,1'-Biphenyl	9.33	154	992474	40.437	nq	99
47) 2-Chloronaphthalene	9.36	162	773473	40.154	nq	97
48) 2-Nitroaniline	9.45	65	236519	40.161	nq	97
49) Acenaphthylene	9.77	152	1138404	39.974	nq	99
50) Dimethylphthalate	9.63	163	887636	40.271	nq	99
51) 2,6-Dinitrotoluene	9.69	165	198205	40.213	nq	97
52) Acenaphthene	9.95	154	679378	37.005	nq	99
53) 3-Nitroaniline	9.86	138	218054	41.311	nq	99
54) 2,4-Dinitrophenol	9.97	184	76446	36.628	nq	# 84
55) Dibenzofuran	10.12	168	1063050	39.814	nq	97
56) 4-Nitrophenol	10.04	139	137483	36.862	nq	95
57) 2,4-Dinitrotoluene	10.10	165	263687	41.402	nq	94
58) Fluorene	10.46	166	779113	40.102	nq	98
59) 2,3,4,6-Tetrachlorophenol	10.24	232	219635	39.215	nq	99
60) Diethylphthalate	10.33	149	862981	40.458	nq	100
61) 4-Chlorophenyl-phenylether	10.45	204	431367	39.594	nq	96
62) 4-Nitroaniline	10.48	138	212581	42.460	nq	98
63) Azobenzene	10.61	77	826129	40.543	nq	97
65) 4,6-Dinitro-2-methylphenol	10.52	198	134010	38.769	nq	98
66) n-Nitrosodiphenylamine	10.57	169	754534	40.136	nq	99
67) 4-Bromophenyl-phenylether	10.94	248	279897	38.575	nq	97
68) Hexachlorobenzene	11.02	284	315102	39.184	nq	99
69) Atrazine	11.10	200	244040	36.948	nq	98
70) Pentachlorophenol	11.21	266	181601	39.685	nq	100
71) Phenanthrene	11.42	178	1167270	38.972	nq	100
72) Anthracene	11.47	178	1172045	38.548	nq	99
73) Carbazole	11.63	167	1144557	38.700	nq	100
74) Di-n-butylphthalate	11.95	149	1334813	38.855	nq	100
75) Fluoranthene	12.61	202	1180222	38.384	nq	98
77) Benzidine	12.73	184	489600	35.968	nq	98
78) Pyrene	12.84	202	1211851	38.661	nq	98
80) Butylbenzylphthalate	13.45	149	668332	50.433	nq	97
81) Benzo(a)anthracene	14.03	228	1066153	40.665	nq	100
82) 3,3'-Dichlorobenzidine	13.99	252	405417	41.885	nq	99
83) Chrysene	14.07	228	991282	39.654	nq	99
84) Bis(2-ethylhexyl)phthalate	14.01	149	777086	44.651	nq	100
85) Di-n-octyl phthalate	14.62	149	1062777	42.672	nq	98
86) Indeno(1,2,3-cd)pyrene	17.01	276	720668	39.543	nq	99
88) Benzo(b)fluoranthene	15.09	252	792456	39.577	nq	98
89) Benzo(k)fluoranthene	15.12	252	766695	40.736	nq	99
90) Benzo(a)pyrene	15.46	252	706943	40.155	nq	99
91) Dibenzo(a,h)anthracene	17.02	278	608044	42.941	nq	99
92) Benzo(g,h,i)perylene	17.46	276	589884	42.592	nq	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.87 min Scan# 813
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:152 Resp: 158874
Ion Ratio Lower Upper
152 100
150 153.7 122.7 184.1
115 58.1 45.2 67.8




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 50.861 ng
RT: 2.66 min Scan# 97
Delta R.T. 0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion: 88 Resp: 206521
Ion Ratio Lower Upper
88 100
58 83.4 67.6 101.4
43 35.4 29.0 43.6



Abundance Ion 88.00 (87.70 to 88.70): BF1
Ion 58.00 (57.70 to 58.70): BF1
Ion 43.00 (42.70 to 43.70): BF1





#3

Pvridine

Concen: 52.665 ng

RT: 3.42 min Scan# 226

Delta R.T. 0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 79 Resp: 554279

Ion Ratio Lower Upper

79 100

52 85.5 67.0 100.6




51 43.4 34.2 51.2



Abundance Ion 79.00 (78.70 to 79.70): BF1

Ion 52.00 (51.70 to 52.70): BF1

Ion 51.00 (50.70 to 51.70): BF1



#4

n-Nitrosodimethylamine

Concen: 48.998 ng

RT: 3.37 min Scan# 218

Delta R.T. 0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 42 Resp: 256571

Ion Ratio Lower Upper

42 100

74 113.0 87.0 130.6



44 6.4 5.5 8.3




Abundance Ion 42.00 (41.70 to 42.70): BF1

Ion 74.00 (73.70 to 74.70): BF1

Ion 44.00 (43.70 to 44.70): BF1





#5

2-Fluorophenol

Concen: 78.876 ng

RT: 5.50 min Scan# 581

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:112 Resp: 732479

Ion Ratio Lower Upper

112 100

64 61.9 48.3 72.5




63 33.6 26.1 39.1



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1



#6

Aniline

Concen: 36.741 ng

RT: 6.53 min Scan# 756

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 93 Resp: 531728

Ion Ratio Lower Upper

93 100

66 82.1 60.0 90.0



65 51.4 37.5 56.3




Abundance Ion 93.00 (92.70 to 93.70): BF1

Ion 66.00 (65.70 to 66.70): BF1

Ion 65.00 (64.70 to 65.70): BF1





#7

Phenol-d6

Concen: 73.188 ng

RT: 6.52 min Scan# 753

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 99 Resp: 875994

Ion Ratio Lower Upper

99 100

42 22.2 17.9 26.9




71 37.3 28.7 43.1



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#8

2-Chlorophenol

Concen: 38.090 ng

RT: 6.66 min Scan# 778

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:128 Resp: 392769

Ion Ratio Lower Upper

128 100

130 32.8 12.7 52.7



64 48.4 31.8 71.8




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 64.00 (63.70 to 64.70): BF1





#9

Benzaldehyde

Concen: 39.098 ng

RT: 6.42 min Scan# 736

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 77 Resp: 262389

Ion Ratio Lower Upper

77 100

105 97.6 80.6 120.6




106 93.0 74.6 114.6



Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 105.00 (104.70 to 105.70): BF1

Ion 106.00 (105.70 to 106.70): BF1



#10

Phenol

Concen: 36.894 ng

RT: 6.53 min Scan# 756

Delta R.T. 0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 94 Resp: 479956

Ion Ratio Lower Upper

94 100

65 44.3 22.8 62.8



66 70.8 43.3 83.3




Abundance Ion 94.00 (93.70 to 94.70): BF1

Ion 65.00 (64.70 to 65.70): BF1

Ion 66.00 (65.70 to 66.70): BF1





#11

bis(2-Chloroethyl)ether

Concen: 38.858 ng

RT: 6.60 min Scan# 768

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion: 93 Resp: 391069

Ion Ratio Lower Upper




93 100

63 82.1 65.2 105.2

95 32.1 11.9 51.9



Abundance Ion 93.00 (92.70 to 93.70): BF1
Ion 63.00 (62.70 to 63.70): BF1
Ion 95.00 (94.70 to 95.70): BF1



#12

1,3-Dichlorobenzene

Concen: 40.037 ng

RT: 6.82 min Scan# 804

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 146 Resp: 478609

Ion Ratio Lower Upper



146 100


148 64.9 51.8 77.6

75 28.2 24.3 36.5



Abundance Ion 146.00 (145.70 to 146.70): E
Ion 148.00 (147.70 to 148.70): E
Ion 75.00 (74.70 to 75.70): BF1









#13
 1,4-Dichlorobenzene
 Concen: 42.916 ng
 RT: 6.89 min Scan# 816
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:146 Resp: 520599
 Ion Ratio Lower Upper
 146 100
 148 62.9 51.2 76.8
 111 41.2 31.5 47.3




Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E








#14
 1,2-Dichlorobenzene
 Concen: 40.568 ng
 RT: 7.05 min Scan# 843
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:146 Resp: 470413
 Ion Ratio Lower Upper
 146 100
 148 65.4 51.5 77.3
 111 41.4 35.5 53.3



Abundance Ion 146.00 (145.70 to 146.70): E
 Ion 148.00 (147.70 to 148.70): E
 Ion 111.00 (110.70 to 111.70): E





#15

Benzyl Alcohol

Concen: 38.809 ng

RT: 7.02 min Scan# 838

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion: 79 Resp: 367527

Ion Ratio Lower Upper

79 100

108 76.6 61.4 92.0

77 64.1 50.4 75.6



#16

2,2'-oxybis(1-Chloropropane)

Concen: 44.553 ng

RT: 7.15 min Scan# 860

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 45 Resp: 764476


Ion Ratio Lower Upper

45 100

77 16.3 0.0 37.1

79 13.2 0.0 33.9









#17
2-Methylphenol
Concen: 45.893 ng
RT: 7.13 min Scan# 858
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:107 Resp: 380346
Ion Ratio Lower Upper
107 100
108 114.3 89.7 134.5
77 58.0 40.6 60.8
79 53.3 40.2 60.2



#18
Hexachloroethane
Concen: 48.314 ng
RT: 7.38 min Scan# 900
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:117 Resp: 204628
Ion Ratio Lower Upper
117 100
119 95.8 76.2 114.2
201 96.1 75.7 113.5






#19
n-Nitroso-di-n-propylamine
Concen: 41.790 ng
RT: 7.29 min Scan# 884
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:	70	Resp:	324994
Ion	Ratio	Lower	Upper
70	100		
42	64.0	53.6	80.4
101	10.1	7.3	10.9
130	21.2	16.8	25.2






Abundance

Ion 70.00 (69.70 to 70.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 101.00 (100.70 to 101.70): BF1

Ion 130.00 (129.70 to 130.70): BF1



#20
3+4-Methylphenols
Concen: 41.949 ng
RT: 7.29 min Scan# 884
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:	107	Resp:	432679
Ion	Ratio	Lower	Upper
107	100		
108	92.9	71.1	111.1
77	82.9	85.4	125.4#
79	32.1	12.4	52.4




Abundance


Ion 107.00 (106.70 to 107.70): BF1

Ion 108.00 (107.70 to 108.70): BF1

Ion 77.00 (76.70 to 77.70): BF1

Ion 79.00 (78.70 to 79.70): BF1





#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.16 min Scan# 1032

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:136 Resp: 563367



Ion Ratio Lower Upper

136 100

137 10.6 9.1 13.7

54 7.4 7.2 10.8

68 4.4 4.4 6.6





Abundance

Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1



#22

Acetophenone

Concen: 44.717 ng

RT: 7.28 min Scan# 883

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Tgt Ion:105 Resp: 637786



Ion Ratio Lower Upper

105 100

71 7.5 4.6 6.8#

51 39.8 32.2 48.2

120 22.2 17.5 26.3




Abundance


Ion 105.00 (104.70 to 105.70): E

Ion 71.00 (70.70 to 71.70): BF1

Ion 51.00 (50.70 to 51.70): BF1

Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 93.947 ng
 RT: 7.43 min Scan# 909
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion: 82 Resp: 990291
 Ion Ratio Lower Upper
 82 100
 128 43.7 35.7 53.5
 54 56.1 45.8 68.8




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#24
 Nitrobenzene
 Concen: 45.605 ng
 RT: 7.45 min Scan# 912
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion: 77 Resp: 486299
 Ion Ratio Lower Upper
 77 100
 123 44.4 38.1 57.1
 65 14.3 11.0 16.4



Abundance Ion 77.00 (76.70 to 77.70): BF1
 Ion 123.00 (122.70 to 123.70): BF1
 Ion 65.00 (64.70 to 65.70): BF1





#25

Isophorone

Concen: 37.224 ng

RT: 7.69 min Scan# 953

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040





Tot Ion: 82 Resp: 635392

Ion Ratio Lower Upper

82 100

95 7.5 5.8 8.6

138 18.1 14.4 21.6



#26

2-Nitrophenol

Concen: 40.679 ng

RT: 7.77 min Scan# 966

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40




Tgt Ion: 139 Resp: 213580


Ion Ratio Lower Upper

139 100

109 31.4 24.1 36.1

65 53.4 43.6 65.4









#27
2,4-Dimethylphenol
Concen: 40.405 ng
RT: 7.81 min Scan# 973
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:122 Resp: 306731
Ion Ratio Lower Upper
122 100
107 115.3 91.0 136.6
121 58.4 46.8 70.2



#28
bis(2-Chloroethoxy)methane
Concen: 39.931 ng
RT: 7.90 min Scan# 988
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion: 93 Resp: 454950
Ion Ratio Lower Upper
93 100
95 32.3 25.8 38.6
123 11.5 9.5 14.3









#29
 2,4-Dichlorophenol
 Concen: 40.376 ng
 RT: 8.02 min Scan# 1008
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:162 Resp: 352450
 Ion Ratio Lower Upper
 162 100
 164 63.6 42.1 82.1
 98 35.9 17.2 57.2




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): BF1






#30
 1,2,4-Trichlorobenzene
 Concen: 40.769 ng
 RT: 8.10 min Scan# 1022
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:180 Resp: 392606
 Ion Ratio Lower Upper
 180 100
 182 96.2 76.2 114.4
 145 29.2 23.4 35.2



Abundance Ion 180.00 (179.70 to 180.70): E
 Ion 182.00 (181.70 to 182.70): E
 Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 40.449 ng
RT: 8.17 min Scan# 1035
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:128 Resp: 1081747
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.4
127 13.5 10.7 16.1



#32
Benzoic acid
Concen: 34.759 ng
RT: 7.95 min Scan# 996
Delta R.T. 0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:122 Resp: 171234
Ion Ratio Lower Upper
122 100
105 125.1 107.9 147.9
77 91.6 76.6 116.6









#33
4-Chloroaniline
Concen: 39.882 ng
RT: 8.22 min Scan# 1043
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:	127	Resp:	461067
Ion	Ratio	Lower	Upper
127	100		
129	32.7	25.8	38.6
65	32.6	23.5	35.3
92	19.5	14.2	21.2



#34
Hexachlorobutadiene
Concen: 40.823 ng
RT: 8.29 min Scan# 1055
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:	225	Resp:	244736
Ion	Ratio	Lower	Upper
225	100		
223	63.2	50.9	76.3
227	62.7	51.4	77.0









#35
Caprolactam
Concen: 41.029 ng
RT: 8.60 min Scan# 1108
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:113 Resp: 117017
Ion Ratio Lower Upper
113 100
55 173.1 162.0 202.0
56 132.7 116.0 156.0



#36
4-Chloro-3-methylphenol
Concen: 40.585 ng
RT: 8.72 min Scan# 1127
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:107 Resp: 352243
Ion Ratio Lower Upper
107 100
144 25.4 20.7 31.1
142 77.4 63.9 95.9









#37
 2-Methylnaphthalene
 Concen: 43.018 ng
 RT: 8.87 min Scan# 1153
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40




Instrument :
 BNA_F
 ClientSampleId :
 SSTCCCC040


Tot Ion:142 Resp: 712861
 Ion Ratio Lower Upper
 142 100
 141 90.1 71.4 107.2
 115 36.1 28.3 42.5



#38
 1-Methylnaphthalene
 Concen: 43.447 ng
 RT: 8.97 min Scan# 1170
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:142 Resp: 690547
 Ion Ratio Lower Upper
 142 100
 141 90.2 74.8 112.2
 116 3.9 3.1 4.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.91 min Scan# 1330

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:164 Resp: 293414

Ion Ratio Lower Upper

164 100

162 101.3 82.2 123.2




160 44.1 35.5 53.3



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#40

1,2,4,5-Tetrachlorobenzene

Concen: 41.463 ng

RT: 9.04 min Scan# 1182

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Tgt Ion:216 Resp: 387851


Ion Ratio Lower Upper

216 100

214 80.1 63.5 95.3

179 21.7 17.0 25.6

108 18.1 14.6 22.0






Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 108.00 (107.70 to 108.70): E









#41
Hexachlorocyclopentadiene
Concen: 43.134 ng
RT: 9.02 min Scan# 1179
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion	Ratio	Lower	Upper
237	100		
235	62.6	42.6	82.6
272	11.8	0.0	32.3




Abundance Ion 236.80 (236.50 to 237.50): E
Ion 234.90 (234.60 to 235.60): E
Ion 271.75 (271.45 to 272.45): E






#42
2,4,6-Tribromophenol
Concen: 79.647 ng
RT: 10.70 min Scan# 1465
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.3	76.9	115.3
141	29.2	23.7	35.5



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#43
 2,4,6-Trichlorophenol
 Concen: 39.681 ng
 RT: 9.15 min Scan# 1201
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:196 Resp: 251761
 Ion Ratio Lower Upper
 196 100
 198 96.1 75.8 113.8
 200 30.4 24.5 36.7




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 200.00 (199.70 to 200.70): E






#44
 2,4,5-Trichlorophenol
 Concen: 39.672 ng
 RT: 9.20 min Scan# 1209
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:196 Resp: 263715
 Ion Ratio Lower Upper
 196 100
 198 95.1 76.2 114.2
 97 44.9 36.4 54.6
 132 25.8 21.7 32.5



Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF1
 Ion 132.00 (131.70 to 132.70): E





#45

2-Fluorobiphenyl

Concen: 80.284 ng

RT: 9.23 min Scan# 1214

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :


BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion:172 Resp: 1615447

Ion	Ratio	Lower	Upper
172	100		
171	36.7	29.6	44.4
170	24.0	19.8	29.8






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#46

1,1'-Biphenyl

Concen: 40.437 ng

RT: 9.33 min Scan# 1232


Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Tgt Ion:154 Resp: 992474

Ion	Ratio	Lower	Upper
154	100		
153	40.9	21.3	61.3
76	10.9	0.0	32.3






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF1









#47
 2-Chloronaphthalene
 Concen: 40.154 ng
 RT: 9.36 min Scan# 1236
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:162 Resp: 773473
 Ion Ratio Lower Upper
 162 100
 127 41.3 31.2 46.8
 164 33.2 27.0 40.6




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 127.00 (126.70 to 127.70): E
 Ion 164.00 (163.70 to 164.70): E






#48
 2-Nitroaniline
 Concen: 40.161 ng
 RT: 9.45 min Scan# 1252
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion: 65 Resp: 236519
 Ion Ratio Lower Upper
 65 100
 92 66.6 52.5 78.7
 138 103.3 79.8 119.6



Abundance Ion 65.00 (64.70 to 65.70): BF1
 Ion 92.00 (91.70 to 92.70): BF1
 Ion 138.00 (137.70 to 138.70): E





#49

Acenaphthylene

Concen: 39.974 ng

RT: 9.77 min Scan# 1307

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:152 Resp: 1138404

Ion Ratio Lower Upper

152 100

151 20.5 16.6 25.0




153 13.1 11.0 16.4



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#50

Dimethylphthalate

Concen: 40.271 ng

RT: 9.63 min Scan# 1283

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:163 Resp: 887636

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.2



164 10.0 8.5 12.7




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#51

2,6-Dinitrotoluene

Concen: 40.213 ng

RT: 9.69 min Scan# 1293

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:165 Resp: 198205

Ion Ratio Lower Upper

165 100

63 59.4 46.2 69.2




89 55.0 42.2 63.2



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF1

Ion 89.00 (88.70 to 89.70): BF1



#52

Acenaphthene

Concen: 37.005 ng

RT: 9.95 min Scan# 1336

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:154 Resp: 679378

Ion Ratio Lower Upper

154 100

153 111.8 89.4 134.0



152 52.8 43.1 64.7




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E









#53
3-Nitroaniline
Concen: 41.311 ng
RT: 9.86 min Scan# 1322
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:138 Resp: 218054
Ion Ratio Lower Upper
138 100
108 10.7 8.2 12.4
92 118.3 93.5 140.3




Abundance Ion 138.00 (137.70 to 138.70): E
Ion 108.00 (107.70 to 108.70): E
Ion 92.00 (91.70 to 92.70): BF1






#54
2,4-Dinitrophenol
Concen: 36.628 ng
RT: 9.97 min Scan# 1340
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:184 Resp: 76446
Ion Ratio Lower Upper
184 100
63 76.0 46.9 70.3#
154 64.6 46.9 70.3



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 154.00 (153.70 to 154.70): E





#55

Dibenzofuran

Concen: 39.814 ng

RT: 10.12 min Scan# 1365

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:168 Resp: 1063050

Ion Ratio Lower Upper

168 100

139 38.9 29.3 43.9




169 14.0 10.6 16.0



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E



#56

4-Nitrophenol

Concen: 36.862 ng

RT: 10.04 min Scan# 1352

Delta R.T. 0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:139 Resp: 137483

Ion Ratio Lower Upper

139 100

109 71.2 52.2 92.2



65 96.1 83.1 123.1




Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF1









#57
2,4-Dinitrotoluene
Concen: 41.402 ng
RT: 10.10 min Scan# 1362
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:	165	Resp:	263687
Ion	Ratio	Lower	Upper
165	100		
63	47.1	34.1	51.1
89	71.2	53.5	80.3
182	2.5	2.0	3.0




Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF1
Ion 89.00 (88.70 to 89.70): BF1
Ion 182.00 (181.70 to 182.70): E






#58
Fluorene
Concen: 40.102 ng
RT: 10.46 min Scan# 1424
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:	166	Resp:	779113
Ion	Ratio	Lower	Upper
166	100		
165	97.4	79.7	119.5
167	13.1	11.1	16.7



Abundance Ion 166.00 (165.70 to 166.70): E
Ion 165.00 (164.70 to 165.70): E
Ion 167.00 (166.70 to 167.70): E









#59
 2,3,4,6-Tetrachlorophenol
 Concen: 39.215 ng
 RT: 10.24 min Scan# 1386
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:	232	Resp:	219635
Ion	Ratio	Lower	Upper
232	100		
131	41.5	34.0	51.0
130	2.2	1.6	2.4
166	29.1	23.4	35.0




Abundance Ion 232.00 (231.70 to 232.70): E
 Ion 131.00 (130.70 to 131.70): E
 Ion 130.00 (129.70 to 130.70): E
 Ion 166.00 (165.70 to 166.70): E






#60
 Diethylphthalate
 Concen: 40.458 ng
 RT: 10.33 min Scan# 1401
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:	149	Resp:	862981
Ion	Ratio	Lower	Upper
149	100		
177	23.1	18.3	27.5
150	12.8	10.2	15.4



Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 177.00 (176.70 to 177.70): E
 Ion 150.00 (149.70 to 150.70): E









#61
 4-Chlorophenyl-phenylether
 Concen: 39.594 ng
 RT: 10.45 min Scan# 1422
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40




Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040


Tot Ion	Ratio	Lower	Upper
204	100		
206	34.5	26.9	40.3
141	56.4	47.9	71.9



#62
 4-Nitroaniline
 Concen: 42.460 ng
 RT: 10.48 min Scan# 1427
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion	Ratio	Lower	Upper
138	100		
92	52.0	31.0	71.0
108	77.4	54.7	94.7





#63

Azobenzene

Concen: 40.543 ng

RT: 10.61 min Scan# 1449

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040

Tot Ion: 77 Resp: 826129


Ion Ratio Lower Upper

77 100

182 27.0 5.7 45.7

105 21.2 0.8 40.8

51 33.2 15.6 55.6






Abundance Ion 77.00 (76.70 to 77.70): BF1

Ion 182.00 (181.70 to 182.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 51.00 (50.70 to 51.70): BF1



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.40 min Scan# 1583

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion: 188 Resp: 560208

Ion Ratio Lower Upper

188 100

94 8.2 6.6 10.0



80 9.0 7.7 11.5




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1









#65
4,6-Dinitro-2-methylphenol
Concen: 38.769 ng
RT: 10.52 min Scan# 1433
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:198 Resp: 134010
Ion Ratio Lower Upper
198 100
51 44.7 27.2 67.2
105 40.4 21.0 61.0



#66
n-Nitrosodiphenylamine
Concen: 40.136 ng
RT: 10.57 min Scan# 1442
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:169 Resp: 754534
Ion Ratio Lower Upper
169 100
168 66.5 53.8 80.6
167 36.0 29.1 43.7









#67
 4-Bromophenyl-phenylether
 Concen: 38.575 ng
 RT: 10.94 min Scan# 1505
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:248 Resp: 279897
 Ion Ratio Lower Upper
 248 100
 250 98.2 76.8 115.2
 141 67.7 57.0 85.6




Abundance Ion 248.00 (247.70 to 248.70): E
 Ion 250.00 (249.70 to 250.70): E
 Ion 141.00 (140.70 to 141.70): E






#68
 Hexachlorobenzene
 Concen: 39.184 ng
 RT: 11.02 min Scan# 1518
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:284 Resp: 315102
 Ion Ratio Lower Upper
 284 100
 142 27.0 21.8 32.6
 249 28.6 23.8 35.6



Abundance Ion 283.80 (283.50 to 284.50): E
 Ion 142.00 (141.70 to 142.70): E
 Ion 249.00 (248.70 to 249.70): E





#69

Atrazine

Concen: 36.948 ng

RT: 11.10 min Scan# 1532

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:200 Resp: 244040

Ion Ratio Lower Upper

200 100

173 25.7 6.6 46.6




215 47.4 28.9 68.9



Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



#70

Pentachlorophenol

Concen: 39.685 ng

RT: 11.21 min Scan# 1551

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:266 Resp: 181601

Ion Ratio Lower Upper

266 100

268 62.8 50.6 76.0



264 61.7 49.3 73.9




Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E









#71
Phenanthrene
Concen: 38.972 ng
RT: 11.42 min Scan# 1587
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:178 Resp: 1167270
Ion Ratio Lower Upper
178 100
176 19.3 15.6 23.4
179 15.6 12.6 18.8



#72
Anthracene
Concen: 38.548 ng
RT: 11.47 min Scan# 1596
Delta R.T. -0.01 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:178 Resp: 1172045
Ion Ratio Lower Upper
178 100
176 19.4 15.7 23.5
179 15.8 12.4 18.6









#73
 Carbazole
 Concen: 38.700 ng
 RT: 11.63 min Scan# 1622
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:167 Resp: 1144557




Ion	Ratio	Lower	Upper
167	100		
166	22.4	17.8	26.6
139	13.1	10.6	16.0




#74
 Di-n-butylphthalate
 Concen: 38.855 ng
 RT: 11.95 min Scan# 1677
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:149 Resp: 1334813

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.5	11.3
104	5.0	3.9	5.9









#75
 Fluoranthene
 Concen: 38.384 ng
 RT: 12.61 min Scan# 1789
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:202 Resp: 1180222




Ion	Ratio	Lower	Upper
202	100		
101	11.0	0.0	29.3
203	17.9	0.0	37.6




#76
 Chrysene-d12
 Concen: 20.000 ng
 RT: 14.04 min Scan# 2033
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:240 Resp: 455559

Ion	Ratio	Lower	Upper
240	100		
120	8.2	8.2	12.2
236	25.2	20.2	30.4





#77

Benzidine

Concen: 35.968 ng

RT: 12.73 min Scan# 1809

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:184 Resp: 489600

Ion Ratio Lower Upper




184 100

185 15.4 13.4 20.0

183 11.7 9.6 14.4



Abundance Ion 184.00 (183.70 to 184.70): E
Ion 185.00 (184.70 to 185.70): E
Ion 183.00 (182.70 to 183.70): E



#78

Pyrene

Concen: 38.661 ng

RT: 12.84 min Scan# 1829

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:202 Resp: 1211851

Ion Ratio Lower Upper



202 100


200 21.2 17.4 26.2

203 17.0 14.6 21.8



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#79
 Terphenyl-d14
 Concen: 91.494 ng
 RT: 12.98 min Scan# 1852
 Delta R.T. -0.01 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Instrument :
 BNA_F
 ClientSampleId :
 SSTDCCC040

Tot Ion:244 Resp: 2195789




Ion	Ratio	Lower	Upper
244	100		
212	7.3	5.8	8.8
122	12.0	9.5	14.3




#80
 Butylbenzylphthalate
 Concen: 50.433 ng
 RT: 13.45 min Scan# 1932
 Delta R.T. -0.00 min
 Lab File: BF112047.D
 Acq: 14 Jan 2019 13:40

Tgt Ion:149 Resp: 668332

Ion	Ratio	Lower	Upper
149	100		
91	68.4	57.2	85.8
206	21.8	17.3	25.9





#81

Benzo(a)anthracene

Concen: 40.665 ng

RT: 14.03 min Scan# 2031

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:228 Resp: 1066153

Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.2	33.2
229	19.7	16.2	24.2

228 100

226 27.6 22.2 33.2

229 19.7 16.2 24.2






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#82

3,3'-Dichlorobenzidine

Concen: 41.885 ng

RT: 13.99 min Scan# 2023

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:252 Resp: 405417

Ion	Ratio	Lower	Upper
252	100		
254	64.0	52.1	78.1
126	11.0	8.6	12.8

252 100

254 64.0 52.1 78.1

126 11.0 8.6 12.8






Abundance

Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#83

Chrysene

Concen: 39.654 ng

RT: 14.07 min Scan# 2037

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:228 Resp: 991282

Ion Ratio Lower Upper

228 100

226 30.4 24.6 37.0




229 19.8 15.7 23.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#84

Bis(2-ethylhexyl)phthalate

Concen: 44.651 ng

RT: 14.01 min Scan# 2027

Delta R.T. -0.01 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:149 Resp: 777086

Ion Ratio Lower Upper

149 100

167 26.9 21.5 32.3



279 4.2 3.0 4.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#85

Di-n-octyl phthalate

Concen: 42.672 ng

RT: 14.62 min Scan# 2131

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:149 Resp: 1062777

Ion Ratio Lower Upper

149 100

167 1.5 1.3 1.9




43 10.7 9.1 13.7



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF1



#86

Indeno(1,2,3-cd)pyrene

Concen: 39.543 ng

RT: 17.01 min Scan# 2537

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:276 Resp: 720668

Ion Ratio Lower Upper

276 100

138 22.0 18.1 27.1



277 25.3 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#87

Pervlene-d12

Concen: 20.000 ng

RT: 15.52 min Scan# 2283

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40

Instrument :

BNA_F

ClientSampleId :

SSTDCCC040


Tot Ion:264 Resp: 334723

Ion Ratio Lower Upper

264 100

260 23.4 19.2 28.8




265 22.5 17.1 25.7



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#88

Benzo(b)fluoranthene

Concen: 39.577 ng

RT: 15.09 min Scan# 2210

Delta R.T. -0.00 min

Lab File: BF112047.D

Acq: 14 Jan 2019 13:40


Tgt Ion:252 Resp: 792456

Ion Ratio Lower Upper

252 100

253 22.3 17.0 25.6



125 8.7 7.1 10.7




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#89
Benzo(k)fluoranthene
Concen: 40.736 ng
RT: 15.12 min Scan# 2215
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40




Instrument :
BNA_F
ClientSampleId :
SSTDCCC040


Tot Ion:252 Resp: 766695
Ion Ratio Lower Upper
252 100
253 21.7 17.8 26.6
125 8.7 7.0 10.4



#90
Benzo(a)pyrene
Concen: 40.155 ng
RT: 15.46 min Scan# 2273
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:252 Resp: 706943
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.2
125 9.6 7.9 11.9









#91
Dibenzo(a,h)anthracene
Concen: 42.941 ng
RT: 17.02 min Scan# 2539
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Instrument :
BNA_F
ClientSampleId :
SSTDCCC040

Tot Ion:278 Resp: 608044
Ion Ratio Lower Upper
278 100
139 14.2 11.8 17.6
279 23.5 19.4 29.2




Abundance Ion 278.00 (277.70 to 278.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 279.00 (278.70 to 279.70): E



#92
Benzo(g,h,i)perylene
Concen: 42.592 ng
RT: 17.46 min Scan# 2614
Delta R.T. -0.00 min
Lab File: BF112047.D
Acq: 14 Jan 2019 13:40

Tgt Ion:276 Resp: 589884
Ion Ratio Lower Upper
276 100
277 23.4 19.0 28.6
138 19.1 16.2 24.2



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 277.00 (276.70 to 277.70): E
Ion 138.00 (137.70 to 138.70): E

