


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF011419\
 Data File : BF112046.D
 Acq On : 14 Jan 2019 13:13
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF010719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jan 11 21:37:26 2019



AutoFind: Scans 1624, 1625, 1626; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.2	36676	PASS
68	69	0.00	2	2.0	740	PASS
69	198	0.00	100	40.9	36434	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.2	43792	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	89064	PASS
199	198	5	9	6.5	5816	PASS
275	198	10	60	28.3	25210	PASS
365	198	1	100	3.7	3336	PASS
441	198	0.01	100	14.0	12487	PASS
442	198	50	100	89.2	79477	PASS
443	442	15	24	20.1	15940	PASS