

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\  
 Data File : BF112112.D  
 Acq On : 18 Jan 2019 2:30  
 Operator : JU/SJ  
 Sample : PB116445BL  
 Misc :  
 ALS Vial : 37 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 PB116445BL

Quant Time: Jan 18 03:30:25 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Jan 16 18:22:05 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|-------|------|----------|-------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.86  | 152  | 303756   | 20.00 | ng    | 0.00     |
| 21) Naphthalene-d8          | 8.15  | 136  | 1106058  | 20.00 | ng    | 0.00     |
| 39) Acenaphthene-d10        | 9.90  | 164  | 507398   | 20.00 | ng    | 0.00     |
| 64) Phenanthrene-d10        | 11.39 | 188  | 819012   | 20.00 | ng    | 0.00     |
| 76) Chrysene-d12            | 14.02 | 240  | 652062   | 20.00 | ng    | -0.01    |
| 87) Perylene-d12            | 15.50 | 264  | 507424   | 20.00 | ng    | 0.00     |
| System Monitoring Compounds |       |      |          |       |       |          |
| 5) 2-Fluorophenol           | 5.50  | 112  | 1224936  | 73.19 | ng    | 0.00     |
| 7) Phenol-d6                | 6.50  | 99   | 1510694  | 72.15 | ng    | 0.00     |
| 23) Nitrobenzene-d5         | 7.42  | 82   | 1336477  | 83.48 | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol    | 10.69 | 330  | 339206   | 61.97 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 9.22  | 172  | 2594173  | 74.96 | ng    | 0.00     |
| 79) Terphenyl-d14           | 12.97 | 244  | 1986626  | 64.81 | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\  
Data File : BF112112.D  
Acq On : 18 Jan 2019 2:30  
Operator : JU/SJ  
Sample : PB116445BL  
Misc :  
ALS Vial : 37 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
PB116445BL

Quant Time: Jan 18 03:30:25 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Jan 16 18:22:05 2019  
Response via : Initial Calibration

