

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\
 Data File : BF112075.D
 Acq On : 17 Jan 2019 8:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 17 10:14:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 16 18:22:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.517	0.499	3.5	112	0.00
3	Pyridine	1.294	1.313	-1.5	113	0.00
4	n-Nitrosodimethylamine	0.526	0.512	2.7	112	-0.01
5 S	2-Fluorophenol	1.102	1.073	2.6	113	0.00
6	Aniline	1.677	1.640	2.2	116	0.00
7 S	Phenol-d6	1.379	1.331	3.5	115	0.00
8	2-Chlorophenol	1.280	1.274	0.5	118	0.00
9	Benzaldehyde	0.780	0.723	7.3	115	0.00
10 C	Phenol	1.507	1.456	3.4	114	0.00
11	bis(2-Chloroethyl)ether	1.198	1.182	1.3	116	0.00
12	1,3-Dichlorobenzene	1.460	1.415	3.1	114	0.00
13 C	1,4-Dichlorobenzene	1.494	1.432	4.1	114	0.00
14	1,2-Dichlorobenzene	1.376	1.338	2.8	115	0.00
15	Benzyl Alcohol	1.027	1.033	-0.6	118	0.00
16	2,2'-oxybis(1-Chloropropane	1.728	1.707	1.2	117	0.00
17	2-Methylphenol	0.961	0.957	0.4	117	0.00
18	Hexachloroethane	0.499	0.504	-1.0	117	0.00
19 P	n-Nitroso-di-n-propylamine	0.840	0.837	0.4	118	0.00
20	3+4-Methylphenols	1.165	1.173	-0.7	117	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	118	0.00
22	Acetophenone	0.463	0.437	5.6	116	0.00
23 S	Nitrobenzene-d5	0.289	0.315	-9.0	123	0.00
24	Nitrobenzene	0.307	0.330	-7.5	122	0.00
25	Isophorone	0.559	0.547	2.1	120	0.00
26 C	2-Nitrophenol	0.145	0.166	-14.5	133	0.00
27	2,4-Dimethylphenol	0.260	0.256	1.5	119	0.00
28	bis(2-Chloroethoxy)methane	0.384	0.373	2.9	119	0.00
29 C	2,4-Dichlorophenol	0.284	0.284	0.0	118	0.00
30	1,2,4-Trichlorobenzene	0.329	0.317	3.6	116	0.00
31	Naphthalene	0.910	0.871	4.3	116	0.00
32	Benzoic acid	0.107	0.095	11.2	109	0.00
33	4-Chloroaniline	0.409	0.392	4.2	118	0.00
34 C	Hexachlorobutadiene	0.183	0.179	2.2	118	0.00
35	Caprolactam	0.099	0.100	-1.0	124	0.00
36 C	4-Chloro-3-methylphenol	0.277	0.275	0.7	120	0.00
37	2-Methylnaphthalene	0.618	0.598	3.2	118	0.00
38	1-Methylnaphthalene	0.594	0.572	3.7	117	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
40	1,2,4,5-Tetrachlorobenzene	0.642	0.613	4.5	119	0.00
41 P	Hexachlorocyclopentadiene	0.210	0.223	-6.2	127	0.00
42 S	2,4,6-Tribromophenol	0.216	0.230	-6.5	125	0.00
43 C	2,4,6-Trichlorophenol	0.384	0.394	-2.6	120	0.00
44	2,4,5-Trichlorophenol	0.409	0.421	-2.9	124	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\
 Data File : BF112075.D
 Acq On : 17 Jan 2019 8:34
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 17 10:14:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 16 18:22:05 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.364	1.188	12.9	115	0.00
46	1,1'-Biphenyl	1.672	1.575	5.8	117	0.00
47	2-Chloronaphthalene	1.309	1.240	5.3	118	0.00
48	2-Nitroaniline	0.295	0.331	-12.2	135	0.00
49	Acenaphthylene	1.910	1.807	5.4	118	0.00
50	Dimethylphthalate	1.459	1.405	3.7	122	0.00
51	2,6-Dinitrotoluene	0.291	0.321	-10.3	130	0.00
52 C	Acenaphthene	1.169	1.119	4.3	121	0.00
53	3-Nitroaniline	0.326	0.354	-8.6	130	0.00
54 P	2,4-Dinitrophenol	0.062	0.080	-29.0#	162#	0.00
55	Dibenzofuran	1.768	1.687	4.6	119	0.00
56 P	4-Nitrophenol	0.202	0.224	-10.9	132	0.00
57	2,4-Dinitrotoluene	0.347	0.399	-15.0	135	0.00
58	Fluorene	1.321	1.250	5.4	119	0.00
59	2,3,4,6-Tetrachlorophenol	0.309	0.328	-6.1	125	0.00
60	Diethylphthalate	1.424	1.384	2.8	122	0.00
61	4-Chlorophenyl-phenylether	0.713	0.680	4.6	119	0.00
62	4-Nitroaniline	0.323	0.352	-9.0	132	0.00
63	Azobenzene	1.363	1.320	3.2	121	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	121	0.00
65	4,6-Dinitro-2-methylphenol	0.071	0.089	-25.4#	152#	0.00
66 c	n-Nitrosodiphenylamine	0.661	0.636	3.8	120	0.00
67	4-Bromophenyl-phenylether	0.232	0.226	2.6	122	0.00
68	Hexachlorobenzene	0.250	0.244	2.4	121	0.00
69	Atrazine	0.213	0.210	1.4	123	0.00
70 C	Pentachlorophenol	0.103	0.114	-10.7	132	0.00
71	Phenanthrene	1.007	0.956	5.1	119	0.00
72	Anthracene	1.020	0.966	5.3	119	0.00
73	Carbazole	1.021	0.953	6.7	118	0.00
74	Di-n-butylphthalate	1.149	1.097	4.5	121	0.00
75 C	Fluoranthene	1.063	0.998	6.1	119	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	125	0.00
77	Benzidine	0.667	0.597	10.5	139	0.00
78	Pyrene	1.320	1.223	7.3	119	0.00
79 S	Terphenyl-d14	0.940	0.831	11.6	117	0.00
80	Butylbenzylphthalate	0.579	0.574	0.9	124	0.00
81	Benzo(a)anthracene	1.147	1.078	6.0	123	0.00
82	3,3'-Dichlorobenzidine	0.429	0.420	2.1	127	0.00
83	Chrysene	1.085	1.018	6.2	122	0.00
84	Bis(2-ethylhexyl)phthalate	0.830	0.810	2.4	125	0.00
85 c	Di-n-octyl phthalate	1.278	1.248	2.3	127	0.00
86	Indeno(1,2,3-cd)pyrene	1.001	0.902	9.9	138	0.00
87 I	Perylene-d12	1.000	1.000	0.0	137	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\
Data File : BF112075.D
Acq On : 17 Jan 2019 8:34
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 17 10:14:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 16 18:22:05 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.135	1.109	2.3	137	0.00
89	Benzo(k)fluoranthene	1.099	1.050	4.5	130	0.00
90 C	Benzo(a)pyrene	1.041	1.009	3.1	137	0.00
91	Dibenzo(a,h)anthracene	0.925	0.843	8.9	138	0.00
92	Benzo(q,h,i)perylene	0.911	0.810	11.1	135	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0