

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\
 Data File : BF112092.D
 Acq On : 17 Jan 2019 16:32
 Operator : JU/SJ
 Sample : K1156-10
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 COMP-7

Manual Integrations
 APPROVED

Sohil
 1/18/2019 11:45:52 AM

Quant Time: Jan 18 00:40:17 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 16 18:22:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	297317	20.00	ng	0.00
21) Naphthalene-d8	8.15	136	1099823	20.00	ng	0.00
39) Acenaphthene-d10	9.90	164	476426	20.00	ng	0.00
64) Phenanthrene-d10	11.39	188	776949	20.00	ng	0.00
76) Chrysene-d12	14.03	240	660151	20.00	ng	0.00
87) Perylene-d12	15.51	264	556088	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	1492399	91.11	ng	0.00
7) Phenol-d6	6.50	99	1909987	93.19	ng	0.00
23) Nitrobenzene-d5	7.42	82	1127914	70.85	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	434932	84.62	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	1943559	59.81	ng	0.00
79) Terphenyl-d14	12.97	244	1620560	52.22	ng	0.00
Target Compounds						
10) Phenol	6.52	94	113337	5.060	ng	Qvalue # 74
31) Naphthalene	8.16	128	282139	5.640	ng	98
37) 2-Methylnaphthalene	8.86	142	160364	4.721	ng	100
38) 1-Methylnaphthalene	8.96	142	138701	4.249	ng	99
50) Dimethylphthalate	9.60	163	186633	5.372	ng	98
52) Acenaphthene	9.93	154	342083	12.287	ng	98
55) Dibenzofuran	10.10	168	403379	9.580	ng	97
58) Fluorene	10.45	166	435762	13.852	ng	99
71) Phenanthrene	11.42	178	3782624	96.741	ng	96
72) Anthracene	11.46	178	1317927	33.274	ng	99
73) Carbazole	11.62	167	347308	8.753	ng	100
75) Fluoranthene	12.61	202	3829405	92.759	ng	96
78) Pvrene	12.84	202	3442846	79.048	ng	97
81) Benzo(a)anthracene	14.02	228	1931320	51.024	ng	98
83) Chrysene	14.06	228	1537079	42.916	ng	97
86) Indeno(1,2,3-cd)pyrene	16.99	276	530911	16.067	ng	93
88) Benzo(b)fluoranthene	15.08	252	1700922m	53.885	ng	
89) Benzo(k)fluoranthene	15.10	252	448423m	14.678	ng	
90) Benzo(a)pvrene	15.44	252	1109363	38.323	ng	96
91) Dibenzo(a,h)anthracene	16.99	278	142921	5.559	ng	# 89
92) Benzo(g,h,i)perylene	17.44	276	445858	17.602	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011719\
Data File : BF112092.D
Acq On : 17 Jan 2019 16:32
Operator : JU/SJ
Sample : K1156-10
Misc :
ALS Vial : 19 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
COMP-7

Manual Integrations
APPROVED

Sohil
1/18/2019 11:45:52 AM

Quant Time: Jan 18 00:40:17 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 16 18:22:05 2019
Response via : Initial Calibration






#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.86 min Scan# 812
 Delta R.T. 0.00 min
 Lab File: BF112092.D
 Acq: 17 Jan 2019 16:32

Instrument :
 BNA_F
 ClientSampleId :
 COMP-7

Tot Ion	Ratio	Lower	Upper
152	100		
150	154.2	127.4	191.0
115	53.0	45.5	68.3

Manual Integrations
 APPROVED

Sohil
 1/18/2019 11:45:52 AM






Abundance

Ion 152.00 (151.70 to 152.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5
 2-Fluorophenol
 Concen: 91.107 ng
 RT: 5.50 min Scan# 580
 Delta R.T. 0.01 min
 Lab File: BF112092.D
 Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Lower	Upper
112	100		
64	54.4	45.7	68.5
63	28.3	23.6	35.4






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 93.194 ng

RT: 6.50 min Scan# 751

Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32

Instrument :

BNA_F

ClientSampled :

COMP-7

Tot Ion: 99 Resp: 1909987

Ion Ratio Lower Upper

99 100

42 19.4 15.1 22.7





71 34.0 26.9 40.3

Manual Integrations

APPROVED

Sohil

1/18/2019 11:45:52 AM



#10

Phenol

Concen: 5.060 ng

RT: 6.52 min Scan# 753

Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32




Tgt Ion: 94 Resp: 113337


Ion Ratio Lower Upper

94 100

65 27.4 20.3 60.3

66 40.3 42.4 82.4#









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.15 min Scan# 1030
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	136	Resp	1099823
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.1	13.7
54	6.4	5.9	8.9
68	6.0	5.1	7.7




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#23
Nitrobenzene-d5
Concen: 70.854 ng
RT: 7.42 min Scan# 907
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	82	Resp	1127914
Ion	Ratio	Lower	Upper
82	100		
128	48.5	37.8	56.8
54	47.9	39.7	59.5









#31
Naphthalene
Concen: 5.640 ng
RT: 8.16 min Scan# 1033
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	Ratio	Resp	Lower	Upper
128	100			
129	10.9	9.6	14.4	
127	13.2	11.1	16.7	




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#37
2-Methylnaphthalene
Concen: 4.721 ng
RT: 8.86 min Scan# 1151
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Resp	Lower	Upper
142	100			
141	88.5	70.9	106.3	
115	31.4	24.9	37.3	









#38
1-Methylnaphthalene
Concen: 4.249 ng
RT: 8.96 min Scan# 1168
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	Ratio	Resp	Lower	Upper
142	100	138701		
141	91.3	72.6	108.8	
116	3.6	2.8	4.2	




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.90 min Scan# 1328
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Resp	Lower	Upper
164	100	476426		
162	101.5	82.2	123.4	
160	46.4	36.2	54.4	






#42
 2,4,6-Tribromophenol
 Concen: 84.624 ng
 RT: 10.69 min Scan# 1463
 Delta R.T. 0.00 min
 Lab File: BF112092.D
 Acq: 17 Jan 2019 16:32

Instrument :
 BNA_F
 ClientSampled :
 COMP-7

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.1	76.6	114.8
141	31.1	24.3	36.5

Manual Integrations
 APPROVED

Sohil
 1/18/2019 11:45:52 AM






Abundance

Ion 329.80 (329.50 to 330.50): E


Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E



#45
 2-Fluorobiphenyl
 Concen: 59.809 ng
 RT: 9.22 min Scan# 1212
 Delta R.T. -0.01 min
 Lab File: BF112092.D
 Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Lower	Upper
172	100		
171	37.6	30.5	45.7
170	24.6	20.2	30.4






Abundance

Ion 172.00 (171.70 to 172.70): E

Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 5.372 ng
RT: 9.60 min Scan# 1278
Delta R.T. -0.02 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	Ratio	Resp	Lower	Upper
163	100	186633		
194	4.5	3.3	4.9	
164	9.9	8.5	12.7	

Manual Integrations
APPROVED

Sohil
1/18/2019 11:45:52 AM






Abundance

Ion 163.00 (162.70 to 163.70): E


Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52
Acenaphthene
Concen: 12.287 ng
RT: 9.93 min Scan# 1334
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Resp	Lower	Upper
154	100	342083		
153	114.2	89.8	134.6	
152	55.5	43.5	65.3	






Abundance

Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#55

Dibenzofuran

Concen: 9.580 ng

RT: 10.10 min Scan# 1363

Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32

Instrument :

BNA_F





ClientSampled :

COMP-7

Tot Ion:	168	Resp:	403379
Ion Ratio	Lower	Upper	
168	100		
139	34.5	29.2	43.8
169	14.9	11.4	17.0

Manual Integrations
APPROVED

Sohil
1/18/2019 11:45:52 AM



#58

Fluorene

Concen: 13.852 ng




RT: 10.45 min Scan# 1421


Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32

Tgt Ion:	166	Resp:	435762
Ion Ratio	Lower	Upper	
166	100		
165	98.0	77.1	115.7
167	14.1	11.3	16.9









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.39 min Scan# 1581
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampled :
COMP-7

Tot Ion	Ratio	Resp	Lower	Upper
188	100	776949		
94	9.9		8.2	12.2
80	10.4		8.9	13.3




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#71
Phenanthrene
Concen: 96.741 ng
RT: 11.42 min Scan# 1586
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	3782624		
176	22.3		16.3	24.5
179	18.0		13.0	19.4





#72

Anthracene

Concen: 33.274 ng

RT: 11.46 min Scan# 1594

Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32

Instrument :

BNA_F

ClientSampled :

COMP-7

Tot Ion:178 Resp: 1317927

Ion Ratio Lower Upper

178 100


176 19.4 15.5 23.3

179 16.4 12.8 19.2




Manual Integrations
APPROVED

Sohil

1/18/2019 11:45:52 AM



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E



#73

Carbazole

Concen: 8.753 ng

RT: 11.62 min Scan# 1620

Delta R.T. -0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32


Tgt Ion:167 Resp: 347308

Ion Ratio Lower Upper



167 100


166 22.4 17.9 26.9

139 12.8 10.2 15.2



Abundance Ion 167.00 (166.70 to 167.70): E
Ion 166.00 (165.70 to 166.70): E
Ion 139.00 (138.70 to 139.70): E





#75

Fluoranthene

Concen: 92.759 ng

RT: 12.61 min Scan# 1789

Delta R.T. 0.01 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32

Instrument :

BNA_F

ClientSampleId :

COMP-7

Tot Ion:202 Resp: 3829405

Ion Ratio Lower Upper

202 100


101 13.2 0.0 31.8

203 20.1 0.0 38.4

Manual Integrations
APPROVED

Sohil




1/18/2019 11:45:52 AM



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng

RT: 14.03 min Scan# 2031

Delta R.T. 0.00 min

Lab File: BF112092.D

Acq: 17 Jan 2019 16:32


Tgt Ion:240 Resp: 660151

Ion Ratio Lower Upper

240 100

120 11.6 9.0 13.6



236 25.2 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 79.048 ng
RT: 12.84 min Scan# 1828
Delta R.T. 0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion:202 Resp: 3442846
Ion Ratio Lower Upper
202 100
200 23.7 17.7 26.5
203 19.7 14.6 22.0




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#79
Terphenyl-d14
Concen: 52.223 ng
RT: 12.97 min Scan# 1850
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion:244 Resp: 1620560
Ion Ratio Lower Upper
244 100
212 6.8 5.9 8.9
122 12.0 9.8 14.6









#81
Benzo(a)anthracene
Concen: 51.024 ng
RT: 14.02 min Scan# 2029
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion:228 Resp: 1931320
Ion Ratio Lower Upper
228 100
226 29.3 22.6 34.0
229 21.2 16.5 24.7




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#83
Chrysene
Concen: 42.916 ng
RT: 14.06 min Scan# 2035
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion:228 Resp: 1537079
Ion Ratio Lower Upper
228 100
226 31.4 25.0 37.6
229 23.7 16.6 24.8






#86
Indeno(1,2,3-cd)pyrene
Concen: 16.067 ng
RT: 16.99 min Scan# 2534
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	Ratio	Lower	Upper
276	100		
138	23.2	22.6	34.0
277	23.9	20.3	30.5

Manual Integrations
APPROVED

Sohil
1/18/2019 11:45:52 AM






Abundance

Ion 276.00 (275.70 to 276.70): E


Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#87
Perylene-d12
Concen: 20.000 ng
RT: 15.51 min Scan# 2282
Delta R.T. 0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Lower	Upper
264	100		
260	23.5	18.2	27.2
265	23.9	17.0	25.4






Abundance

Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#88
Benzo(b)fluoranthene
Concen: 53.885 ng m
RT: 15.08 min Scan# 2209
Delta R.T. 0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampleId :
COMP-7

Tot Ion	Ratio	Lower	Upper
252	100		
253	22.9	18.0	27.0
125	12.2	8.7	13.1




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#89
Benzo(k)fluoranthene
Concen: 14.678 ng m
RT: 15.10 min Scan# 2213
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	18.0	27.0
125	11.7	9.1	13.7









#90
Benzo(a)pyrene
Concen: 38.323 ng
RT: 15.44 min Scan# 2271
Delta R.T. 0.00 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Instrument :
BNA_F
ClientSampled :
COMP-7

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.6	17.5	26.3
125	13.2	9.0	13.4




Manual Integrations
APPROVED


Sohil
1/18/2019 11:45:52 AM



#91
Dibenzo(a,h)anthracene
Concen: 5.559 ng
RT: 16.99 min Scan# 2534
Delta R.T. -0.01 min
Lab File: BF112092.D
Acq: 17 Jan 2019 16:32

Tgt Ion	Ratio	Lower	Upper
278	100		
139	22.5	13.7	20.5#
279	28.4	19.0	28.6





#92
 Benzo(a,h,i)perylene
 Concen: 17.602 ng
 RT: 17.44 min Scan# 2610
 Delta R.T. 0.00 min
 Lab File: BF112092.D
 Acq: 17 Jan 2019 16:32

Instrument :
 BNA_F
 ClientSampleId :
 COMP-7

Tot Ion:	276	Resp:	445858
Ion Ratio	Lower	Upper	
276	100		
277	24.5	19.4	29.0
138	25.4	19.1	28.7

Manual Integrations
 APPROVED

Sohil
 1/18/2019 11:45:52 AM

