

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011819\
 Data File : BF112119.D
 Acq On : 18 Jan 2019 11:30
 Operator : JU/SJ
 Sample : K1157-07
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-04-011519

Manual Integrations
 APPROVED

Sohil
 1/21/2019 12:00:38 PM

Quant Time: Jan 18 12:30:38 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jan 16 18:22:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	343863	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	1271951	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	608180	20.00	ng	-0.01
64) Phenanthrene-d10	11.38	188	1013550	20.00	ng	0.00
76) Chrysene-d12	14.02	240	712983	20.00	ng	-0.01
87) Perylene-d12	15.49	264	744213	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	2244908	118.49	ng	0.00
7) Phenol-d6	6.51	99	2828975	119.35	ng	0.00
23) Nitrobenzene-d5	7.42	82	1818084	98.75	ng	0.00
42) 2,4,6-Tribromophenol	10.69	330	809546	123.39	ng	0.00
45) 2-Fluorobiphenyl	9.22	172	3194090	77.00	ng	0.00
79) Terphenyl-d14	12.96	244	2588209	77.23	ng	-0.01
Target Compounds						
10) Phenol	6.52	94	61106	2.359	ng	94
50) Dimethylphthalate	9.60	163	318783	7.187	ng	98
71) Phenanthrene	11.40	178	132549	2.599	ng	97
75) Fluoranthene	12.59	202	191357	3.553	ng	97
78) Pyrene	12.82	202	177182	3.767	ng	98
83) Chrysene	14.04	228	88952	2.300	ng	97
88) Benzo(b)fluoranthene	15.06	252	117377m	2.779	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011819\
Data File : BF112119.D
Acq On : 18 Jan 2019 11:30
Operator : JU/SJ
Sample : K1157-07
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-04-011519

Manual Integrations
APPROVED

Sohil
1/21/2019 12:00:38 PM

Quant Time: Jan 18 12:30:38 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jan 16 18:22:05 2019
Response via : Initial Calibration

