

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011819\
 Data File : BF112131.D
 Acq On : 18 Jan 2019 18:20
 Operator : JU/SJ
 Sample : K1191-08 2X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SITE4-WEST

Manual Integrations
 APPROVED

Sohil
 1/21/2019 12:00:53 PM

Quant Time: Jan 19 00:21:33 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 18 23:33:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	278758	20.00	ng	0.00
21) Naphthalene-d8	8.14	136	1103396	20.00	ng	0.00
39) Acenaphthene-d10	9.89	164	562197	20.00	ng	0.00
64) Phenanthrene-d10	11.38	188	1008409	20.00	ng	0.00
76) Chrysene-d12	14.02	240	673069	20.00	ng	0.00
87) Perylene-d12	15.50	264	629536	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	737763	48.04	ng	0.00
7) Phenol-d6	6.50	99	1042086	54.23	ng	0.00
23) Nitrobenzene-d5	7.41	82	646902	40.51	ng	-0.02
42) 2,4,6-Tribromophenol	10.68	330	155196	25.59	ng	-0.01
45) 2-Fluorobiphenyl	9.21	172	1430978	37.32	ng	0.00
79) Terphenyl-d14	12.96	244	1307702	41.33	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.60	163	122981	3.000	ng	97
58) Fluorene	10.44	166	83726	2.255	ng	98
71) Phenanthrene	11.40	178	811934	15.999	ng	97
72) Anthracene	11.45	178	203771	3.964	ng	98
75) Fluoranthene	12.59	202	858733	16.026	ng	99
78) Pyrene	12.82	202	819107	18.446	ng	99
81) Benzo(a)anthracene	14.01	228	333931	8.653	ng	99
83) Chrysene	14.04	228	297534	8.148	ng	97
86) Indeno(1,2,3-cd)pyrene	16.97	276	115972	3.442	ng	# 92
88) Benzo(b)fluoranthene	15.07	252	351180m	9.827	ng	
89) Benzo(k)fluoranthene	15.09	252	80734m	2.334	ng	
90) Benzo(a)pyrene	15.43	252	240720	7.345	ng	# 95
92) Benzo(a,h,i)perylene	17.42	276	105846	3.691	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF011819\
Data File : BF112131.D
Acq On : 18 Jan 2019 18:20
Operator : JU/SJ
Sample : K1191-08 2X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SITE4-WEST

Manual Integrations
APPROVED

Sohil
1/21/2019 12:00:53 PM

Quant Time: Jan 19 00:21:33 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF011619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 18 23:33:58 2019
Response via : Initial Calibration

