

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012519\
 Data File : BF112238.D
 Acq On : 25 Jan 2019 18:07
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 28 03:07:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 02:58:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2	1,4-Dioxane	0.375	0.425	-13.3	88	0.00
3	Pyridine	0.954	1.168	-22.4	95	0.00
4	n-Nitrosodimethylamine	0.401	0.502	-25.2#	92	0.00
5 S	2-Fluorophenol	0.942	1.177	-24.9	85	0.00
6	Aniline	1.556	1.812	-16.5	84	0.00
7 S	Phenol-d6	1.308	1.521	-16.3	84	0.00
8	2-Chlorophenol	1.267	1.349	-6.5	79	0.00
9	Benzaldehyde	0.684	0.744	-8.8	80	0.00
10 C	Phenol	1.435	1.700	-18.5	86	0.00
11	bis(2-Chloroethyl)ether	1.130	1.260	-11.5	82	0.00
12	1,3-Dichlorobenzene	1.475	1.529	-3.7	79	0.00
13 C	1,4-Dichlorobenzene	1.518	1.548	-2.0	79	0.00
14	1,2-Dichlorobenzene	1.421	1.265	11.0	69	0.00
15	Benzyl Alcohol	1.103	1.011	8.3	70	0.00
16	2,2'-oxybis(1-Chloropropane	1.451	1.224	15.6	64	0.00
17	2-Methylphenol	1.004	0.881	12.3	68	0.00
18	Hexachloroethane	0.533	0.489	8.3	70	0.00
19 P	n-Nitroso-di-n-propylamine	0.902	0.811	10.1	69	0.00
20	3+4-Methylphenols	1.284	1.182	7.9	71	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
22	Acetophenone	0.487	0.495	-1.6	71	0.00
23 S	Nitrobenzene-d5	0.347	0.352	-1.4	70	0.00
24	Nitrobenzene	0.347	0.350	-0.9	69	0.00
25	Isophorone	0.545	0.549	-0.7	71	0.00
26 C	2-Nitrophenol	0.185	0.190	-2.7	72	0.00
27	2,4-Dimethylphenol	0.255	0.255	0.0	73	0.00
28	bis(2-Chloroethoxy)methane	0.371	0.372	-0.3	75	0.00
29 C	2,4-Dichlorophenol	0.286	0.288	-0.7	74	0.00
30	1,2,4-Trichlorobenzene	0.328	0.328	0.0	76	0.00
31	Naphthalene	0.935	0.939	-0.4	77	0.00
32	Benzoic acid	0.146	0.132	9.6	65	0.00
33	4-Chloroaniline	0.407	0.406	0.2	76	0.00
34 C	Hexachlorobutadiene	0.193	0.191	1.0	76	0.00
35	Caprolactam	0.101	0.100	1.0	77	0.00
36 C	4-Chloro-3-methylphenol	0.302	0.307	-1.7	78	0.00
37	2-Methylnaphthalene	0.640	0.637	0.5	76	0.00
38	1-Methylnaphthalene	0.614	0.615	-0.2	77	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	71	0.00
40	1,2,4,5-Tetrachlorobenzene	0.657	0.635	3.3	74	0.00
41 P	Hexachlorocyclopentadiene	0.195	0.199	-2.1	75	0.00
42 S	2,4,6-Tribromophenol	0.233	0.252	-8.2	88	0.00
43 C	2,4,6-Trichlorophenol	0.405	0.402	0.7	74	0.00
44	2,4,5-Trichlorophenol	0.423	0.430	-1.7	75	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012519\
 Data File : BF112238.D
 Acq On : 25 Jan 2019 18:07
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 28 03:07:56 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 02:58:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.414	1.425	-0.8	79	0.00
46	1,1'-Biphenyl	1.748	1.754	-0.3	77	0.00
47	2-Chloronaphthalene	1.356	1.359	-0.2	77	0.00
48	2-Nitroaniline	0.370	0.385	-4.1	78	0.00
49	Acenaphthylene	1.987	1.992	-0.3	76	0.00
50	Dimethylphthalate	1.554	1.554	0.0	78	0.00
51	2,6-Dinitrotoluene	0.348	0.353	-1.4	78	0.00
52 C	Acenaphthene	1.187	1.215	-2.4	77	0.00
53	3-Nitroaniline	0.377	0.391	-3.7	79	0.00
54 P	2,4-Dinitrophenol	0.112	0.114	-1.8	76	0.00
55	Dibenzofuran	1.814	1.858	-2.4	78	0.00
56 P	4-Nitrophenol	0.199	0.215	-8.0	78	0.00
57	2,4-Dinitrotoluene	0.443	0.480	-8.4	81	0.00
58	Fluorene	1.358	1.426	-5.0	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.320	0.338	-5.6	75	0.00
60	Diethylphthalate	1.542	1.586	-2.9	78	0.00
61	4-Chlorophenyl-phenylether	0.738	0.748	-1.4	78	0.00
62	4-Nitroaniline	0.370	0.399	-7.8	83	0.00
63	Azobenzene	1.356	1.429	-5.4	78	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
65	4,6-Dinitro-2-methylphenol	0.112	0.116	-3.6	78	0.00
66 c	n-Nitrosodiphenylamine	0.674	0.672	0.3	78	0.00
67	4-Bromophenyl-phenylether	0.235	0.234	0.4	78	0.00
68	Hexachlorobenzene	0.252	0.250	0.8	78	0.00
69	Atrazine	0.217	0.217	0.0	78	0.00
70 C	Pentachlorophenol	0.098	0.103	-5.1	83	0.00
71	Phenanthrene	1.040	1.050	-1.0	78	0.00
72	Anthracene	1.036	1.068	-3.1	88	0.00
73	Carbazole	1.022	1.052	-2.9	80	0.00
74	Di-n-butylphthalate	1.268	1.275	-0.6	79	0.00
75 C	Fluoranthene	1.115	1.096	1.7	71	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	69	0.00
77	Benzidine	0.550	0.524	4.7	64	0.00
78	Pyrene	1.385	1.326	4.3	72	0.00
79 S	Terphenyl-d14	1.062	1.003	5.6	73	0.00
80	Butylbenzylphthalate	0.670	0.653	2.5	72	0.00
81	Benzo(a)anthracene	1.176	1.156	1.7	71	0.00
82	3,3'-Dichlorobenzidine	0.440	0.424	3.6	70	0.00
83	Chrysene	1.122	1.099	2.0	74	0.00
84	Bis(2-ethylhexyl)phthalate	0.951	0.912	4.1	71	0.00
85 c	Di-n-octyl phthalate	1.463	1.452	0.8	73	0.00
86	Indeno(1,2,3-cd)pyrene	0.889	0.788	11.4	76	0.00
87 I	Perylene-d12	1.000	1.000	0.0	70	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012519\
Data File : BF112238.D
Acq On : 25 Jan 2019 18:07
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Jan 28 03:07:56 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 02:58:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.196	1.285	-7.4	76	0.00
89	Benzo(k)fluoranthene	1.146	1.114	2.8	71	0.00
90 C	Benzo(a)pyrene	1.076	1.081	-0.5	73	0.00
91	Dibenzo(a,h)anthracene	0.867	0.836	3.6	78	0.00
92	Benzo(q,h,i)perylene	0.818	0.779	4.8	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0