

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012721\
 Data File : BF123046.D
 Acq On : 27 Jan 2021 16:38
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012721

Quant Time: Jan 28 01:06:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.645	0.656	-1.7	102	0.00
3	Pyridine	1.725	1.754	-1.7	100	0.00
4	n-Nitrosodimethylamine	0.726	0.739	-1.8	99	0.00
5 S	2-Fluorophenol	1.297	1.303	-0.5	100	0.00
6	Aniline	2.163	2.211	-2.2	100	0.00
7 S	Phenol-d6	1.757	1.779	-1.3	101	0.00
8	2-Chlorophenol	1.405	1.427	-1.6	100	0.00
9	Benzaldehyde	0.803	0.821	-2.2	100	0.00
10 C	Phenol	1.743	1.775	-1.8	98	0.00
11	bis(2-Chloroethyl)ether	1.450	1.460	-0.7	99	0.00
12	1,3-Dichlorobenzene	1.639	1.650	-0.7	100	0.00
13 C	1,4-Dichlorobenzene	1.707	1.652	3.2	100	0.00
14	1,2-Dichlorobenzene	1.597	1.541	3.5	100	0.00
15	Benzyl Alcohol	1.232	1.295	-5.1	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.237	2.267	-1.3	99	0.00
17	2-Methylphenol	1.196	1.213	-1.4	99	0.00
18	Hexachloroethane	0.598	0.608	-1.7	100	0.00
19 P	n-Nitroso-di-n-propylamine	1.057	1.069	-1.1	99	0.00
20	3+4-Methylphenols	1.409	1.484	-5.3	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
22	Acetophenone	0.521	0.520	0.2	99	0.00
23 S	Nitrobenzene-d5	0.397	0.415	-4.5	101	0.00
24	Nitrobenzene	0.409	0.424	-3.7	100	0.00
25	Isophorone	0.728	0.744	-2.2	100	0.00
26 C	2-Nitrophenol	0.169	0.193	-14.2	102	0.00
27	2,4-Dimethylphenol	0.306	0.316	-3.3	101	0.00
28	bis(2-Chloroethoxy)methane	0.497	0.504	-1.4	100	0.00
29 C	2,4-Dichlorophenol	0.319	0.331	-3.8	100	0.00
30	1,2,4-Trichlorobenzene	0.360	0.364	-1.1	100	0.00
31	Naphthalene	1.097	1.102	-0.5	100	0.00
32	Benzoic acid	0.224	0.249	-11.2	101	0.00
33	4-Chloroaniline	0.487	0.490	-0.6	99	0.00
34 C	Hexachlorobutadiene	0.208	0.213	-2.4	99	0.00
35	Caprolactam	0.108	0.111	-2.8	99	0.00
36 C	4-Chloro-3-methylphenol	0.332	0.344	-3.6	99	0.00
37	2-Methylnaphthalene	0.728	0.734	-0.8	100	0.00
38	1-Methylnaphthalene	0.690	0.699	-1.3	100	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
40	1,2,4,5-Tetrachlorobenzene	0.622	0.622	0.0	99	0.00
41 P	Hexachlorocyclopentadiene	0.295	0.301	-2.0	94	0.00
42 S	2,4,6-Tribromophenol	0.217	0.227	-4.6	101	0.00
43 C	2,4,6-Trichlorophenol	0.427	0.456	-6.8	101	0.00
44	2,4,5-Trichlorophenol	0.446	0.464	-4.0	100	0.00
45 S	2-Fluorobiphenyl	1.345	1.332	1.0	100	0.00
46	1,1'-Biphenyl	1.640	1.645	-0.3	99	0.00
47	2-Chloronaphthalene	1.373	1.385	-0.9	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012721\
 Data File : BF123046.D
 Acq On : 27 Jan 2021 16:38
 Operator : JU/CG
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012721

Quant Time: Jan 28 01:06:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 28 00:48:07 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.416	0.452	-8.7	101	0.00
49	Acenaphthylene	2.010	2.018	-0.4	99	0.00
50	Dimethylphthalate	1.569	1.572	-0.2	100	0.00
51	2,6-Dinitrotoluene	0.334	0.355	-6.3	99	0.00
52 C	Acenaphthene	1.291	1.311	-1.5	100	0.00
53	3-Nitroaniline	0.395	0.416	-5.3	100	0.00
54 P	2,4-Dinitrophenol	0.126	0.147	-16.7	106	0.00
55	Dibenzofuran	1.879	1.849	1.6	101	0.00
56 P	4-Nitrophenol	0.286	0.316	-10.5	103	0.00
57	2,4-Dinitrotoluene	0.435	0.472	-8.5	101	0.00
58	Fluorene	1.501	1.357	9.6	99	0.00
59	2,3,4,6-Tetrachlorophenol	0.378	0.403	-6.6	103	0.00
60	Diethylphthalate	1.543	1.543	0.0	99	0.00
61	4-Chlorophenyl-phenylether	0.705	0.674	4.4	99	0.00
62	4-Nitroaniline	0.397	0.411	-3.5	99	0.00
63	Azobenzene	1.510	1.504	0.4	98	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.112	-10.9	102	0.00
66 c	n-Nitrosodiphenylamine	0.721	0.734	-1.8	99	0.00
67	4-Bromophenyl-phenylether	0.255	0.262	-2.7	99	0.00
68	Hexachlorobenzene	0.273	0.276	-1.1	97	0.00
69	Atrazine	0.199	0.201	-1.0	99	0.00
70 C	Pentachlorophenol	0.148	0.162	-9.5	101	0.00
71	Phenanthrene	1.242	1.197	3.6	99	0.00
72	Anthracene	1.192	1.187	0.4	99	0.00
73	Carbazole	1.106	1.110	-0.4	99	0.00
74	Di-n-butylphthalate	1.311	1.334	-1.8	99	0.00
75 C	Fluoranthene	1.241	1.231	0.8	100	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.453	0.522	-15.2	94	0.00
78	Pyrene	1.510	1.506	0.3	98	0.00
79 S	Terphenyl-d14	1.018	1.026	-0.8	98	0.00
80	Butylbenzylphthalate	0.650	0.690	-6.2	100	0.00
81	Benzo(a)anthracene	1.354	1.358	-0.3	100	0.00
82	3,3'-Dichlorobenzidine	0.426	0.453	-6.3	99	0.00
83	Chrysene	1.355	1.356	-0.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.863	0.890	-3.1	99	0.00
85 c	Di-n-octyl phthalate	1.450	1.562	-7.7	99	0.00
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.332	1.422	-6.8	102	0.00
88	Benzo(b)fluoranthene	1.444	1.470	-1.8	98	0.00
89	Benzo(k)fluoranthene	1.341	1.370	-2.2	103	0.00
90 C	Benzo(a)pyrene	1.268	1.333	-5.1	101	0.00
91	Dibenzo(a,h)anthracene	1.096	1.166	-6.4	101	0.00
92	Benzo(g,h,i)perylene	1.062	1.125	-5.9	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012721\
Data File : BF123046.D
Acq On : 27 Jan 2021 16:38
Operator : JU/CG
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF012721

Quant Time: Jan 28 01:06:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jan 28 00:48:07 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0