

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012724\
 Data File : BF137181.D
 Acq On : 26 Jan 2024 22:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012724

Quant Time: Jan 27 02:16:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 02:14:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	41.712	-4.3	98	0.00
3	Pyridine	40.000	42.805	-7.0	104	0.00
4	n-Nitrosodimethylamine	40.000	39.185	2.0	96	0.00
5 S	2-Fluorophenol	80.000	95.683	-19.6	100	0.00
6	Aniline	40.000	43.200	-8.0	102	0.00
7 S	Phenol-d6	80.000	88.794	-11.0	96	0.00
8	2-Chlorophenol	40.000	45.051	-12.6	101	0.00
9	Benzaldehyde	40.000	52.779	-31.9#	104	0.00
10 C	Phenol	40.000	45.392	-13.5	98	0.00
11	bis(2-Chloroethyl)ether	40.000	40.031	-0.1	99	0.00
12	1,3-Dichlorobenzene	40.000	42.437	-6.1	100	0.00
13 C	1,4-Dichlorobenzene	40.000	40.752	-1.9	94	0.00
14	1,2-Dichlorobenzene	40.000	42.191	-5.5	98	0.00
15	Benzyl Alcohol	40.000	44.537	-11.3	100	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.385	-1.0	95	0.00
17	2-Methylphenol	40.000	43.161	-7.9	94	0.00
18	Hexachloroethane	40.000	43.139	-7.8	100	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	42.128	-5.3	98	0.00
20	3+4-Methylphenols	40.000	44.948	-12.4	99	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	98	0.00
22	Acetophenone	40.000	42.963	-7.4	102	0.00
23 S	Nitrobenzene-d5	80.000	85.264	-6.6	100	0.00
24	Nitrobenzene	40.000	43.582	-9.0	101	0.00
25	Isophorone	40.000	40.897	-2.2	100	0.00
26 C	2-Nitrophenol	40.000	43.032	-7.6	107	0.00
27	2,4-Dimethylphenol	40.000	43.224	-8.1	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	42.144	-5.4	99	0.00
29 C	2,4-Dichlorophenol	40.000	45.303	-13.3	100	0.00
30	1,2,4-Trichlorobenzene	40.000	40.738	-1.8	98	0.00
31	Naphthalene	40.000	40.722	-1.8	98	0.00
32	Benzoic acid	40.000	48.196	-20.5	137	0.00
33	4-Chloroaniline	40.000	45.215	-13.0	97	0.00
34 C	Hexachlorobutadiene	40.000	40.522	-1.3	98	0.00
35	Caprolactam	40.000	44.737	-11.8	100	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.551	-8.9	102	0.00
37	2-Methylnaphthalene	40.000	41.181	-3.0	98	0.00
38	1-Methylnaphthalene	40.000	40.565	-1.4	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.891	-2.2	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	43.318	-8.3	98	0.00
42 S	2,4,6-Tribromophenol	80.000	86.230	-7.8	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.277	-8.2	98	0.00
44	2,4,5-Trichlorophenol	40.000	47.071	-17.7	109	0.00
45 S	2-Fluorobiphenyl	80.000	81.141	-1.4	99	0.00
46	1,1'-Biphenyl	40.000	41.230	-3.1	98	0.00
47	2-Chloronaphthalene	40.000	41.531	-3.8	101	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012724\
 Data File : BF137181.D
 Acq On : 26 Jan 2024 22:45
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF012724

Quant Time: Jan 27 02:16:39 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 27 02:14:33 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	46.874	-17.2	100	0.00
49	Acenaphthylene	40.000	41.081	-2.7	99	0.00
50	Dimethylphthalate	40.000	40.912	-2.3	97	0.00
51	2,6-Dinitrotoluene	40.000	44.319	-10.8	101	0.00
52 C	Acenaphthene	40.000	40.880	-2.2	99	0.00
53	3-Nitroaniline	40.000	42.628	-6.6	100	0.00
54 P	2,4-Dinitrophenol	40.000	44.721	-11.8	125	0.00
55	Dibenzofuran	40.000	41.066	-2.7	98	0.00
56 P	4-Nitrophenol	40.000	43.066	-7.7	111	0.00
57	2,4-Dinitrotoluene	40.000	47.059	-17.6	103	0.00
58	Fluorene	40.000	40.860	-2.1	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.866	-7.2	100	0.00
60	Diethylphthalate	40.000	41.580	-3.9	99	0.00
61	4-Chlorophenyl-phenylether	40.000	40.623	-1.6	98	0.00
62	4-Nitroaniline	40.000	44.464	-11.2	99	0.00
63	Azobenzene	40.000	40.741	-1.9	97	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.992	-7.5	113	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.458	-3.6	99	0.00
67	4-Bromophenyl-phenylether	40.000	41.932	-4.8	100	0.00
68	Hexachlorobenzene	40.000	40.604	-1.5	98	0.00
69	Atrazine	40.000	42.123	-5.3	99	0.00
70 C	Pentachlorophenol	40.000	43.463	-8.7	103	0.00
71	Phenanthrene	40.000	40.768	-1.9	97	0.00
72	Anthracene	40.000	41.340	-3.4	100	0.00
73	Carbazole	40.000	41.387	-3.5	96	0.00
74	Di-n-butylphthalate	40.000	42.014	-5.0	98	0.00
75 C	Fluoranthene	40.000	40.052	-0.1	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	94	0.00
77	Benzydine	40.000	42.823	-7.1	106	0.00
78	Pyrene	40.000	42.876	-7.2	96	0.00
79 S	Terphenyl-d14	80.000	85.432	-6.8	97	0.00
80	Butylbenzylphthalate	40.000	41.628	-4.1	100	0.00
81	Benzo(a)anthracene	40.000	40.772	-1.9	94	0.00
82	3,3'-Dichlorobenzidine	40.000	42.515	-6.3	92	0.00
83	Chrysene	40.000	39.601	1.0	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	45.112	-12.8	99	0.00
85 c	Di-n-octyl phthalate	40.000	41.444	-3.6	96	0.00
86 I	Perylene-d12	20.000	20.000	0.0	92	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.176	-10.4	91	0.00
88	Benzo(b)fluoranthene	40.000	40.617	-1.5	94	0.00
89	Benzo(k)fluoranthene	40.000	38.948	2.6	88	0.00
90 C	Benzo(a)pyrene	40.000	40.808	-2.0	91	0.00
91	Dibenzo(a,h)anthracene	40.000	44.290	-10.7	92	0.00
92	Benzo(g,h,i)perylene	40.000	44.365	-10.9	92	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF012724\
Data File : BF137181.D
Acq On : 26 Jan 2024 22:45
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF012724

Quant Time: Jan 27 02:16:39 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 27 02:14:33 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0