

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012919\
 Data File : BF112292.D
 Acq On : 29 Jan 2019 13:52
 Operator : JU/SJ
 Sample : K1264-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 RH-CONCRETE

Manual Integrations
 APPROVED

mohammad
 1/30/2019 11:08:16 AM

Quant Time: Jan 30 00:19:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.84	152	176267	20.00	ng	0.00
21) Naphthalene-d8	8.12	136	664870	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	302770	20.00	ng	-0.01
64) Phenanthrene-d10	11.36	188	489427	20.00	ng	0.00
76) Chrysene-d12	14.01	240	326706	20.00	ng	0.00
87) Perylene-d12	15.47	264	394769	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	966275	116.44	ng	0.00
7) Phenol-d6	6.49	99	1211676	105.08	ng	0.00
23) Nitrobenzene-d5	7.40	82	762392	66.07	ng	-0.01
42) 2,4,6-Tribromophenol	10.67	330	314170	89.15	ng	-0.01
45) 2-Fluorobiphenyl	9.19	172	1511926	70.63	ng	-0.01
79) Terphenyl-d14	12.94	244	1119964	64.57	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.14	128	63280	2.035	ng	97
37) 2-Methylnaphthalene	8.83	142	45928	2.158	ng	97
50) Dimethylphthalate	9.59	163	90825	3.861	ng	96
52) Acenaphthene	9.90	154	100633	5.599	ng	98
55) Dibenzofuran	10.08	168	187032	6.810	ng	99
58) Fluorene	10.42	166	59522	2.896	ng	# 93
71) Phenanthrene	11.39	178	2614614	102.757	ng	99
72) Anthracene	11.44	178	164545	6.492	ng	97
73) Carbazole	11.59	167	143116	5.724	ng	# 94
75) Fluoranthene	12.59	202	2809921	102.962	ng	99
78) Pyrene	12.82	202	2214494	97.903	ng	99
81) Benzo(a)anthracene	14.00	228	652577	33.979	ng	99
83) Chrysene	14.03	228	608858	33.212	ng	97
84) Bis(2-ethylhexyl)phthalate	13.97	149	60387	3.887	ng	97
86) Indeno(1,2,3-cd)pyrene	16.94	276	266999	18.380	ng	# 92
88) Benzo(b)fluoranthene	15.04	252	714098m	30.247	ng	
89) Benzo(k)fluoranthene	15.07	252	242481m	10.723	ng	
90) Benzo(a)pyrene	15.41	252	236217	11.120	ng	# 95
91) Dibenzo(a,h)anthracene	16.94	278	70876	4.140	ng	# 79
92) Benzo(a,h,i)perylene	17.39	276	290487	17.984	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF012919\
Data File : BF112292.D
Acq On : 29 Jan 2019 13:52
Operator : JU/SJ
Sample : K1264-01
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM

Quant Time: Jan 30 00:19:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.84 min Scan# 808
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
152	100		
150	156.1	127.4	191.0
115	57.2	45.5	68.3

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol
Concen: 116.439 ng
RT: 5.47 min Scan# 576
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
112	100		
64	57.0	45.7	68.5
63	30.4	23.6	35.4






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 105.077 ng

RT: 6.49 min Scan# 748

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F

ClientSampleId :

RH-CONCRETE

Tot Ion: 99 Resp: 1211676

Ion Ratio Lower Upper

99 100

42 18.2 15.1 22.7


71 35.4 26.9 40.3

Manual Integrations




APPROVED

mohammad

1/30/2019 11:08:16 AM



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.12 min Scan# 1025

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Tgt Ion: 136 Resp: 664870

Ion Ratio Lower Upper




136 100


137 11.0 9.1 13.7

54 6.8 5.9 8.9

68 5.3 5.1 7.7

Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
Nitrobenzene-d5
Concen: 66.071 ng
RT: 7.40 min Scan# 903
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
82	100		
128	46.4	37.8	56.8
54	47.4	39.7	59.5




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#31
Naphthalene
Concen: 2.035 ng
RT: 8.14 min Scan# 1029
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.5	9.6	14.4
127	12.7	11.1	16.7









#37
2-Methylnaphthalene
Concen: 2.158 ng
RT: 8.83 min Scan# 1147
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
142	100		
141	85.1	70.9	106.3
115	30.3	24.9	37.3




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 9.87 min Scan# 1324
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
164	100		
162	102.9	82.2	123.4
160	46.1	36.2	54.4





#42

2,4,6-Tribromophenol

Concen: 89.148 ng

RT: 10.67 min Scan# 1459

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F


ClientSampleId :

RH-CONCRETE




Tot Ion:	330	Resp:	314170
Ion Ratio	Lower	Upper	
330	100		
332	95.9	76.6	114.8
141	32.9	24.3	36.5

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 70.627 ng


RT: 9.19 min Scan# 1208

Delta R.T. -0.01 min



Lab File: BF112292.D


Acq: 29 Jan 2019 13:52

Tgt Ion:	172	Resp:	1511926
Ion Ratio	Lower	Upper	
172	100		
171	36.6	30.5	45.7
170	24.5	20.2	30.4



Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 3.861 ng
RT: 9.59 min Scan# 1275
Delta R.T. -0.02 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE




Tot Ion:	163	Resp:	90825
Ion Ratio	Lower	Upper	
163	100		
194	4.7	3.3	4.9
164	8.6	8.5	12.7

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM






Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E




#52
Acenaphthene
Concen: 5.599 ng
RT: 9.90 min Scan# 1329
Delta R.T. -0.02 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion:	154	Resp:	100633
Ion Ratio	Lower	Upper	
154	100		
153	113.8	89.8	134.6
152	55.5	43.5	65.3

Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E









#55
Dibenzofuran
Concen: 6.810 ng
RT: 10.08 min Scan# 1359
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
168	100		
139	36.3	29.2	43.8
169	12.9	11.4	17.0




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#58
Fluorene
Concen: 2.896 ng
RT: 10.42 min Scan# 1417
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
166	100		
165	102.4	77.1	115.7
167	18.4	11.3	16.9#









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.36 min Scan# 1577
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Resp Lower	Upper
188	100		
94	10.4	8.2	12.2
80	11.2	8.9	13.3




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#71
Phenanthrene
Concen: 102.757 ng
RT: 11.39 min Scan# 1582
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Resp Lower	Upper
178	100		
176	20.7	16.3	24.5
179	16.9	13.0	19.4





#72

Anthracene

Concen: 6.492 ng

RT: 11.44 min Scan# 1590

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F

ClientSampleId :

RH-CONCRETE

Tot Ion:178 Resp: 164545

Ion Ratio Lower Upper

178 100

176 19.2 15.5 23.3


179 18.6 12.8 19.2

Manual Integrations

APPROVED

mohammad




1/30/2019 11:08:16 AM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#73

Carbazole

Concen: 5.724 ng

RT: 11.59 min Scan# 1616

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52


Tgt Ion:167 Resp: 143116

Ion Ratio Lower Upper

167 100

166 21.7 17.9 26.9



139 18.1 10.2 15.2#




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E





#75

Fluoranthene

Concen: 102.962 ng

RT: 12.59 min Scan# 1786

Delta R.T. 0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F


ClientSampleId :

RH-CONCRETE




Tot Ion:	202	Resp:	2809921
Ion	Ratio	Lower	Upper
202	100		
101	12.1	0.0	31.8
203	18.9	0.0	38.4

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#76

Chrysene-d12

Concen: 20.000 ng


RT: 14.01 min Scan# 2027

Delta R.T. 0.00 min



Lab File: BF112292.D


Acq: 29 Jan 2019 13:52

Tgt Ion:	240	Resp:	326706
Ion	Ratio	Lower	Upper
240	100		
120	12.7	9.0	13.6
236	25.9	20.7	31.1



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#78
Pvrene
Concen: 97.903 ng
RT: 12.82 min Scan# 1824
Delta R.T. 0.00 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
202	100		
200	22.5	17.7	26.5
203	18.5	14.6	22.0




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#79
Terphenyl-d14
Concen: 64.565 ng
RT: 12.94 min Scan# 1846
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.4	5.9	8.9
122	13.1	9.8	14.6









#81
Benzo(a)anthracene
Concen: 33.979 ng
RT: 14.00 min Scan# 2025
Delta R.T. 0.00 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
228	100		
226	27.8	22.6	34.0
229	20.9	16.5	24.7




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#83
Chrysene
Concen: 33.212 ng
RT: 14.03 min Scan# 2031
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.9	25.0	37.6
229	22.4	16.6	24.8





#84

Bis(2-ethylhexyl)phthalate

Concen: 3.887 ng

RT: 13.97 min Scan# 2021

Delta R.T. 0.00 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F


ClientSampleId :

RH-CONCRETE




Tot Ion:	149	Resp:	60387
Ion Ratio	Lower	Upper	
149	100		
167	30.0	22.5	33.7
279	5.4	3.8	5.6

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 167.00 (166.70 to 167.70): E
Ion 279.00 (278.70 to 279.70): E



#86

Indeno(1,2,3-cd)pyrene

Concen: 18.380 ng

RT: 16.94 min Scan# 2525




Delta R.T. -0.02 min


Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Tgt Ion:	276	Resp:	266999
Ion Ratio	Lower	Upper	
276	100		
138	21.8	22.6	34.0#
277	24.0	20.3	30.5

Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E









#87
Pervlene-d12
Concen: 20.000 ng
RT: 15.47 min Scan# 2276
Delta R.T. 0.00 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
264	100		
260	23.8	18.2	27.2
265	22.5	17.0	25.4




Manual Integrations
APPROVED


mohammad
1/30/2019 11:08:16 AM



#88
Benzo(b)fluoranthene
Concen: 30.247 ng m
RT: 15.04 min Scan# 2203
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	18.0	27.0
125	13.3	8.7	13.1#





#89

Benzo(k)fluoranthene

Concen: 10.723 ng m

RT: 15.07 min Scan# 2207

Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Instrument :

BNA_F





ClientSampleId :

RH-CONCRETE

Tot Ion:	252	Resp:	242481
Ion Ratio	Lower	Upper	
252	100		
253	22.6	18.0	27.0
125	13.8	9.1	13.7#

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM



#90

Benzo(a)pyrene

Concen: 11.120 ng




RT: 15.41 min Scan# 2265


Delta R.T. -0.01 min

Lab File: BF112292.D

Acq: 29 Jan 2019 13:52

Tgt Ion:	252	Resp:	236217
Ion Ratio	Lower	Upper	
252	100		
253	23.8	17.5	26.3
125	14.2	9.0	13.4#









#91
Dibenzo(a,h)anthracene
Concen: 4.140 ng
RT: 16.94 min Scan# 2525
Delta R.T. -0.02 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Instrument :
BNA_F
ClientSampleId :
RH-CONCRETE

Tot Ion	Ratio	Lower	Upper
278	100		
139	0.0	13.7	20.5#
279	27.9	19.0	28.6

Manual Integrations
APPROVED

mohammad
1/30/2019 11:08:16 AM



#92
Benzo(g,h,i)perylene
Concen: 17.984 ng
RT: 17.39 min Scan# 2601
Delta R.T. -0.01 min
Lab File: BF112292.D
Acq: 29 Jan 2019 13:52

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.4	29.0
138	22.7	19.1	28.7

