




Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF013019\
Data File : BF112317.D
Acq On : 30 Jan 2019 12:33
Operator : JU/SJ
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 30 18:13:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration



(70) Pentachlorophenol (C)




11.175min (-0.012) 58868.05ng

response 156660

Ion	Exp%	Act%
265.70	100	100
268.00	64.00	63.97
264.00	63.60	62.24
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA F\Data\BF013019\
 Data File : BF112317.D
 Acq On : 30 Jan 2019 12:33
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jan 30 18:13:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration



(77) Benzidine

12.698min (+0.000) 152242.29ng

response 1252879

Ion	Exp%	Act%
184.00	100	100
185.00	17.40	14.37
183.00	12.50	12.59
0.00	0.00	0.00