

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013023\
 Data File : BF132215.D
 Acq On : 30 Jan 2023 17:28
 Operator : CG\JU
 Sample : 01318-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DR-1-012523

Quant Time: Jan 31 00:31:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jan 28 02:25:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.078	152	211983	20.000	ng	0.00
21) Naphthalene-d8	8.366	136	794756	20.000	ng	0.00
39) Acenaphthene-d10	10.131	164	430034	20.000	ng	0.00
64) Phenanthrene-d10	11.630	188	851245	20.000	ng	0.00
76) Chrysene-d12	14.289	240	497789	20.000	ng	0.00
86) Perylene-d12	15.877	264	409675	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.713	112	1512101	118.682	ng	0.02
7) Phenol-d6	6.695	99	1735110	108.613	ng	0.00
23) Nitrobenzene-d5	7.642	82	1270818	83.712	ng	0.00
42) 2,4,6-Tribromophenol	10.925	330	633754	159.153	ng	0.00
45) 2-Fluorobiphenyl	9.442	172	2253622	77.112	ng	0.00
79) Terphenyl-d14	13.224	244	2954649	93.254	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF013023\
Data File : BF132215.D
Acq On : 30 Jan 2023 17:28
Operator : CG\JU
Sample : 01318-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DR-1-012523

Quant Time: Jan 31 00:31:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF012723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jan 28 02:25:45 2023
Response via : Initial Calibration

