

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF020419\
 Data File : BF112386.D
 Acq On : 4 Feb 2019 18:00
 Operator : JU/SJ
 Sample : K1292-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 WC-1

Quant Time: Feb 05 01:19:49 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	139927	20.00	ng	-0.01
21) Naphthalene-d8	8.11	136	526164	20.00	ng	-0.02
39) Acenaphthene-d10	9.86	164	256280	20.00	ng	-0.02
64) Phenanthrene-d10	11.35	188	511871	20.00	ng	-0.02
76) Chrysene-d12	14.00	240	461242	20.00	ng	-0.01
87) Perylene-d12	15.46	264	348427	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.48	112	995783	151.16	ng	0.00
7) Phenol-d6	6.48	99	1119236	122.27	ng	-0.01
23) Nitrobenzene-d5	7.39	82	861176	94.31	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	434065	145.51	ng	-0.02
45) 2-Fluorobiphenyl	9.19	172	1708973	94.31	ng	-0.02
79) Terphenyl-d14	12.94	244	1986186	81.10	ng	-0.01


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF020419\
Data File : BF112386.D
Acq On : 4 Feb 2019 18:00
Operator : JU/SJ
Sample : K1292-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
WC-1

Quant Time: Feb 05 01:19:49 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.83 min Scan# 807
Delta R.T. -0.01 min
Lab File: BF112386.D
Acq: 4 Feb 2019 18:00

Instrument :
BNA_F
ClientSampleId :
WC-1

Tot Ion:152 Resp: 139927
Ion Ratio Lower Upper
152 100
150 150.9 127.4 191.0
115 54.0 45.5 68.3




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5

2-Fluorophenol
Concen: 151.158 ng
RT: 5.48 min Scan# 577
Delta R.T. 0.00 min
Lab File: BF112386.D
Acq: 4 Feb 2019 18:00

Tgt Ion:112 Resp: 995783
Ion Ratio Lower Upper
112 100
64 53.0 45.7 68.5
63 28.8 23.6 35.4



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 122.268 ng

RT: 6.48 min Scan# 747

Delta R.T. -0.01 min

Lab File: BF112386.D

Acq: 4 Feb 2019 18:00

Instrument :

BNA_F

ClientSampled :

WC-1


Tot Ion: 99 Resp: 1119236

Ion Ratio Lower Upper




99 100

42 18.5 15.1 22.7

71 36.8 26.9 40.3



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.11 min Scan# 1024

Delta R.T. -0.02 min

Lab File: BF112386.D

Acq: 4 Feb 2019 18:00

Tgt Ion: 136 Resp: 526164


Ion Ratio Lower Upper

136 100



137 11.1 9.1 13.7


54 6.4 5.9 8.9

68 4.8 5.1 7.7#



Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1









#23
 Nitrobenzene-d5
 Concen: 94.307 ng
 RT: 7.39 min Scan# 902
 Delta R.T. -0.02 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Instrument :
 BNA_F
 ClientSampleId :
 WC-1

Tot Ion: 82 Resp: 861176
 Ion Ratio Lower Upper
 82 100
 128 47.1 37.8 56.8
 54 48.8 39.7 59.5




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): E
 Ion 54.00 (53.70 to 54.70): BF1






#39
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.86 min Scan# 1322
 Delta R.T. -0.02 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Tgt Ion:164 Resp: 256280
 Ion Ratio Lower Upper
 164 100
 162 103.1 82.2 123.4
 160 44.7 36.2 54.4



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#42
 2,4,6-Tribromophenol
 Concen: 145.512 ng
 RT: 10.66 min Scan# 1458
 Delta R.T. -0.02 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Instrument :
 BNA_F
 ClientSampled :
 WC-1

Tot Ion:330 Resp: 434065
 Ion Ratio Lower Upper
 330 100
 332 97.2 76.6 114.8
 141 32.0 24.3 36.5




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#45
 2-Fluorobiphenyl
 Concen: 94.313 ng
 RT: 9.19 min Scan# 1207
 Delta R.T. -0.02 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Tgt Ion:172 Resp: 1708973
 Ion Ratio Lower Upper
 172 100
 171 37.0 30.5 45.7
 170 24.4 20.2 30.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.35 min Scan# 1575
Delta R.T. -0.02 min
Lab File: BF112386.D
Acq: 4 Feb 2019 18:00




Instrument :
BNA_F
ClientSampleId :
WC-1


Tot Ion:188 Resp: 511871
Ion Ratio Lower Upper
188 100
94 8.8 8.2 12.2
80 10.0 8.9 13.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.00 min Scan# 2025
Delta R.T. -0.01 min
Lab File: BF112386.D
Acq: 4 Feb 2019 18:00

Tgt Ion:240 Resp: 461242
Ion Ratio Lower Upper
240 100
120 9.3 9.0 13.6
236 25.8 20.7 31.1









#79
 Terphenyl-d14
 Concen: 81.104 ng
 RT: 12.94 min Scan# 1845
 Delta R.T. -0.01 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Instrument :
 BNA_F
 ClientSampled :
 WC-1

Tot Ion:244 Resp: 1986186

Ion	Ratio	Lower	Upper
244	100		
212	7.3	5.9	8.9
122	11.8	9.8	14.6



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.46 min Scan# 2274
 Delta R.T. -0.01 min
 Lab File: BF112386.D
 Acq: 4 Feb 2019 18:00

Tgt Ion:264 Resp: 348427

Ion	Ratio	Lower	Upper
264	100		
260	23.6	18.2	27.2
265	21.8	17.0	25.4

