


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020519\
 Data File : BF112392.D
 Acq On : 5 Feb 2019 13:15
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 28 10:38:18 2019



AutoFind: Scans 1617, 1618, 1619; Background Corrected with Scan 1612

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	35.8	36760	PASS
68	69	0.00	2	1.5	575	PASS
69	198	0.00	100	38.2	39149	PASS
70	69	0.00	2	0.3	107	PASS
127	198	10	80	49.9	51189	PASS
197	198	0.00	2	0.5	519	PASS
198	198	100	100	100.0	102597	PASS
199	198	5	9	6.5	6688	PASS
275	198	10	60	27.1	27837	PASS
365	198	1	100	3.6	3724	PASS
441	198	0.01	100	12.7	12987	PASS
442	198	50	100	75.2	77195	PASS
443	442	15	24	19.4	14985	PASS