

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF020719\
 Data File : BF112454.D
 Acq On : 7 Feb 2019 22:32
 Operator : JU/SJ
 Sample : K1396-01
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 23935

Manual Integrations
 APPROVED

Sohil
 2/11/2019 4:50:10 PM

Quant Time: Feb 09 04:02:13 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	130977	20.00	ng	-0.01
21) Naphthalene-d8	8.12	136	492035	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	216243	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	343749	20.00	ng	-0.01
76) Chrysene-d12	14.01	240	267207	20.00	ng	0.00
87) Perylene-d12	15.48	264	249738	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	707316	114.71	ng	0.00
7) Phenol-d6	6.49	99	872117	101.78	ng	0.00
23) Nitrobenzene-d5	7.39	82	531860	62.28	ng	-0.02
42) 2,4,6-Tribromophenol	10.66	330	193878	77.03	ng	-0.02
45) 2-Fluorobiphenyl	9.19	172	847234	55.41	ng	-0.02
79) Terphenyl-d14	12.94	244	538984	37.99	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.58	163	85116	5.066	ng	98
71) Phenanthrene	11.38	178	53014	2.966	ng	97
75) Fluoranthene	12.57	202	148313	7.738	ng	97
78) Pyrene	12.80	202	118567	6.409	ng	98
81) Benzo(a)anthracene	14.00	228	54227	3.452	ng	97
83) Chrysene	14.03	228	74791	4.988	ng	97
84) Bis(2-ethylhexyl)phthalate	13.97	149	59151	4.656	ng	98
86) Indeno(1,2,3-cd)pyrene	16.97	276	34363	2.892	ng	94
88) Benzo(b)fluoranthene	15.06	252	98455m	6.592	ng	
90) Benzo(a)pyrene	15.42	252	53745	3.999	ng	# 66
92) Benzo(g,h,i)perylene	17.43	276	34439	3.370	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF020719\
Data File : BF112454.D
Acq On : 7 Feb 2019 22:32
Operator : JU/SJ
Sample : K1396-01
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
23935

Manual Integrations
APPROVED

Sohil
2/11/2019 4:50:10 PM

Quant Time: Feb 09 04:02:13 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration

