




Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020821\
Data File : BF123141.D
Acq On : 8 Feb 2021 9:58
Operator : JU/CG
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 08 17:11:01 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 29 12:29:00 2021
Response via : Initial Calibration



TIC: BF123141.D\data.ms

(77) Benzidine




12.774min (+ 0.006) 0.00 ng

response 1789131

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.70	14.72
183.00	12.10	12.21
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF020821\
 Data File : BF123141.D
 Acq On : 8 Feb 2021 9:58
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Feb 08 17:11:01 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012721.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 29 12:29:00 2021
 Response via : Initial Calibration



TIC: BF123141.D\data.ms

(70) Pentachlorophenol (C)

11.233min (-0.000) 32661.55 ng

response 367050

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.10	65.23
264.00	62.30	62.72
0.00	0.00	0.00