



Data Path : Z:\svoasrv\HPCHEM1\BNA\_F\Data\BF020821\  
 Data File : BF123141.D  
 Acq On : 8 Feb 2021 9:58  
 Operator : JU/CG  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA\_F\METHODS\8270-BF012721.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Jan 29 12:29:00 2021



AutoFind: Scans 1631, 1632, 1633; Background Corrected with Scan 1625

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 35.0         | 108779     | PASS                |
| 68             | 69              | 0.00            | 2               | 2.0          | 2149       | PASS                |
| 69             | 198             | 0.00            | 100             | 35.2         | 109539     | PASS                |
| 70             | 69              | 0.00            | 2               | 0.7          | 717        | PASS                |
| 127            | 198             | 10              | 80              | 47.5         | 147541     | PASS                |
| 197            | 198             | 0.00            | 2               | 0.2          | 502        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 310763     | PASS                |
| 199            | 198             | 5               | 9               | 6.7          | 20672      | PASS                |
| 275            | 198             | 10              | 60              | 29.7         | 92171      | PASS                |
| 365            | 198             | 1               | 100             | 3.6          | 11207      | PASS                |
| 441            | 198             | 0.01            | 100             | 16.4         | 50835      | PASS                |
| 442            | 442             | 50              | 100             | 100.0        | 329941     | PASS                |
| 443            | 442             | 15              | 24              | 19.0         | 62720      | PASS                |