

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021119\
 Data File : BF112481.D
 Acq On : 11 Feb 2019 15:28
 Operator : JU/SJ
 Sample : K1356-07
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SU-03-020819-B

Manual Integrations
 APPROVED

Sohil
 2/12/2019 11:57:03 AM

Quant Time: Feb 11 17:35:09 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jan 28 10:38:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	161917	20.00	ng	-0.02
21) Naphthalene-d8	8.12	136	632232	20.00	ng	-0.01
39) Acenaphthene-d10	9.87	164	303253	20.00	ng	-0.02
64) Phenanthrene-d10	11.36	188	529228	20.00	ng	-0.01
76) Chrysene-d12	14.00	240	334435	20.00	ng	0.00
87) Perylene-d12	15.47	264	334061	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.47	112	892613	117.10	ng	0.00
7) Phenol-d6	6.49	99	1109382	104.73	ng	0.00
23) Nitrobenzene-d5	7.40	82	711588	64.85	ng	-0.01
42) 2,4,6-Tribromophenol	10.66	330	322593	91.39	ng	-0.02
45) 2-Fluorobiphenyl	9.19	172	1429740	66.68	ng	-0.02
79) Terphenyl-d14	12.94	244	1158617	65.25	ng	-0.01
Target Compounds						
10) Phenol	6.49	94	82839	7.128	ng	90
50) Dimethylphthalate	9.58	163	138939	5.897	ng	98
75) Fluoranthene	12.57	202	64685	2.192	ng	99
78) Pyrene	12.80	202	62936	2.718	ng	99
88) Benzo(b)fluoranthene	15.04	252	43373m	2.171	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF021119\
Data File : BF112481.D
Acq On : 11 Feb 2019 15:28
Operator : JU/SJ
Sample : K1356-07
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SU-03-020819-B

Manual Integrations
APPROVED

Sohil
2/12/2019 11:57:03 AM

Quant Time: Feb 11 17:35:09 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF012519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jan 28 10:38:18 2019
Response via : Initial Calibration

