


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021319\
 Data File : BF112527.D
 Acq On : 13 Feb 2019 9:42
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 28 10:38:18 2019



AutoFind: Scans 1618, 1619, 1620; Background Corrected with Scan 1613

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.5	36544	PASS
68	69	0.00	2	1.8	724	PASS
69	198	0.00	100	40.1	40197	PASS
70	69	0.00	2	0.3	124	PASS
127	198	10	80	52.3	52437	PASS
197	198	0.00	2	0.2	160	PASS
198	198	100	100	100.0	100245	PASS
199	198	5	9	7.1	7109	PASS
275	198	10	60	28.2	28253	PASS
365	198	1	100	3.3	3317	PASS
441	198	0.01	100	12.0	12011	PASS
442	198	50	100	76.6	76820	PASS
443	442	15	24	19.0	14611	PASS