


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF021919\
 Data File : BF112611.D
 Acq On : 19 Feb 2019 10:34
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF012519.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Jan 28 10:38:18 2019



AutoFind: Scans 1615, 1616, 1617; Background Corrected with Scan 1610

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.7	36864	PASS
68	69	0.00	2	0.7	266	PASS
69	198	0.00	100	41.2	40232	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	53.6	52408	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	97709	PASS
199	198	5	9	6.9	6752	PASS
275	198	10	60	28.3	27612	PASS
365	198	1	100	4.0	3870	PASS
441	198	0.01	100	12.2	11905	PASS
442	198	50	100	78.6	76800	PASS
443	442	15	24	19.4	14884	PASS