

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022223\
 Data File : BF132508.D
 Acq On : 22 Feb 2023 16:30
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022223

Quant Time: Feb 23 01:01:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 23 00:59:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	42.318	-5.8	97	0.00
3	Pyridine	40.000	41.579	-3.9	96	0.00
4	n-Nitrosodimethylamine	40.000	41.925	-4.8	94	0.00
5 S	2-Fluorophenol	80.000	81.967	-2.5	97	0.00
6	Aniline	40.000	42.561	-6.4	97	0.00
7 S	Phenol-d6	80.000	81.161	-1.5	96	0.00
8	2-Chlorophenol	40.000	41.385	-3.5	97	0.00
9	Benzaldehyde	40.000	39.018	2.5	98	0.00
10 C	Phenol	40.000	41.483	-3.7	97	0.00
11	bis(2-Chloroethyl)ether	40.000	41.460	-3.7	96	0.00
12	1,3-Dichlorobenzene	40.000	40.820	-2.1	96	0.00
13 C	1,4-Dichlorobenzene	40.000	40.702	-1.8	96	0.00
14	1,2-Dichlorobenzene	40.000	38.650	3.4	95	0.00
15	Benzyl Alcohol	40.000	42.124	-5.3	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.865	-2.2	95	0.00
17	2-Methylphenol	40.000	40.679	-1.7	95	0.00
18	Hexachloroethane	40.000	40.623	-1.6	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.023	2.4	94	0.00
20	3+4-Methylphenols	40.000	36.557	8.6	97	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	96	0.00
22	Acetophenone	40.000	39.534	1.2	95	0.00
23 S	Nitrobenzene-d5	80.000	82.108	-2.6	96	0.00
24	Nitrobenzene	40.000	41.621	-4.1	96	0.00
25	Isophorone	40.000	40.798	-2.0	96	0.00
26 C	2-Nitrophenol	40.000	42.176	-5.4	96	0.00
27	2,4-Dimethylphenol	40.000	40.580	-1.4	94	0.00
28	bis(2-Chloroethoxy)methane	40.000	41.322	-3.3	97	0.00
29 C	2,4-Dichlorophenol	40.000	41.480	-3.7	96	0.00
30	1,2,4-Trichlorobenzene	40.000	40.867	-2.2	95	0.00
31	Naphthalene	40.000	40.435	-1.1	97	0.00
32	Benzoic acid	40.000	40.969	-2.4	96	0.00
33	4-Chloroaniline	40.000	41.735	-4.3	96	0.00
34 C	Hexachlorobutadiene	40.000	41.389	-3.5	96	0.00
35	Caprolactam	40.000	42.579	-6.4	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.129	-2.8	95	0.00
37	2-Methylnaphthalene	40.000	40.014	-0.0	95	0.00
38	1-Methylnaphthalene	40.000	40.199	-0.5	96	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.106	-2.8	96	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.908	-12.3	94	0.00
42 S	2,4,6-Tribromophenol	80.000	83.343	-4.2	96	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.033	-5.1	96	0.00
44	2,4,5-Trichlorophenol	40.000	42.461	-6.2	97	0.00
45 S	2-Fluorobiphenyl	80.000	75.424	5.7	96	0.00
46	1,1'-Biphenyl	40.000	40.521	-1.3	96	0.00
47	2-Chloronaphthalene	40.000	40.814	-2.0	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022223\
 Data File : BF132508.D
 Acq On : 22 Feb 2023 16:30
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF022223

Quant Time: Feb 23 01:01:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Feb 23 00:59:42 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.480	-3.7	95	0.00
49	Acenaphthylene	40.000	40.286	-0.7	95	0.00
50	Dimethylphthalate	40.000	41.063	-2.7	96	0.00
51	2,6-Dinitrotoluene	40.000	41.699	-4.2	95	0.00
52 C	Acenaphthene	40.000	41.233	-3.1	96	0.00
53	3-Nitroaniline	40.000	41.955	-4.9	96	0.00
54 P	2,4-Dinitrophenol	40.000	43.060	-7.7	95	0.00
55	Dibenzofuran	40.000	40.704	-1.8	96	0.00
56 P	4-Nitrophenol	40.000	44.875	-12.2	99	0.00
57	2,4-Dinitrotoluene	40.000	41.728	-4.3	95	0.00
58	Fluorene	40.000	40.485	-1.2	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.281	-5.7	95	0.00
60	Diethylphthalate	40.000	40.957	-2.4	95	0.00
61	4-Chlorophenyl-phenylether	40.000	40.887	-2.2	97	0.00
62	4-Nitroaniline	40.000	42.694	-6.7	96	0.00
63	Azobenzene	40.000	41.542	-3.9	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.373	-13.4	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.526	-3.8	97	0.00
67	4-Bromophenyl-phenylether	40.000	41.780	-4.5	97	0.00
68	Hexachlorobenzene	40.000	41.499	-3.7	96	0.00
69	Atrazine	40.000	41.749	-4.4	96	0.00
70 C	Pentachlorophenol	40.000	43.593	-9.0	93	0.00
71	Phenanthrene	40.000	40.521	-1.3	96	0.00
72	Anthracene	40.000	40.636	-1.6	96	0.00
73	Carbazole	40.000	40.639	-1.6	96	0.00
74	Di-n-butylphthalate	40.000	41.062	-2.7	95	0.00
75 C	Fluoranthene	40.000	40.736	-1.8	95	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	43.193	-8.0	80	0.00
78	Pyrene	40.000	44.104	-10.3	95	0.00
79 S	Terphenyl-d14	80.000	84.851	-6.1	94	0.00
80	Butylbenzylphthalate	40.000	43.718	-9.3	92	0.00
81	Benzo(a)anthracene	40.000	42.009	-5.0	91	0.00
82	3,3'-Dichlorobenzidine	40.000	39.766	0.6	89	0.00
83	Chrysene	40.000	41.208	-3.0	93	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.023	-7.6	91	0.00
85 c	Di-n-octyl phthalate	40.000	40.896	-2.2	88	0.00
86 I	Perylene-d12	20.000	20.000	0.0	93	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	45.261	-13.2	96	0.00
88	Benzo(b)fluoranthene	40.000	41.486	-3.7	93	0.00
89	Benzo(k)fluoranthene	40.000	38.985	2.5	89	0.00
90 C	Benzo(a)pyrene	40.000	41.215	-3.0	92	0.00
91	Dibenzo(a,h)anthracene	40.000	45.155	-12.9	95	0.00
92	Benzo(g,h,i)perylene	40.000	42.115	-5.3	96	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022223\
Data File : BF132508.D
Acq On : 22 Feb 2023 16:30
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF022223

Quant Time: Feb 23 01:01:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Feb 23 00:59:42 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0