

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022521\
 Data File : BF123297.D
 Acq On : 25 Feb 2021 13:52
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 25 14:39:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.802	0.770	4.0	100	-0.02
3	Pyridine	1.954	1.609	17.7	78	-0.02
4	n-Nitrosodimethylamine	0.903	0.784	13.2	85	-0.02
5 S	2-Fluorophenol	1.389	1.332	4.1	106	0.00
6	Aniline	2.230	2.102	5.7	104	0.00
7 S	Phenol-d6	1.888	1.812	4.0	106	0.00
8	2-Chlorophenol	1.493	1.462	2.1	107	0.00
9	Benzaldehyde	1.075	1.011	6.0	110	0.00
10 C	Phenol	2.057	1.960	4.7	106	0.00
11	bis(2-Chloroethyl)ether	1.476	1.429	3.2	109	0.00
12	1,3-Dichlorobenzene	1.553	1.519	2.2	106	0.00
13 C	1,4-Dichlorobenzene	1.582	1.547	2.2	106	0.00
14	1,2-Dichlorobenzene	1.472	1.454	1.2	108	0.00
15	Benzyl Alcohol	1.472	1.498	-1.8	110	0.00
16	2,2'-oxybis(1-Chloropropane	2.024	1.913	5.5	109	0.00
17	2-Methylphenol	1.256	1.217	3.1	108	0.00
18	Hexachloroethane	0.572	0.598	-4.5	111	0.00
19 P	n-Nitroso-di-n-propylamine	1.123	1.147	-2.1	111	0.00
20	3+4-Methylphenols	1.649	1.525	7.5	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	108	0.00
22	Acetophenone	0.519	0.507	2.3	109	0.00
23 S	Nitrobenzene-d5	0.437	0.437	0.0	110	0.00
24	Nitrobenzene	0.461	0.460	0.2	111	0.00
25	Isophorone	0.817	0.819	-0.2	112	0.00
26 C	2-Nitrophenol	0.193	0.201	-4.1	111	0.00
27	2,4-Dimethylphenol	0.299	0.290	3.0	107	0.00
28	bis(2-Chloroethoxy)methane	0.428	0.425	0.7	111	0.00
29 C	2,4-Dichlorophenol	0.308	0.301	2.3	108	0.00
30	1,2,4-Trichlorobenzene	0.335	0.334	0.3	109	0.00
31	Naphthalene	1.096	1.076	1.8	109	0.00
32	Benzoic acid	0.220	0.191	13.2	90	0.01
33	4-Chloroaniline	0.446	0.432	3.1	108	0.00
34 C	Hexachlorobutadiene	0.189	0.194	-2.6	112	0.00
35	Caprolactam	0.103	0.099	3.9	108	0.00
36 C	4-Chloro-3-methylphenol	0.367	0.364	0.8	110	0.00
37	2-Methylnaphthalene	0.730	0.706	3.3	107	0.00
38	1-Methylnaphthalene	0.683	0.664	2.8	108	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
40	1,2,4,5-Tetrachlorobenzene	0.598	0.598	0.0	111	0.00
41 P	Hexachlorocyclopentadiene	0.280	0.278	0.7	99	0.00
42 S	2,4,6-Tribromophenol	0.204	0.207	-1.5	109	0.00
43 C	2,4,6-Trichlorophenol	0.424	0.421	0.7	108	0.00
44	2,4,5-Trichlorophenol	0.454	0.446	1.8	108	0.00
45 S	2-Fluorobiphenyl	1.392	1.351	2.9	110	0.00
46	1,1'-Biphenyl	1.659	1.612	2.8	109	0.00
47	2-Chloronaphthalene	1.309	1.282	2.1	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022521\
 Data File : BF123297.D
 Acq On : 25 Feb 2021 13:52
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 25 14:39:00 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 19 16:08:53 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.469	0.486	-3.6	113	0.00
49	Acenaphthylene	1.986	1.930	2.8	109	0.00
50	Dimethylphthalate	1.496	1.494	0.1	113	0.00
51	2,6-Dinitrotoluene	0.345	0.351	-1.7	111	0.00
52 C	Acenaphthene	1.369	1.356	0.9	110	0.00
53	3-Nitroaniline	0.383	0.378	1.3	108	0.00
54 P	2,4-Dinitrophenol	0.181	0.181	0.0	105	0.00
55	Dibenzofuran	1.854	1.780	4.0	108	0.00
56 P	4-Nitrophenol	0.305	0.281	7.9	96	0.00
57	2,4-Dinitrotoluene	0.464	0.473	-1.9	111	0.00
58	Fluorene	1.459	1.437	1.5	110	0.00
59	2,3,4,6-Tetrachlorophenol	0.362	0.361	0.3	107	0.00
60	Diethylphthalate	1.481	1.491	-0.7	113	0.00
61	4-Chlorophenyl-phenylether	0.694	0.688	0.9	111	0.00
62	4-Nitroaniline	0.385	0.375	2.6	107	0.00
63	Azobenzene	1.707	1.663	2.6	111	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.135	0.145	-7.4	107	0.00
66 c	n-Nitrosodiphenylamine	0.692	0.676	2.3	109	0.00
67	4-Bromophenyl-phenylether	0.218	0.219	-0.5	111	0.00
68	Hexachlorobenzene	0.230	0.230	0.0	109	0.00
69	Atrazine	0.219	0.216	1.4	109	0.00
70 C	Pentachlorophenol	0.145	0.149	-2.8	106	0.00
71	Phenanthrene	1.183	1.142	3.5	108	0.00
72	Anthracene	1.180	1.150	2.5	109	0.00
73	Carbazole	1.111	1.062	4.4	107	0.00
74	Di-n-butylphthalate	1.243	1.227	1.3	111	0.00
75 C	Fluoranthene	1.254	1.197	4.5	105	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzydine	0.646	0.595	7.9	100	0.00
78	Pyrene	1.561	1.628	-4.3	107	0.00
79 S	Terphenyl-d14	1.040	1.093	-5.1	108	0.00
80	Butylbenzylphthalate	0.660	0.698	-5.8	108	0.00
81	Benzo(a)anthracene	1.359	1.347	0.9	101	0.00
82	3,3'-Dichlorobenzidine	0.452	0.455	-0.7	102	0.00
83	Chrysene	1.335	1.320	1.1	100	0.00
84	Bis(2-ethylhexyl)phthalate	0.909	0.913	-0.4	105	0.00
85 c	Di-n-octyl phthalate	1.552	1.461	5.9	98	0.00
86 I	Perylene-d12	1.000	1.000	0.0	86	0.00
87	Indeno(1,2,3-cd)pyrene	1.382	1.249	9.6	79	0.00
88	Benzo(b)fluoranthene	1.405	1.517	-8.0	94	0.00
89	Benzo(k)fluoranthene	1.274	1.323	-3.8	87	0.00
90 C	Benzo(a)pyrene	1.253	1.278	-2.0	88	0.00
91	Dibenzo(a,h)anthracene	1.186	1.064	10.3	79	-0.01
92	Benzo(g,h,i)perylene	1.092	0.993	9.1	76	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022521\
Data File : BF123297.D
Acq On : 25 Feb 2021 13:52
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 25 14:39:00 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF021921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 19 16:08:53 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0