

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022719\
 Data File : BF112729.D
 Acq On : 27 Feb 2019 19:56
 Operator : JU/SJ
 Sample : K1627-07 2X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-04-(0-2)

Manual Integrations
 APPROVED

Sohil
 2/28/2019 12:47:13 PM

Quant Time: Feb 28 03:13:35 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 27 16:08:50 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.74	152	189750	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	664975	20.00	ng	0.00
39) Acenaphthene-d10	9.77	164	269416	20.00	ng	0.00
64) Phenanthrene-d10	11.25	188	517754	20.00	ng	0.00
76) Chrysene-d12	13.87	240	503195	20.00	ng	0.00
87) Perylene-d12	15.27	264	397020	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	559111	45.84	ng	0.00
7) Phenol-d6	6.36	99	684351	44.66	ng	0.00
23) Nitrobenzene-d5	7.29	82	385950	34.73	ng	-0.01
42) 2,4,6-Tribromophenol	10.55	330	138406	48.39	ng	0.00
45) 2-Fluorobiphenyl	9.09	172	591549	28.99	ng	0.00
79) Terphenyl-d14	12.83	244	610856	23.53	ng	0.00
Target Compounds						
50) Dimethylphthalate	9.49	163	44877	2.259	ng	Qvalue # 99
71) Phenanthrene	11.27	178	169805	6.592	ng	99
75) Fluoranthene	12.45	202	383962	13.912	ng	97
78) Pyrene	12.67	202	312062	7.356	ng	98
81) Benzo(a)anthracene	13.86	228	171853	4.928	ng	98
83) Chrysene	13.89	228	176291	5.161	ng	97
84) Bis(2-ethylhexyl)phthalate	13.87	149	358234	16.311	ng	99
86) Indeno(1,2,3-cd)pyrene	16.62	276	60130	2.065	ng	98
88) Benzo(b)fluoranthene	14.87	252	194053m	7.556	ng	
89) Benzo(k)fluoranthene	14.90	252	62956m	2.706	ng	
90) Benzo(a)pyrene	15.22	252	108759	4.788	ng	96
92) Benzo(a,h,i)perylene	17.03	276	56353	2.895	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022719\
Data File : BF112729.D
Acq On : 27 Feb 2019 19:56
Operator : JU/SJ
Sample : K1627-07 2X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-04-(0-2)

Manual Integrations
APPROVED

Sohil
2/28/2019 12:47:13 PM

Quant Time: Feb 28 03:13:35 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 27 16:08:50 2019
Response via : Initial Calibration

