


2,4-Dinitrophenol


Response Ratio



Resp Ratio = $2.47 \times 10^{-1} \times \text{Amt} - 7.11 \times 10^{-2}$
Coef of Det (r^2) = 0.998 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM021119.M
Calibration Table Last Updated: Mon Feb 11 16:02:12 2019

4,6-Dinitro-2-methylphenol




Resp Ratio = 1.40×10^{-1} * Amt
RF Rel Std Dev = 11.5% Curve Fit: Avg RF

Method Name: Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\8270-BM021119.M
Calibration Table Last Updated: Mon Feb 11 16:02:12 2019

2-Fluorobiphenyl

Response Ratio



$R = -6.91e-002 A^2 + 1.47e+000 A + 1.13e-001$
Coef of Det (r^2) = 0.999 Curve Fit: Quadratic

Method Name: Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
Calibration Table Last Updated: Fri Mar 01 15:40:43 2019