

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022719\
 Data File : BF112719.D
 Acq On : 27 Feb 2019 15:04
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 28 01:58:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 27 16:08:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.644	0.646	-0.3	107	0.00
3	Pyridine	1.724	1.804	-4.6	112	0.00
4	n-Nitrosodimethylamine	0.754	0.763	-1.2	106	0.00
5 S	2-Fluorophenol	1.286	1.233	4.1	104	0.00
6	Aniline	2.223	2.178	2.0	104	0.00
7 S	Phenol-d6	1.615	1.560	3.4	105	0.00
8	2-Chlorophenol	1.431	1.414	1.2	106	0.00
9	Benzaldehyde	1.164	1.141	2.0	105	0.00
10 C	Phenol	1.885	1.903	-1.0	107	0.00
11	bis(2-Chloroethyl)ether	1.447	1.430	1.2	106	0.00
12	1,3-Dichlorobenzene	1.479	1.453	1.8	107	0.00
13 C	1,4-Dichlorobenzene	1.483	1.449	2.3	106	0.00
14	1,2-Dichlorobenzene	1.359	1.325	2.5	107	0.00
15	Benzyl Alcohol	1.232	1.233	-0.1	106	0.00
16	2,2'-oxybis(1-Chloropropane	2.414	2.384	1.2	106	0.00
17	2-Methylphenol	1.161	1.142	1.6	107	0.00
18	Hexachloroethane	0.568	0.564	0.7	106	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	1.004	1.9	107	0.00
20	3+4-Methylphenols	1.311	1.262	3.7	106	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	106	0.00
22	Acetophenone	0.468	0.469	-0.2	106	0.00
23 S	Nitrobenzene-d5	0.334	0.345	-3.3	108	0.00
24	Nitrobenzene	0.372	0.379	-1.9	107	0.00
25	Isophorone	0.720	0.713	1.0	108	0.00
26 C	2-Nitrophenol	0.155	0.170	-9.7	109	0.00
27	2,4-Dimethylphenol	0.288	0.285	1.0	107	0.00
28	bis(2-Chloroethoxy)methane	0.450	0.446	0.9	108	0.00
29 C	2,4-Dichlorophenol	0.279	0.280	-0.4	106	0.00
30	1,2,4-Trichlorobenzene	0.300	0.298	0.7	108	0.00
31	Naphthalene	0.993	0.963	3.0	106	0.00
32	Benzoic acid	0.202	0.176	12.9	92	0.00
33	4-Chloroaniline	0.421	0.414	1.7	106	0.00
34 C	Hexachlorobutadiene	0.169	0.172	-1.8	109	0.00
35	Caprolactam	0.098	0.101	-3.1	108	0.00
36 C	4-Chloro-3-methylphenol	0.309	0.310	-0.3	107	0.00
37	2-Methylnaphthalene	0.641	0.629	1.9	107	0.00
38	1-Methylnaphthalene	0.621	0.611	1.6	107	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	0.583	0.569	2.4	107	0.00
41 P	Hexachlorocyclopentadiene	0.319	0.319	0.0	107	0.00
42 S	2,4,6-Tribromophenol	0.212	0.215	-1.4	109	0.00
43 C	2,4,6-Trichlorophenol	0.401	0.407	-1.5	108	0.00
44	2,4,5-Trichlorophenol	0.434	0.437	-0.7	108	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022719\
 Data File : BF112719.D
 Acq On : 27 Feb 2019 15:04
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 28 01:58:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 27 16:08:50 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.274	1.215	4.6	107	0.00
46	1,1'-Biphenyl	1.631	1.632	-0.1	107	0.00
47	2-Chloronaphthalene	1.274	1.258	1.3	110	0.00
48	2-Nitroaniline	0.387	0.421	-8.8	111	0.00
49	Acenaphthylene	2.162	2.091	3.3	108	0.00
50	Dimethylphthalate	1.475	1.439	2.4	108	0.00
51	2,6-Dinitrotoluene	0.307	0.330	-7.5	110	0.00
52 C	Acenaphthene	1.265	1.226	3.1	107	0.00
53	3-Nitroaniline	0.374	0.391	-4.5	108	0.00
54 P	2,4-Dinitrophenol	0.108	0.115	-6.5	112	0.00
55	Dibenzofuran	1.838	1.763	4.1	107	0.00
56 P	4-Nitrophenol	0.304	0.331	-8.9	109	0.00
57	2,4-Dinitrotoluene	0.382	0.431	-12.8	114	0.00
58	Fluorene	1.304	1.238	5.1	106	0.00
59	2,3,4,6-Tetrachlorophenol	0.361	0.370	-2.5	109	0.00
60	Diethylphthalate	1.558	1.528	1.9	109	0.00
61	4-Chlorophenyl-phenylether	0.607	0.589	3.0	108	0.00
62	4-Nitroaniline	0.346	0.363	-4.9	111	0.00
63	Azobenzene	1.595	1.546	3.1	107	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
65	4,6-Dinitro-2-methylphenol	0.076	0.083	-9.2	114	0.00
66 c	n-Nitrosodiphenylamine	0.641	0.622	3.0	107	0.00
67	4-Bromophenyl-phenylether	0.211	0.208	1.4	109	0.00
68	Hexachlorobenzene	0.237	0.237	0.0	110	0.00
69	Atrazine	0.196	0.190	3.1	107	0.00
70 C	Pentachlorophenol	0.140	0.143	-2.1	107	0.00
71	Phenanthrene	0.995	0.992	0.3	108	0.00
72	Anthracene	1.042	1.005	3.6	108	0.00
73	Carbazole	1.016	0.971	4.4	107	0.00
74	Di-n-butylphthalate	1.218	1.179	3.2	109	0.00
75 C	Fluoranthene	1.066	1.067	-0.1	108	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
77	Benzidine	1.038	1.016	2.1	103	0.00
78	Pyrene	1.686	1.699	-0.8	108	0.00
79 S	Terphenyl-d14	1.032	1.008	2.3	108	0.00
80	Butylbenzylphthalate	0.744	0.775	-4.2	109	0.00
81	Benzo(a)anthracene	1.386	1.385	0.1	107	0.00
82	3,3'-Dichlorobenzidine	0.524	0.536	-2.3	106	0.00
83	Chrysene	1.358	1.352	0.4	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.873	0.882	-1.0	108	0.00
85 c	Di-n-octyl phthalate	1.499	1.548	-3.3	111	0.00
86	Indeno(1,2,3-cd)pyrene	1.158	1.056	8.8	105	0.00
87 I	Perylene-d12	1.000	1.000	0.0	109	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF022719\
Data File : BF112719.D
Acq On : 27 Feb 2019 15:04
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 28 01:58:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 27 16:08:50 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.294	1.366	-5.6	108	0.00
89	Benzo(k)fluoranthene	1.172	1.102	6.0	109	0.00
90 C	Benzo(a)pyrene	1.144	1.143	0.1	108	0.00
91	Dibenzo(a,h)anthracene	0.976	0.918	5.9	106	0.00
92	Benzo(q,h,i)perylene	0.980	0.912	6.9	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0