


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF022719\
 Data File : BF112718.D
 Acq On : 27 Feb 2019 14:37
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 27 16:08:50 2019



AutoFind: Scans 1603, 1604, 1605; Background Corrected with Scan 1598

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	42.7	55436	PASS
68	69	0.00	2	1.5	710	PASS
69	198	0.00	100	37.6	48900	PASS
70	69	0.00	2	0.5	250	PASS
127	198	10	80	50.3	65349	PASS
197	198	0.00	2	0.3	389	PASS
198	198	100	100	100.0	129954	PASS
199	198	5	9	6.7	8762	PASS
275	198	10	60	26.2	34034	PASS
365	198	1	100	3.1	3998	PASS
441	198	0.01	100	13.6	17627	PASS
442	198	50	100	87.8	114141	PASS
443	442	15	24	18.4	20953	PASS