

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022723\
 Data File : BF132556.D
 Acq On : 27 Feb 2023 17:04
 Operator : CG\JU
 Sample : 01701-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 B-19-SB02

Quant Time: Feb 27 18:00:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.001	152	141477	20.000	ng	0.00
21) Naphthalene-d8	8.289	136	506948	20.000	ng	0.00
39) Acenaphthene-d10	10.048	164	255432	20.000	ng	0.00
64) Phenanthrene-d10	11.548	188	405194	20.000	ng	# 0.00
76) Chrysene-d12	14.201	240	307065	20.000	ng	# 0.00
86) Perylene-d12	15.754	264	234023	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.631	112	928370	106.499	ng	0.00
7) Phenol-d6	6.631	99	1203070	105.749	ng	0.00
23) Nitrobenzene-d5	7.566	82	790773	74.007	ng	0.00
42) 2,4,6-Tribromophenol	10.848	330	270092	97.911	ng	0.00
45) 2-Fluorobiphenyl	9.366	172	1245458	74.243	ng	0.00
79) Terphenyl-d14	13.136	244	1123285	61.850	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022723\
Data File : BF132556.D
Acq On : 27 Feb 2023 17:04
Operator : CG\JU
Sample : 01701-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
B-19-SB02

Quant Time: Feb 27 18:00:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 23:17:30 2023
Response via : Initial Calibration

