

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022723\
 Data File : BF132545.D
 Acq On : 27 Feb 2023 11:00
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 27 13:18:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2	1,4-Dioxane	0.682	0.731	-7.2	87	-0.01
3	Pyridine	1.809	1.801	0.4	81	0.00
4	n-Nitrosodimethylamine	0.683	0.726	-6.3	84	-0.01
5 S	2-Fluorophenol	1.232	1.243	-0.9	84	0.00
6	Aniline	2.164	2.249	-3.9	84	0.00
7 S	Phenol-d6	1.608	1.596	0.7	82	0.00
8	2-Chlorophenol	1.279	1.303	-1.9	84	0.00
9	Benzaldehyde	0.868	0.842	3.0	86	0.00
10 C	Phenol	1.844	1.844	0.0	83	0.00
11	bis(2-Chloroethyl)ether	1.423	1.422	0.1	82	0.00
12	1,3-Dichlorobenzene	1.373	1.396	-1.7	84	0.00
13 C	1,4-Dichlorobenzene	1.370	1.402	-2.3	85	0.00
14	1,2-Dichlorobenzene	1.315	1.274	3.1	84	0.00
15	Benzyl Alcohol	1.239	1.348	-8.8	87	0.00
16	2,2'-oxybis(1-Chloropropane	1.815	1.847	-1.8	83	0.00
17	2-Methylphenol	1.194	1.216	-1.8	84	0.00
18	Hexachloroethane	0.535	0.564	-5.4	86	0.00
19 P	n-Nitroso-di-n-propylamine	0.990	0.964	2.6	83	0.00
20	3+4-Methylphenols	1.542	1.459	5.4	88	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
22	Acetophenone	0.504	0.516	-2.4	85	0.00
23 S	Nitrobenzene-d5	0.422	0.436	-3.3	83	0.00
24	Nitrobenzene	0.451	0.478	-6.0	85	0.00
25	Isophorone	0.808	0.839	-3.8	85	0.00
26 C	2-Nitrophenol	0.199	0.209	-5.0	83	0.00
27	2,4-Dimethylphenol	0.314	0.322	-2.5	82	0.00
28	bis(2-Chloroethoxy)methane	0.473	0.488	-3.2	84	0.00
29 C	2,4-Dichlorophenol	0.312	0.328	-5.1	84	0.00
30	1,2,4-Trichlorobenzene	0.339	0.356	-5.0	84	0.00
31	Naphthalene	1.024	1.049	-2.4	85	0.00
32	Benzoic acid	0.227	0.267	-17.6	88	0.00
33	4-Chloroaniline	0.439	0.465	-5.9	84	0.00
34 C	Hexachlorobutadiene	0.215	0.239	-11.2	89	0.00
35	Caprolactam	0.103	0.111	-7.8	83	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.363	-5.8	85	0.00
37	2-Methylnaphthalene	0.698	0.721	-3.3	85	0.00
38	1-Methylnaphthalene	0.652	0.674	-3.4	86	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
40	1,2,4,5-Tetrachlorobenzene	0.654	0.702	-7.3	88	0.00
41 P	Hexachlorocyclopentadiene	0.355	0.408	-14.9	85	0.00
42 S	2,4,6-Tribromophenol	0.216	0.243	-12.5	91	0.00
43 C	2,4,6-Trichlorophenol	0.443	0.465	-5.0	85	0.00
44	2,4,5-Trichlorophenol	0.518	0.550	-6.2	85	0.00
45 S	2-Fluorobiphenyl	1.313	1.279	2.6	87	0.00
46	1,1'-Biphenyl	1.514	1.552	-2.5	85	0.00
47	2-Chloronaphthalene	1.200	1.243	-3.6	86	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022723\
 Data File : BF132545.D
 Acq On : 27 Feb 2023 11:00
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 27 13:18:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.442	0.460	-4.1	84	0.00
49	Acenaphthylene	1.910	1.962	-2.7	85	0.00
50	Dimethylphthalate	1.456	1.518	-4.3	86	0.00
51	2,6-Dinitrotoluene	0.332	0.352	-6.0	85	0.00
52 C	Acenaphthene	1.213	1.277	-5.3	86	0.00
53	3-Nitroaniline	0.371	0.393	-5.9	85	0.00
54 P	2,4-Dinitrophenol	0.195	0.220	-12.8	84	0.00
55	Dibenzofuran	1.768	1.814	-2.6	85	0.00
56 P	4-Nitrophenol	0.277	0.291	-5.1	81	0.00
57	2,4-Dinitrotoluene	0.441	0.477	-8.2	87	0.00
58	Fluorene	1.353	1.400	-3.5	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.384	0.423	-10.2	87	0.00
60	Diethylphthalate	1.422	1.524	-7.2	87	0.00
61	4-Chlorophenyl-phenylether	0.702	0.739	-5.3	88	0.00
62	4-Nitroaniline	0.364	0.387	-6.3	84	0.00
63	Azobenzene	1.506	1.611	-7.0	87	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
65	4,6-Dinitro-2-methylphenol	0.135	0.151	-11.9	87	0.00
66 c	n-Nitrosodiphenylamine	0.624	0.638	-2.2	86	0.00
67	4-Bromophenyl-phenylether	0.226	0.243	-7.5	89	0.00
68	Hexachlorobenzene	0.238	0.252	-5.9	89	0.00
69	Atrazine	0.195	0.211	-8.2	89	0.00
70 C	Pentachlorophenol	0.142	0.170	-19.7	93	0.00
71	Phenanthrene	1.036	1.061	-2.4	87	0.00
72	Anthracene	1.067	1.088	-2.0	87	0.00
73	Carbazole	0.962	0.979	-1.8	86	0.00
74	Di-n-butylphthalate	1.128	1.196	-6.0	88	0.00
75 C	Fluoranthene	1.133	1.192	-5.2	89	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
77	Benzydine	0.487	0.522	-7.2	78	0.00
78	Pyrene	1.818	1.877	-3.2	88	0.00
79 S	Terphenyl-d14	1.183	1.213	-2.5	89	0.00
80	Butylbenzylphthalate	0.718	0.772	-7.5	89	0.00
81	Benzo(a)anthracene	1.496	1.567	-4.7	90	0.00
82	3,3'-Dichlorobenzidine	0.441	0.450	-2.0	90	0.00
83	Chrysene	1.399	1.437	-2.7	91	0.00
84	Bis(2-ethylhexyl)phthalate	0.881	0.951	-7.9	90	0.00
85 c	Di-n-octyl phthalate	1.288	1.382	-7.3	92	0.00
86 I	Perylene-d12	1.000	1.000	0.0	86	0.00
87	Indeno(1,2,3-cd)pyrene	1.310	1.505	-14.9	90	0.00
88	Benzo(b)fluoranthene	1.340	1.485	-10.8	92	0.00
89	Benzo(k)fluoranthene	1.312	1.321	-0.7	85	0.00
90 C	Benzo(a)pyrene	1.254	1.329	-6.0	87	0.00
91	Dibenzo(a,h)anthracene	1.068	1.226	-14.8	89	0.00
92	Benzo(g,h,i)perylene	1.112	1.292	-16.2	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF022723\
Data File : BF132545.D
Acq On : 27 Feb 2023 11:00
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Feb 27 13:18:38 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 23:17:30 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0