

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030323\
 Data File : BF132623.D
 Acq On : 03 Mar 2023 12:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 02:14:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	114	0.00
2	1,4-Dioxane	0.682	0.636	6.7	102	-0.03
3	Pyridine	1.809	1.660	8.2	100	-0.02
4	n-Nitrosodimethylamine	0.683	0.616	9.8	96	-0.03
5 S	2-Fluorophenol	1.232	1.192	3.2	109	0.00
6	Aniline	2.164	2.262	-4.5	114	0.00
7 S	Phenol-d6	1.608	1.560	3.0	109	0.00
8	2-Chlorophenol	1.279	1.264	1.2	110	0.00
9	Benzaldehyde	0.868	0.867	0.1	120	0.00
10 C	Phenol	1.844	1.815	1.6	110	0.00
11	bis(2-Chloroethyl)ether	1.423	1.427	-0.3	111	0.00
12	1,3-Dichlorobenzene	1.373	1.352	1.5	110	0.00
13 C	1,4-Dichlorobenzene	1.370	1.362	0.6	112	0.00
14	1,2-Dichlorobenzene	1.315	1.227	6.7	109	0.00
15	Benzyl Alcohol	1.239	1.253	-1.1	109	0.00
16	2,2'-oxybis(1-Chloropropane	1.815	1.775	2.2	108	0.00
17	2-Methylphenol	1.194	1.216	-1.8	113	0.00
18	Hexachloroethane	0.535	0.530	0.9	109	0.00
19 P	n-Nitroso-di-n-propylamine	0.990	0.891	10.0	103	0.00
20	3+4-Methylphenols	1.542	1.369	11.2	112	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	115	0.00
22	Acetophenone	0.504	0.469	6.9	107	0.00
23 S	Nitrobenzene-d5	0.422	0.400	5.2	106	0.00
24	Nitrobenzene	0.451	0.434	3.8	107	0.00
25	Isophorone	0.808	0.815	-0.9	114	0.00
26 C	2-Nitrophenol	0.199	0.208	-4.5	114	0.00
27	2,4-Dimethylphenol	0.314	0.312	0.6	111	0.00
28	bis(2-Chloroethoxy)methane	0.473	0.484	-2.3	115	0.00
29 C	2,4-Dichlorophenol	0.312	0.317	-1.6	112	0.00
30	1,2,4-Trichlorobenzene	0.339	0.344	-1.5	113	0.00
31	Naphthalene	1.024	0.990	3.3	111	0.00
32	Benzoic acid	0.227	0.259	-14.1	119	0.00
33	4-Chloroaniline	0.439	0.464	-5.7	117	0.00
34 C	Hexachlorobutadiene	0.215	0.222	-3.3	114	0.00
35	Caprolactam	0.103	0.111	-7.8	116	0.00
36 C	4-Chloro-3-methylphenol	0.343	0.352	-2.6	113	0.00
37	2-Methylnaphthalene	0.698	0.695	0.4	113	0.00
38	1-Methylnaphthalene	0.652	0.642	1.5	113	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	118	0.00
40	1,2,4,5-Tetrachlorobenzene	0.654	0.661	-1.1	116	0.00
41 P	Hexachlorocyclopentadiene	0.355	0.341	3.9	99	0.00
42 S	2,4,6-Tribromophenol	0.216	0.226	-4.6	118	0.00
43 C	2,4,6-Trichlorophenol	0.443	0.455	-2.7	116	0.00
44	2,4,5-Trichlorophenol	0.518	0.519	-0.2	112	0.00
45 S	2-Fluorobiphenyl	1.313	1.166	11.2	111	0.00
46	1,1'-Biphenyl	1.514	1.483	2.0	113	0.00
47	2-Chloronaphthalene	1.200	1.192	0.7	115	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030323\
 Data File : BF132623.D
 Acq On : 03 Mar 2023 12:07
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 04 02:14:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 24 23:17:30 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.442	0.421	4.8	107	0.00
49	Acenaphthylene	1.910	1.874	1.9	113	0.00
50	Dimethylphthalate	1.456	1.485	-2.0	118	0.00
51	2,6-Dinitrotoluene	0.332	0.344	-3.6	116	0.00
52 C	Acenaphthene	1.213	1.220	-0.6	115	0.00
53	3-Nitroaniline	0.371	0.397	-7.0	120	0.00
54 P	2,4-Dinitrophenol	0.195	0.209	-7.2	111	0.00
55	Dibenzofuran	1.768	1.732	2.0	114	0.00
56 P	4-Nitrophenol	0.277	0.284	-2.5	111	0.00
57	2,4-Dinitrotoluene	0.441	0.462	-4.8	118	0.00
58	Fluorene	1.353	1.320	2.4	113	0.00
59	2,3,4,6-Tetrachlorophenol	0.384	0.401	-4.4	115	0.00
60	Diethylphthalate	1.422	1.455	-2.3	116	0.00
61	4-Chlorophenyl-phenylether	0.702	0.705	-0.4	117	0.00
62	4-Nitroaniline	0.364	0.387	-6.3	118	0.00
63	Azobenzene	1.506	1.485	1.4	112	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	119	0.00
65	4,6-Dinitro-2-methylphenol	0.135	0.150	-11.1	119	0.00
66 c	n-Nitrosodiphenylamine	0.624	0.625	-0.2	116	0.00
67	4-Bromophenyl-phenylether	0.226	0.235	-4.0	120	0.00
68	Hexachlorobenzene	0.238	0.248	-4.2	120	0.00
69	Atrazine	0.195	0.195	0.0	115	0.00
70 C	Pentachlorophenol	0.142	0.166	-16.9	125	0.00
71	Phenanthrene	1.036	1.006	2.9	114	0.00
72	Anthracene	1.067	1.044	2.2	115	0.00
73	Carbazole	0.962	0.943	2.0	115	0.00
74	Di-n-butylphthalate	1.128	1.149	-1.9	117	0.00
75 C	Fluoranthene	1.133	1.107	2.3	114	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	112	0.00
77	Benzydine	0.487	0.652	-33.9#	121	0.00
78	Pyrene	1.818	1.962	-7.9	114	0.00
79 S	Terphenyl-d14	1.183	1.236	-4.5	113	0.00
80	Butylbenzylphthalate	0.718	0.792	-10.3	113	0.00
81	Benzo(a)anthracene	1.496	1.543	-3.1	109	0.00
82	3,3'-Dichlorobenzidine	0.441	0.437	0.9	108	0.00
83	Chrysene	1.399	1.427	-2.0	112	0.00
84	Bis(2-ethylhexyl)phthalate	0.881	0.991	-12.5	116	0.00
85 c	Di-n-octyl phthalate	1.288	1.425	-10.6	117	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.310	1.532	-16.9	113	0.00
88	Benzo(b)fluoranthene	1.340	1.350	-0.7	103	0.00
89	Benzo(k)fluoranthene	1.312	1.352	-3.0	107	0.00
90 C	Benzo(a)pyrene	1.254	1.300	-3.7	105	0.00
91	Dibenzo(a,h)anthracene	1.068	1.245	-16.6	111	0.00
92	Benzo(g,h,i)perylene	1.112	1.317	-18.4	114	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF030323\
Data File : BF132623.D
Acq On : 03 Mar 2023 12:07
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Mar 04 02:14:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF022223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 24 23:17:30 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0