

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030420\  
 Data File : BF119202.D  
 Acq On : 4 Mar 2020 18:53  
 Operator : CG/JU  
 Sample : L1774-02  
 Misc :  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 VNJ-223

Quant Time: Mar 05 06:00:54 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Feb 26 16:03:05 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min)  |
|-----------------------------|-------|------|----------|--------|-------|-----------|
| 1) 1,4-Dichlorobenzene-d4   | 6.74  | 152  | 129324   | 20.00  | ng    | 0.00      |
| 21) Naphthalene-d8          | 8.02  | 136  | 482562   | 20.00  | ng    | -0.01     |
| 39) Acenaphthene-d10        | 9.77  | 164  | 246316   | 20.00  | ng    | -0.01     |
| 64) Phenanthrene-d10        | 11.26 | 188  | 397856   | 20.00  | ng    | 0.00      |
| 76) Chrysene-d12            | 13.90 | 240  | 247990   | 20.00  | ng    | 0.00      |
| 87) Perylene-d12            | 15.33 | 264  | 283415   | 20.00  | ng    | 0.00      |
| System Monitoring Compounds |       |      |          |        |       |           |
| 5) 2-Fluorophenol           | 5.38  | 112  | 783836   | 101.29 | ng    | 0.01      |
| 7) Phenol-d6                | 6.40  | 99   | 949034   | 100.84 | ng    | 0.00      |
| 23) Nitrobenzene-d5         | 7.30  | 82   | 626669   | 68.57  | ng    | -0.02     |
| 42) 2,4,6-Tribromophenol    | 10.57 | 330  | 284085   | 88.16  | ng    | 0.00      |
| 45) 2-Fluorobiphenyl        | 9.10  | 172  | 1182259  | 70.71  | ng    | -0.01     |
| 79) Terphenyl-d14           | 12.84 | 244  | 1079558  | 74.95  | ng    | -0.01     |
| Target Compounds            |       |      |          |        |       |           |
| 50) Dimethylphthalate       | 9.49  | 163  | 253350   | 13.451 | ng    | Qvalue 98 |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030420\  
Data File : BF119202.D  
Acq On : 4 Mar 2020 18:53  
Operator : CG/JU  
Sample : L1774-02  
Misc :  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
BNA\_F  
ClientSampleId :  
VNJ-223

Quant Time: Mar 05 06:00:54 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022020.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Feb 26 16:03:05 2020  
Response via : Initial Calibration

