


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF030420\
 Data File : BF119186.D
 Acq On : 4 Mar 2020 9:52
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 26 16:03:05 2020



AutoFind: Scans 1602, 1603, 1604; Background Corrected with Scan 1597

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.5	41618	PASS
68	69	0.00	2	1.7	946	PASS
69	198	0.00	100	34.7	54448	PASS
70	69	0.00	2	0.4	243	PASS
127	198	10	80	46.4	72674	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	156778	PASS
199	198	5	9	6.8	10598	PASS
275	198	10	60	31.5	49365	PASS
365	198	1	100	4.3	6722	PASS
441	198	0.01	100	12.7	19953	PASS
442	198	50	100	80.4	125976	PASS
443	442	15	24	18.9	23800	PASS