


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF030520\
 Data File : BF119208.D
 Acq On : 5 Mar 2020 10:22
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF022020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Feb 26 16:03:05 2020



AutoFind: Scans 1600, 1601, 1602; Background Corrected with Scan 1594

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.9	47869	PASS
68	69	0.00	2	1.2	734	PASS
69	198	0.00	100	33.7	62384	PASS
70	69	0.00	2	0.5	310	PASS
127	198	10	80	45.6	84488	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	185130	PASS
199	198	5	9	6.6	12253	PASS
275	198	10	60	32.6	60389	PASS
365	198	1	100	4.3	7945	PASS
441	198	0.01	100	13.1	24232	PASS
442	198	50	100	85.0	157296	PASS
443	442	15	24	18.8	29531	PASS