

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030619\
 Data File : BF112893.D
 Acq On : 6 Mar 2019 20:33
 Operator : JU/SJ
 Sample : PB117655BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB117655BL

Quant Time: Mar 07 04:37:37 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	240744	20.00	ng	-0.01
21) Naphthalene-d8	8.01	136	782077	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	369373	20.00	ng	0.00
64) Phenanthrene-d10	11.24	188	842576	20.00	ng	0.00
76) Chrysene-d12	13.86	240	515834	20.00	ng	-0.01
87) Perylene-d12	15.26	264	526387	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.35	112	1133801	73.26	ng	0.00
7) Phenol-d6	6.37	99	1451483	74.66	ng	0.00
23) Nitrobenzene-d5	7.29	82	887503	67.90	ng	-0.02
42) 2,4,6-Tribromophenol	10.55	330	373007	95.12	ng	-0.01
45) 2-Fluorobiphenyl	9.09	172	1443675	59.90	ng	-0.01
79) Terphenyl-d14	12.82	244	1394580	52.41	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030619\
Data File : BF112893.D
Acq On : 6 Mar 2019 20:33
Operator : JU/SJ
Sample : PB117655BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB117655BL

Quant Time: Mar 07 04:37:37 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

