

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030619\
 Data File : BF112913.D
 Acq On : 7 Mar 2019 6:11
 Operator : JU/SJ
 Sample : K1705-13 2X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EXT-027-SW024-022719

Manual Integrations
 APPROVED

Sohil
 3/7/2019 11:03:11 AM

Quant Time: Mar 07 06:45:45 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	176868	20.00	ng	0.00
21) Naphthalene-d8	8.01	136	598110	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	273292	20.00	ng	0.00
64) Phenanthrene-d10	11.24	188	564543	20.00	ng	0.00
76) Chrysene-d12	13.86	240	556067	20.00	ng	-0.01
87) Perylene-d12	15.27	264	402851	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	522061	45.91	ng	0.00
7) Phenol-d6	6.36	99	635174	44.47	ng	0.00
23) Nitrobenzene-d5	7.29	82	348202	34.84	ng	-0.02
42) 2,4,6-Tribromophenol	10.55	330	169857	58.54	ng	-0.01
45) 2-Fluorobiphenyl	9.09	172	621657	31.68	ng	-0.01
79) Terphenyl-d14	12.82	244	719700	25.09	ng	-0.01
Target Compounds						
71) Phenanthrene	11.26	178	100598	3.582	ng	98
75) Fluoranthene	12.45	202	155153	5.156	ng	95
78) Pyrene	12.67	202	134138	2.861	ng	99
88) Benzo(b)fluoranthene	14.87	252	62505m	2.399	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF030619\
Data File : BF112913.D
Acq On : 7 Mar 2019 6:11
Operator : JU/SJ
Sample : K1705-13 2X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EXT-027-SW024-022719

Manual Integrations
APPROVED

Sohil
3/7/2019 11:03:11 AM

Quant Time: Mar 07 06:45:45 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

