



Benzaldehyde

Response Ratio



$R = -1.245e-001 A^2 + 1.120e+000 A + 5.911e-002$
Coef of Det (r^2) = 0.993511 Curve Fit: Quadratic
Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF031022.M
Calibration Table Last Updated: Thu Mar 10 16:51:53 2022


Benzoic acid



$\text{Response} = 2.296\text{e-}001 * \text{Amt} - 7.604\text{e-}002$
 Coef of Det (r^2) = 0.996623 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF031022.M
 Calibration Table Last Updated: Thu Mar 10 16:51:53 2022

Hexachlorocyclopentadiene

Response Ratio




$$\text{Response} = 2.854\text{e-}001 * \text{Amt} - 8.480\text{e-}002$$

Coef of Det (r^2) = 0.998648 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF031022.M

Calibration Table Last Updated: Thu Mar 10 16:51:53 2022

2,4-Dinitrophenol



Response = 1.678e-001 * Amt - 3.227e-002
 Coef of Det (r^2) = 0.996781 Curve Fit: Linear
 Method Name: Z:\svoasrv\HPCHEM1\BNA F\Methods\8270-BF031022.M
 Calibration Table Last Updated: Thu Mar 10 16:51:53 2022