

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031022\
 Data File : BF127291.D
 Acq On : 10 Mar 2022 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF031022

Quant Time: Mar 10 17:51:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	97	0.00
2	1,4-Dioxane	40.000	38.666	3.3	98	0.00
3	Pyridine	40.000	38.376	4.1	92	0.00
4	n-Nitrosodimethylamine	40.000	40.185	-0.5	94	0.00
5 S	2-Fluorophenol	80.000	76.119	4.9	96	0.00
6	Aniline	40.000	40.187	-0.5	95	0.00
7 S	Phenol-d6	80.000	77.550	3.1	95	0.00
8	2-Chlorophenol	40.000	39.703	0.7	96	0.00
9	Benzaldehyde	40.000	45.464	-13.7	104	0.00
10 C	Phenol	40.000	38.659	3.4	93	0.00
11	bis(2-Chloroethyl)ether	40.000	39.835	0.4	94	0.00
12	1,3-Dichlorobenzene	40.000	39.254	1.9	96	0.00
13 C	1,4-Dichlorobenzene	40.000	39.992	0.0	96	0.00
14	1,2-Dichlorobenzene	40.000	38.223	4.4	92	0.00
15	Benzyl Alcohol	40.000	40.942	-2.4	95	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.688	3.3	91	0.00
17	2-Methylphenol	40.000	38.165	4.6	92	0.00
18	Hexachloroethane	40.000	39.722	0.7	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.864	2.8	91	0.00
20	3+4-Methylphenols	40.000	37.374	6.6	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	99	0.00
22	Acetophenone	40.000	36.905	7.7	91	0.00
23 S	Nitrobenzene-d5	80.000	76.440	4.5	94	0.00
24	Nitrobenzene	40.000	39.017	2.5	95	0.00
25	Isophorone	40.000	39.075	2.3	97	0.00
26 C	2-Nitrophenol	40.000	41.829	-4.6	104	0.00
27	2,4-Dimethylphenol	40.000	40.254	-0.6	100	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.512	3.7	93	0.00
29 C	2,4-Dichlorophenol	40.000	39.159	2.1	94	0.00
30	1,2,4-Trichlorobenzene	40.000	40.890	-2.2	100	0.00
31	Naphthalene	40.000	39.434	1.4	99	0.00
32	Benzoic acid	40.000	41.173	-2.9	108	0.00
33	4-Chloroaniline	40.000	38.821	2.9	97	0.00
34 C	Hexachlorobutadiene	40.000	40.338	-0.8	98	0.00
35	Caprolactam	40.000	42.245	-5.6	97	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.434	-1.1	97	0.00
37	2-Methylnaphthalene	40.000	41.081	-2.7	101	0.00
38	1-Methylnaphthalene	40.000	40.009	-0.0	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.644	-4.1	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.002	-5.0	98	0.00
42 S	2,4,6-Tribromophenol	80.000	83.157	-3.9	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.178	-5.4	98	0.00
44	2,4,5-Trichlorophenol	40.000	41.558	-3.9	98	0.00
45 S	2-Fluorobiphenyl	80.000	80.562	-0.7	99	0.00
46	1,1'-Biphenyl	40.000	40.105	-0.3	98	0.00
47	2-Chloronaphthalene	40.000	40.408	-1.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031022\
 Data File : BF127291.D
 Acq On : 10 Mar 2022 17:22
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF031022

Quant Time: Mar 10 17:51:10 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Mar 10 16:51:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.347	-5.9	100	0.00
49	Acenaphthylene	40.000	39.950	0.1	94	0.00
50	Dimethylphthalate	40.000	42.966	-7.4	105	0.00
51	2,6-Dinitrotoluene	40.000	41.818	-4.5	98	0.00
52 C	Acenaphthene	40.000	39.963	0.1	97	0.00
53	3-Nitroaniline	40.000	39.309	1.7	94	0.00
54 P	2,4-Dinitrophenol	40.000	40.832	-2.1	94	0.00
55	Dibenzofuran	40.000	40.320	-0.8	97	0.00
56 P	4-Nitrophenol	40.000	42.633	-6.6	93	0.00
57	2,4-Dinitrotoluene	40.000	41.286	-3.2	96	0.00
58	Fluorene	40.000	40.102	-0.3	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.652	-4.1	97	0.00
60	Diethylphthalate	40.000	41.389	-3.5	100	0.00
61	4-Chlorophenyl-phenylether	40.000	41.106	-2.8	97	0.00
62	4-Nitroaniline	40.000	41.088	-2.7	96	0.00
63	Azobenzene	40.000	41.707	-4.3	96	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	98	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.760	-11.9	98	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.084	-0.2	99	0.00
67	4-Bromophenyl-phenylether	40.000	40.137	-0.3	96	0.00
68	Hexachlorobenzene	40.000	40.035	-0.1	94	0.00
69	Atrazine	40.000	41.634	-4.1	96	0.00
70 C	Pentachlorophenol	40.000	44.013	-10.0	98	0.00
71	Phenanthrene	40.000	38.858	2.9	97	0.00
72	Anthracene	40.000	40.308	-0.8	100	0.00
73	Carbazole	40.000	39.302	1.7	97	0.00
74	Di-n-butylphthalate	40.000	40.399	-1.0	100	0.00
75 C	Fluoranthene	40.000	40.354	-0.9	99	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
77	Benzydine	40.000	37.648	5.9	108	0.00
78	Pyrene	40.000	40.441	-1.1	100	0.00
79 S	Terphenyl-d14	80.000	81.267	-1.6	101	0.00
80	Butylbenzylphthalate	40.000	41.643	-4.1	99	0.00
81	Benzo(a)anthracene	40.000	40.194	-0.5	97	0.00
82	3,3'-Dichlorobenzidine	40.000	38.964	2.6	98	0.00
83	Chrysene	40.000	39.380	1.5	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.595	-1.5	96	0.00
85 c	Di-n-octyl phthalate	40.000	45.021	-12.6	109	0.00
86 I	Perylene-d12	20.000	20.000	0.0	107	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.149	9.6	94	0.00
88	Benzo(b)fluoranthene	40.000	39.931	0.2	106	0.00
89	Benzo(k)fluoranthene	40.000	41.225	-3.1	112	0.00
90 C	Benzo(a)pyrene	40.000	40.531	-1.3	105	0.00
91	Dibenzo(a,h)anthracene	40.000	36.248	9.4	94	0.00
92	Benzo(g,h,i)perylene	40.000	37.733	5.7	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF031022\
Data File : BF127291.D
Acq On : 10 Mar 2022 17:22
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF031022

Quant Time: Mar 10 17:51:10 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF031022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Mar 10 16:51:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0