

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031119\
 Data File : BF112977.D
 Acq On : 12 Mar 2019 1:33
 Operator : JU/SJ
 Sample : K1822-02 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 3-8-BOD

Manual Integrations
 APPROVED

Sohil
 3/12/2019 4:29:05 PM

Quant Time: Mar 12 07:11:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	181236	20.00	ng	0.00
21) Naphthalene-d8	8.02	136	665570	20.00	ng	0.00
39) Acenaphthene-d10	9.78	164	201137	20.00	ng	0.01
64) Phenanthrene-d10	11.29	188	235214	20.00	ng	0.04
76) Chrysene-d12	13.89	240	469938	20.00	ng	0.01
87) Perylene-d12	15.28	264	453669	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	17314	1.49	ng	-0.01
7) Phenol-d6	6.36	99	29013	1.98	ng	0.00
23) Nitrobenzene-d5	7.29	82	36997	3.33	ng	-0.02
42) 2,4,6-Tribromophenol	10.57	330	3848	1.80	ng	0.02
45) 2-Fluorobiphenyl	9.09	172	42453	1.34	ng	0.00
79) Terphenyl-d14	12.85	244	27497	1.13	ng	0.02
Target Compounds						
31) Naphthalene	8.04	128	267461	8.094	ng	99
37) 2-Methylnaphthalene	8.73	142	290968	13.635	ng	99
38) 1-Methylnaphthalene	8.83	142	149697m	7.242	ng	
46) 1,1'-Biphenyl	9.20	154	68020	4.146	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031119\
Data File : BF112977.D
Acq On : 12 Mar 2019 1:33
Operator : JU/SJ
Sample : K1822-02 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
3-8-BOD

Manual Integrations
APPROVED

Sohil
3/12/2019 4:29:05 PM

Quant Time: Mar 12 07:11:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration

