

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031119\
 Data File : BF112992.D
 Acq On : 12 Mar 2019 8:57
 Operator : JU/SJ
 Sample : K1817-11
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PST-028-B009-030619

Quant Time: Mar 12 09:42:44 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Mar 01 15:40:43 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.73	152	98753	20.00	ng	-0.01
21) Naphthalene-d8	8.01	136	392033	20.00	ng	-0.01
39) Acenaphthene-d10	9.76	164	213248	20.00	ng	0.00
64) Phenanthrene-d10	11.24	188	470250	20.00	ng	0.00
76) Chrysene-d12	13.86	240	345892	20.00	ng	-0.01
87) Perylene-d12	15.27	264	260627	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.33	112	588155	92.65	ng	-0.02
7) Phenol-d6	6.36	99	779938	97.80	ng	0.00
23) Nitrobenzene-d5	7.29	82	448791	68.50	ng	-0.02
42) 2,4,6-Tribromophenol	10.55	330	277465	122.56	ng	-0.01
45) 2-Fluorobiphenyl	9.09	172	827016	59.33	ng	-0.01
79) Terphenyl-d14	12.82	244	957719	53.67	ng	-0.01
Target Compounds						
10) Phenol	6.37	94	25304	2.719	ng	93
50) Dimethylphthalate	9.48	163	91845	5.841	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF031119\
Data File : BF112992.D
Acq On : 12 Mar 2019 8:57
Operator : JU/SJ
Sample : K1817-11
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PST-028-B009-030619

Quant Time: Mar 12 09:42:44 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF022719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Mar 01 15:40:43 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.73 min Scan# 789
Delta R.T. -0.01 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Instrument :
BNA_F
ClientSampleId :
PST-028-B009-030619

Tot Ion:152 Resp: 98753
Ion Ratio Lower Upper
152 100
150 160.1 124.2 186.2
115 59.4 48.2 72.4




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 92.645 ng
RT: 5.33 min Scan# 551
Delta R.T. -0.02 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Tgt Ion:112 Resp: 588155
Ion Ratio Lower Upper
112 100
64 58.9 47.6 71.4
63 29.3 24.4 36.6



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 97.802 ng

RT: 6.36 min Scan# 727

Delta R.T. -0.01 min

Lab File: BF112992.D

Acq: 12 Mar 2019 8:57

Instrument :

BNA_F

ClientSampleId :

PST-028-B009-030619


Tot Ion: 99 Resp: 779938

Ion Ratio Lower Upper




99 100

42 17.8 16.3 24.5

71 30.7 26.2 39.4



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 2.719 ng

RT: 6.37 min Scan# 729

Delta R.T. -0.01 min

Lab File: BF112992.D

Acq: 12 Mar 2019 8:57


Tgt Ion: 94 Resp: 25304

Ion Ratio Lower Upper



94 100


65 26.8 10.5 50.5

66 38.2 22.6 62.6



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.01 min Scan# 1007
Delta R.T. -0.01 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Instrument :
BNA_F
ClientSampleId :
PST-028-B009-030619

Tot Ion:136	Resp:	392033
Ion Ratio	Lower	Upper
136	100	
137	11.0	8.9 13.3
54	8.1	7.3 10.9
68	6.5	5.4 8.2




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF1
Ion 68.00 (67.70 to 68.70): BF1






#23
Nitrobenzene-d5
Concen: 68.501 ng
RT: 7.29 min Scan# 884
Delta R.T. -0.02 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Tgt Ion: 82	Resp:	448791
Ion Ratio	Lower	Upper
82	100	
128	46.0	36.3 54.5
54	50.6	43.7 65.5



Abundance Ion 82.00 (81.70 to 82.70): BF1
Ion 128.00 (127.70 to 128.70): E
Ion 54.00 (53.70 to 54.70): BF1





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.76 min Scan# 1305

Delta R.T. -0.01 min

Lab File: BF112992.D

Acq: 12 Mar 2019 8:57

Instrument :

BNA_F

ClientSampleId :

PST-028-B009-030619


Tot Ion:164 Resp: 213248

Ion Ratio Lower Upper

164 100

162 101.3 80.6 120.8




160 44.8 36.0 54.0



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#42

2,4,6-Tribromophenol

Concen: 122.561 ng

RT: 10.55 min Scan# 1438

Delta R.T. -0.01 min

Lab File: BF112992.D

Acq: 12 Mar 2019 8:57


Tgt Ion:330 Resp: 277465

Ion Ratio Lower Upper

330 100

332 97.1 76.2 114.4



141 28.5 27.0 40.6




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#45
2-Fluorobiphenyl
Concen: 59.333 ng
RT: 9.09 min Scan# 1190
Delta R.T. -0.01 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Instrument :
BNA_F
ClientSampleId :
PST-028-B009-030619

Tot Ion:172 Resp: 827016
Ion Ratio Lower Upper
172 100
171 34.7 29.0 43.4
170 23.9 20.0 30.0




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#50
Dimethylphthalate
Concen: 5.841 ng
RT: 9.48 min Scan# 1257
Delta R.T. -0.02 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Tgt Ion:163 Resp: 91845
Ion Ratio Lower Upper
163 100
194 4.1 3.0 4.4
164 9.3 8.4 12.6



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E









#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.24 min Scan# 1556
Delta R.T. -0.01 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57




Instrument :
BNA_F
ClientSampleId :
PST-028-B009-030619


Tot Ion:188 Resp: 470250
Ion Ratio Lower Upper
188 100
94 9.3 9.0 13.4
80 9.7 9.2 13.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 13.86 min Scan# 2002
Delta R.T. -0.01 min
Lab File: BF112992.D
Acq: 12 Mar 2019 8:57

Tgt Ion:240 Resp: 345892
Ion Ratio Lower Upper
240 100
120 12.0 8.9 13.3
236 25.6 20.7 31.1









#79
 Terphenyl-d14
 Concen: 53.674 ng
 RT: 12.82 min Scan# 1824
 Delta R.T. -0.01 min
 Lab File: BF112992.D
 Acq: 12 Mar 2019 8:57

Instrument :
 BNA_F
 ClientSampleId :
 PST-028-B009-030619

Tot Ion:244 Resp: 957719
 Ion Ratio Lower Upper
 244 100
 212 6.4 5.6 8.4
 122 11.8 10.5 15.7



#87
 Perylene-d12
 Concen: 20.000 ng
 RT: 15.27 min Scan# 2241
 Delta R.T. -0.01 min
 Lab File: BF112992.D
 Acq: 12 Mar 2019 8:57

Tgt Ion:264 Resp: 260627
 Ion Ratio Lower Upper
 264 100
 260 23.6 18.7 28.1
 265 21.7 17.2 25.8

